

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062419\
 Data File : BN006580.D
 Acq On : 25 Jun 2019 20:52
 Operator : HP/JU
 Sample : K3388-02DL 5X
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A4215DL

Manual Integrations
 APPROVED

mohammad
 6/26/2019 8:12:58 AM

Quant Time: Jun 26 01:22:33 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 25 07:15:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	5430	0.40	ng/ul	0.00
2) Naphthalene-d8	10.41	136	19861	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.28	164	9399	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.04	188	17006	0.40	ng/ul	0.00
16) Chrysene-d12	21.26	240	10035	0.40	ng/ul	0.00
20) Perylene-d12	23.49	264	11167m	0.40	ng/ul	0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.01	152	996	0.03	ng/ul	0.00
14) Fluoranthene-d10	19.08	212	1593	0.03	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.46	128	3459	0.063	ng/ul#	92
5) 2-Methylnaphthalene	12.08	142	1664	0.044	ng/ul	98
7) Acenaphthylene	14.00	152	13031	0.249	ng/ul	99
8) Acenaphthene	14.34	153	1976	0.055	ng/ul	98
9) Fluorene	15.34	166	2631	0.062	ng/ul#	95
12) Phenanthrene	17.08	178	27169	0.498	ng/ul	99
13) Anthracene	17.17	178	9349	0.179	ng/ul	96
15) Fluoranthene	19.11	202	42399	0.703	ng/ul	98
17) Pyrene	19.48	202	39040	0.739	ng/ul	98
18) Benzo(a)anthracene	21.24	228	18264	0.412	ng/ul	91
19) Chrysene	21.29	228	20783m	0.499	ng/ul	
21) Benzo(b)fluoranthene	22.83	252	43983m	0.945	ng/ul	
22) Benzo(k)fluoranthene	22.86	252	13817m	0.305	ng/ul	
23) Benzo(a)pyrene	23.40	252	34169	0.788	ng/ul#	91
24) Indeno(1,2,3-cd)pyrene	25.72	276	39328	0.838	ng/ul#	87
25) Dibenzo(a,h)anthracene	25.72	278	8210	0.219	ng/ul#	84
26) Benzo(a,h,i)perylene	26.41	276	41640	1.059	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062419\
Data File : BN006580.D
Acq On : 25 Jun 2019 20:52
Operator : HP/JU
Sample : K3388-02DL 5X
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A4215DL

Manual Integrations
APPROVED

mohammad
6/26/2019 8:12:58 AM

Quant Time: Jun 26 01:22:33 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 25 07:15:41 2019
Response via : Initial Calibration

