

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062423\
 Data File : BN026017.D
 Acq On : 24 Jun 2023 22:53
 Operator : MA/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4313

Quant Time: Jun 25 00:57:28 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 22 01:27:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.980	152	6059	0.400	ng/ul	0.00
4) Naphthalene-d8	10.792	136	20314	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.616	164	12684	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.359	188	25964	0.400	ng/ul	0.00
17) Chrysene-d12	21.541	240	14958	0.400	ng/ul	0.00
23) Perylene-d12	23.999	264	13162	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.331	96	2508	0.368	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.376	152	12428	0.398	ng/ul	0.00
18) Fluoranthene-d10	19.387	212	25558	0.449	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.364	88	2815	0.389	ng/ul	98
5) Naphthalene	10.842	128	22697	0.390	ng/ul	100
7) 2-Methylnaphthalene	12.447	142	14620	0.394	ng/ul	100
8) 1-Methylnaphthalene	12.667	142	15637	0.407	ng/ul	99
10) Acenaphthylene	14.339	152	24225	0.408	ng/ul	99
11) Acenaphthene	14.677	153	19063	0.392	ng/ul	98
12) Fluorene	15.662	166	21255	0.382	ng/ul	98
14) Pentachlorophenol	17.017	266	2322	0.249	ng/ul	99
15) Phenanthrene	17.402	178	32814	0.383	ng/ul	100
16) Anthracene	17.495	178	29229	0.383	ng/ul	99
19) Fluoranthene	19.414	202	35237	0.441	ng/ul	99
20) Pyrene	19.777	202	35818	0.431	ng/ul	99
21) Benzo(a)anthracene	21.523	228	23384	0.379	ng/ul	99
22) Chrysene	21.581	228	24109	0.375	ng/ul	100
24) Benzo(b)fluoranthene	23.247	252	21291	0.379	ng/ul	99
25) Benzo(k)fluoranthene	23.297	252	21791	0.386	ng/ul	100
26) Benzo(a)pyrene	23.888	252	18901	0.387	ng/ul	97
27) Indeno(1,2,3-cd)pyrene	26.564	276	23677	0.444	ng/ul#	93
28) Dibenzo(a,h)anthracene	26.581	278	18401	0.456	ng/ul	96
29) Benzo(g,h,i)perylene	27.349	276	20570	0.434	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062423\
Data File : BN026017.D
Acq On : 24 Jun 2023 22:53
Operator : MA/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4313

Quant Time: Jun 25 00:57:28 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 22 01:27:41 2023
Response via : Initial Calibration

