


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062518\
 Data File : BN001902.D
 Acq On : 25 Jun 2018 09:16
 Operator : JU/SJ
 Sample : DFTPP82
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP82

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SOM-EPA-BN061918.M
 Title : SVOA CALIBRATION
 Last Update : Tue Jun 26 04:45:44 2018



AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2468

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.1	245312	PASS
68	69	0.00	2	1.6	4021	PASS
69	198	0.00	100	35.9	244032	PASS
70	69	0.00	2	0.6	1399	PASS
127	198	10	80	45.0	305301	PASS
197	198	0.00	2	0.4	3016	PASS
198	198	100	100	100.0	679018	PASS
199	198	5	9	6.7	45245	PASS
275	198	10	60	25.5	173285	PASS
365	198	1	100	3.0	20148	PASS
441	443	0.01	100	82.5	89816	PASS
442	198	50	100	84.6	574378	PASS
443	442	15	24	19.0	108882	PASS

m/z	Abundance
36.00	302.0
37.00	633.0
38.00	2348.0
39.00	15104.0
39.90	2839.0
41.10	588.0
44.00	13892.0
45.00	573.0
49.00	1528.0
50.00	53720.0
51.00	222272.0
52.00	10666.0
53.00	567.0
55.00	1432.0
56.00	6898.0
57.00	16416.0
58.00	659.0
61.00	2596.0
62.00	2806.0
63.00	9071.0
64.00	1435.0
65.00	4382.0
65.90	397.0
67.00	420.0
68.00	3477.0
69.00	219648.0
70.00	1185.0
73.00	1578.0
74.00	21688.0
75.00	34872.0
76.00	10957.0
77.00	249728.0
78.00	17656.0
79.00	14477.0
80.00	12987.0
81.00	16400.0
81.90	4300.0
83.00	3994.0
83.90	715.0
84.90	3453.0
86.00	5314.0
87.00	2362.0
87.90	822.0
89.00	396.0
91.00	3631.0
92.00	4170.0
93.00	27168.0
94.00	2173.0
94.90	585.0
96.00	1519.0
96.90	472.0
98.00	18848.0
98.90	19128.0
99.90	1617.0
101.00	9937.0
102.00	633.0
102.90	2930.0
103.90	5704.0
105.00	5849.0
106.00	1601.0
107.00	73800.0
108.00	11882.0
109.00	1738.0
110.00	137024.0
111.00	20816.0
111.90	2808.0
112.90	758.0
114.70	447.0
116.00	3719.0
117.00	52720.0
117.90	3557.0
119.00	772.0
120.00	999.0
122.00	4880.0
123.00	7404.0
124.00	2899.0
124.90	3336.0
127.00	258496.0
128.00	19000.0
129.00	96392.0
130.00	8744.0
131.00	1992.0
132.10	902.0
132.90	547.0
133.90	3088.0
134.90	7812.0
135.90	3066.0
137.00	5742.0
137.80	964.0
138.90	540.0
140.00	1093.0

Instrument :
BNA_N
ClientSampleId :
DFTPP82

141.00	12782.0
142.00	3987.0
143.00	2164.0
144.00	576.0
144.80	581.0
146.00	2070.0
147.00	6292.0
147.90	12998.0
149.00	3231.0
149.90	938.0
151.10	1735.0
151.90	1225.0
153.00	3683.0
153.90	2897.0
155.00	7637.0
156.00	11555.0
157.10	1557.0
157.90	2114.0
158.90	2159.0
160.00	3514.0
160.90	4970.0
161.90	1706.0
162.90	418.0
164.00	811.0
164.90	4500.0
166.00	3484.0
167.00	21768.0
168.00	16736.0
169.00	2900.0
170.00	1001.0
170.90	1129.0
171.90	1773.0
173.00	2574.0
174.00	6162.0
175.00	8736.0
176.00	2974.0
176.90	4879.0
178.00	1695.0
178.90	19672.0
180.00	15042.0
180.90	6381.0
181.90	967.0
182.90	648.0
184.00	1149.0
185.00	9071.0
186.00	66840.0
187.00	19856.0
188.10	2376.0
188.90	4404.0
189.90	647.0
191.00	2381.0
192.00	5970.0
192.90	6129.0
193.90	1576.0
195.00	853.0
196.00	14920.0
196.90	2612.0
197.90	528576.0
198.90	38120.0
199.90	2432.0
201.50	2477.0
202.90	4015.0
204.00	17520.0
205.00	30064.0
206.00	135104.0
207.00	19560.0
208.00	5738.0
208.90	1562.0
209.80	1892.0
211.00	5482.0
212.00	901.0
212.90	423.0
214.90	1686.0
216.00	2658.0
216.90	34464.0
218.00	4757.0
218.90	406.0
221.00	22784.0
223.00	7882.0
224.00	73912.0
225.00	18568.0
226.00	2380.0
227.00	28424.0
228.00	4290.0
229.00	6109.0
230.10	754.0
231.10	2077.0
231.90	455.0
233.00	549.0
233.90	2426.0
235.10	1936.0
235.90	1303.0
237.00	3205.0
237.90	645.0
238.80	940.0

Instrument :
BNA_N
ClientSampleId :
DFTPP82

239.90	1122.0
240.90	1455.0
242.00	3936.0
243.00	4239.0
244.00	53768.0
245.00	7936.0
246.00	9073.0
246.90	2336.0
247.80	533.0
248.90	1913.0
250.00	521.0
250.70	401.0
251.90	551.0
252.90	764.0
254.00	1940.0
255.00	277888.0
256.00	44152.0
256.90	3677.0
258.00	15425.0
258.90	2687.0
259.80	477.0
260.10	510.0
260.90	647.0
263.90	321.0
265.00	6230.0
265.90	999.0
270.90	626.0
271.90	570.0
273.00	7594.0
274.00	20816.0
275.00	121776.0
276.00	17920.0
276.90	10002.0
277.90	1865.0
279.00	486.0
281.10	661.0
282.10	456.0
283.00	1485.0
284.00	805.0
284.90	1657.0
289.90	438.0
292.00	632.0
293.00	2198.0
294.00	828.0
295.10	988.0
296.00	30872.0
297.00	4182.0
298.00	346.0
301.90	547.0
302.90	3980.0
304.00	1124.0
308.10	507.0
308.90	428.0
309.90	691.0
313.90	1581.0
314.90	3357.0
316.00	2031.0
316.90	733.0
320.90	1049.0
322.00	484.0
323.00	11128.0
324.00	2007.0
325.00	468.0
327.00	2811.0
328.00	834.0
331.90	582.0
333.00	904.0
334.00	7112.0
335.00	2178.0
336.00	327.0
341.00	1331.0
345.90	2451.0
346.80	558.0
352.00	2806.0
353.00	1673.0
354.00	3241.0
355.10	502.0
359.00	305.0
364.90	11694.0
365.90	2358.0
370.90	603.0
372.00	4764.0
373.10	1335.0
383.00	1690.0
384.00	446.0
390.00	945.0
391.00	489.0
392.00	366.0
401.10	327.0
402.00	2282.0
403.00	2707.0
403.90	1371.0
405.00	489.0
421.00	2587.0
422.00	2239.0

Instrument :

BNA_N

ClientSampleId :

DFTPP82

423.10	18408.0
424.00	6283.0
425.00	1126.0
441.00	52784.0
442.00	344128.0
443.00	64888.0
444.00	6839.0
445.00	404.0

Instrument :
BNA_N
ClientSampleId :
DFTPP82

m/z	Abundance
37.00	895.0
38.10	3224.0
39.00	20456.0
40.00	2462.0
41.00	858.0
43.20	338.0
44.00	12734.0
44.90	942.0
49.00	2135.0
50.00	68160.0
51.00	292096.0
52.00	14923.0
53.00	523.0
55.00	1369.0
56.00	8463.0
57.00	20720.0
57.90	782.0
58.90	313.0
61.00	3868.0
62.00	4641.0
63.00	11678.0
64.00	1971.0
65.00	5797.0
66.00	475.0
66.90	491.0
68.00	5149.0
69.00	289920.0
70.00	1907.0
73.00	2218.0
74.00	29736.0
75.00	45400.0
76.00	14127.0
77.00	331776.0
78.00	23584.0
79.00	20616.0
80.00	15456.0
81.00	22984.0
82.00	5701.0
83.00	5274.0
83.90	716.0
85.00	4007.0
86.00	7099.0
87.00	3077.0
87.90	1067.0
88.90	619.0
90.90	5003.0
92.00	6193.0
93.00	35208.0
94.00	1819.0
94.90	756.0
96.00	2110.0
97.00	559.0
98.00	25072.0
99.00	23752.0
100.00	2461.0
101.00	13575.0
101.90	730.0
103.00	4435.0
104.00	8466.0
105.00	6934.0
106.00	2575.0
107.00	98848.0
108.00	16339.0
109.00	2306.0
110.00	186048.0
111.00	28376.0
111.90	3999.0
112.90	1007.0
115.00	425.0
116.00	5540.0
117.00	73736.0
117.90	5417.0
118.90	1026.0
119.90	884.0
121.10	392.0
122.00	5413.0
123.00	10499.0
123.90	4551.0
124.90	4705.0
126.00	1147.0
127.00	358976.0
128.00	26384.0
129.00	131520.0
130.00	11406.0
130.90	2221.0
131.90	1235.0
132.90	549.0
133.90	4200.0
135.00	10481.0
136.00	5073.0
137.00	6132.0

Instrument :
BNA_N
ClientSampleId :
DFTPP82

137.90	1049.0
139.00	605.0
139.90	2173.0
141.00	16369.0
141.90	4886.0
142.90	3632.0
143.90	965.0
145.00	1290.0
146.00	2965.0
147.00	8927.0
148.00	20080.0
148.90	3734.0
149.90	1130.0
151.00	2152.0
151.90	1501.0
152.90	5828.0
153.90	4038.0
155.00	10021.0
156.00	15544.0
157.00	3152.0
157.90	3538.0
158.90	2334.0
159.90	5723.0
161.00	8852.0
162.00	2161.0
162.90	858.0
164.00	1256.0
165.00	6882.0
166.00	5230.0
167.00	34032.0
168.00	20496.0
169.00	3193.0
170.00	1145.0
170.90	1610.0
172.00	3570.0
172.90	4112.0
174.00	7674.0
175.00	15402.0
176.00	3964.0
176.90	6130.0
177.90	2189.0
179.00	27040.0
180.00	20824.0
181.00	9317.0
182.00	1288.0
182.80	616.0
183.90	2441.0
185.00	13318.0
186.00	101648.0
187.00	32208.0
188.00	3343.0
189.00	6552.0
189.90	1014.0
191.00	3099.0
191.90	8371.0
193.00	9301.0
194.00	2123.0
195.10	956.0
196.00	22152.0
196.90	3299.0
197.90	796288.0
198.90	52800.0
199.90	3793.0
201.40	2731.0
202.90	5736.0
204.00	26184.0
205.00	45800.0
206.00	201088.0
207.00	26216.0
208.00	7329.0
209.00	2375.0
211.00	7561.0
212.00	1276.0
212.90	641.0
213.90	311.0
215.00	2781.0
216.00	3905.0
216.90	54176.0
218.00	6380.0
218.90	869.0
221.00	33008.0
223.00	12250.0
224.00	109528.0
225.00	28504.0
226.00	2932.0
227.00	46632.0
228.00	6146.0
229.00	9599.0
230.00	1367.0
231.00	3834.0
232.10	731.0
232.90	633.0
233.90	2858.0
234.90	3521.0
235.90	2154.0

Instrument :
BNA_N
ClientSampleId :
DFTPP82

236.90	4220.0
238.00	748.0
238.90	1756.0
240.00	1461.0
241.00	2481.0
242.00	5876.0
243.00	5999.0
244.00	84000.0
245.00	10673.0
246.00	15626.0
247.00	3089.0
248.00	735.0
249.00	3700.0
249.90	514.0
251.00	855.0
252.00	856.0
253.00	1782.0
255.00	447424.0
256.00	65824.0
257.00	4769.0
258.00	24736.0
259.00	4321.0
259.80	642.0
261.10	698.0
263.90	414.0
264.90	8616.0
265.90	1192.0
269.70	347.0
271.00	822.0
271.90	882.0
272.90	13834.0
274.00	34408.0
275.00	204672.0
276.00	26200.0
277.00	15133.0
278.00	2394.0
279.00	692.0
281.00	491.0
281.90	735.0
283.00	2099.0
284.00	930.0
285.00	2999.0
286.00	476.0
288.90	602.0
289.70	339.0
289.90	338.0
290.80	558.0
292.10	569.0
293.00	3590.0
293.90	1249.0
295.00	1005.0
296.00	52672.0
297.00	7374.0
297.80	583.0
300.90	637.0
301.90	954.0
303.00	5980.0
304.00	1998.0
305.00	383.0
308.10	685.0
309.00	499.0
309.90	584.0
312.90	551.0
314.00	2078.0
315.00	5696.0
316.00	3835.0
317.00	416.0
320.90	1755.0
322.00	903.0
323.00	16776.0
324.00	3425.0
324.90	525.0
327.00	3239.0
328.00	1518.0
332.00	1484.0
333.00	1453.0
334.00	10847.0
335.00	2827.0
335.90	315.0
339.00	604.0
341.00	2394.0
342.00	624.0
346.00	3597.0
346.90	626.0
350.90	421.0
352.00	4988.0
353.00	3979.0
354.00	5647.0
355.10	1064.0
359.00	365.0
364.90	24592.0
366.00	3499.0
370.00	624.0
370.90	997.0
372.00	9663.0

Instrument :
BNA_N
ClientSampleId :
DFTPP82

373.00	2364.0
373.90	499.0
383.00	2747.0
383.90	698.0
384.90	319.0
389.90	991.0
391.00	974.0
391.90	707.0
401.00	861.0
402.00	4179.0
403.00	5595.0
404.00	2060.0
405.00	652.0
421.00	4993.0
422.00	4745.0
423.00	34520.0
424.00	10639.0
425.10	1403.0
441.00	100592.0
442.00	636672.0
443.00	124096.0
444.00	12718.0
445.00	358.0

Instrument :
BNA_N
ClientSampleId :
DFTPP82

m/z	Abundance
37.00	884.0
38.10	2452.0
39.00	15317.0
40.00	2347.0
40.90	763.0
43.00	403.0
44.00	13675.0
45.00	723.0
49.10	1224.0
50.00	52936.0
51.00	221568.0
52.00	11162.0
53.00	610.0
55.00	1042.0
55.90	6472.0
57.00	15235.0
58.00	726.0
61.00	2941.0
62.00	2933.0
63.00	9094.0
63.90	1131.0
65.00	4355.0
66.00	442.0
66.90	456.0
68.00	3439.0
69.00	222528.0
70.00	1106.0
72.90	1863.0
74.00	21912.0
75.00	35472.0
76.00	11447.0
77.00	253568.0
78.00	16816.0
79.00	16432.0
80.00	13011.0
81.00	17808.0
82.00	4704.0
83.00	4143.0
83.80	575.0
85.00	3139.0
86.00	4908.0
87.00	2246.0
88.00	769.0
88.90	520.0
90.90	3883.0
91.90	4733.0
93.00	26352.0
93.90	2181.0
94.90	857.0
96.00	1407.0
96.80	607.0
98.00	21896.0
99.00	18624.0
100.00	1209.0
101.00	10733.0
102.00	1163.0
103.00	4368.0
103.90	6903.0
105.00	5490.0
106.00	2025.0
107.00	78936.0
108.00	12735.0
109.00	2161.0
110.00	151168.0
111.00	22520.0
112.00	2370.0
112.90	961.0
115.00	554.0
116.00	3900.0
117.00	58128.0
117.90	4140.0
118.80	542.0
119.00	545.0
119.90	955.0
122.00	5400.0
123.00	8323.0
123.90	3204.0
125.00	3420.0
126.10	1056.0
127.00	298432.0
128.00	24736.0
129.00	111536.0
130.00	9630.0
130.90	1656.0
132.00	1137.0
133.00	395.0
133.90	3069.0
135.00	9413.0
135.90	4435.0
137.00	5306.0
138.00	944.0

Instrument :
BNA_N
ClientSampleId :
DFTPP82

139.00	779.0
140.00	1276.0
141.00	14128.0
141.90	4737.0
143.00	3475.0
144.00	802.0
145.00	820.0
146.00	2209.0
147.00	7844.0
148.00	15511.0
148.90	3589.0
149.90	835.0
151.10	1916.0
151.70	1318.0
153.00	4371.0
154.00	4120.0
155.00	7799.0
156.00	13380.0
157.00	2526.0
157.90	3032.0
158.90	2035.0
159.90	5342.0
161.00	8156.0
162.00	1866.0
163.00	549.0
163.80	680.0
165.00	5538.0
166.00	4388.0
167.00	30120.0
168.00	17288.0
169.00	2751.0
170.10	1445.0
171.00	1297.0
171.90	2476.0
173.00	3586.0
174.00	6304.0
175.00	13169.0
176.10	3003.0
176.90	5745.0
177.90	2012.0
178.90	22808.0
179.90	16592.0
181.00	7086.0
182.10	1437.0
183.00	608.0
183.90	2186.0
185.00	11131.0
186.00	90784.0
187.00	25600.0
188.00	3001.0
189.00	5525.0
189.90	1048.0
191.00	2442.0
191.90	8174.0
193.00	8533.0
193.90	1657.0
195.10	808.0
196.00	19448.0
196.90	3139.0
197.90	712192.0
198.90	44816.0
199.90	4060.0
201.60	2775.0
203.00	4595.0
204.00	23864.0
205.00	41480.0
206.00	180288.0
207.00	24336.0
207.90	5956.0
209.00	2008.0
211.00	6725.0
212.00	1325.0
212.90	605.0
214.90	2040.0
216.00	3536.0
216.90	46256.0
218.00	6150.0
219.00	614.0
221.00	30344.0
223.00	10016.0
224.00	101664.0
225.00	24936.0
226.00	3001.0
227.00	40504.0
228.00	5698.0
229.00	9771.0
229.90	1043.0
231.00	3620.0
232.00	1035.0
233.00	576.0
234.00	2416.0
235.00	3000.0
235.90	2009.0
237.00	4156.0
237.90	512.0

Instrument :
BNA_N
ClientSampleId :
DFTPP82

239.00	1238.0
239.90	1391.0
240.90	1832.0
242.00	4960.0
243.00	5014.0
244.00	81168.0
245.00	11254.0
246.00	14353.0
246.90	3171.0
248.00	720.0
248.90	2802.0
250.00	355.0
250.90	866.0
252.00	924.0
252.90	1428.0
254.10	2811.0
255.00	415040.0
256.00	57072.0
257.00	4344.0
258.00	22064.0
258.90	3980.0
259.80	529.0
260.90	985.0
264.20	719.0
265.00	9210.0
265.90	1296.0
270.00	611.0
270.90	1023.0
271.90	1100.0
273.00	13047.0
274.00	31112.0
275.00	193408.0
276.00	25696.0
277.00	15039.0
278.00	2222.0
279.00	702.0
281.10	674.0
282.00	482.0
283.00	1371.0
284.00	1332.0
285.00	3507.0
285.90	753.0
287.90	380.0
289.10	435.0
289.80	550.0
290.90	361.0
291.90	522.0
293.00	3621.0
293.90	1108.0
294.90	1526.0
296.00	50960.0
297.00	7394.0
297.90	454.0
300.90	788.0
302.00	867.0
303.00	5730.0
304.00	1277.0
308.00	835.0
308.80	354.0
309.90	599.0
313.00	590.0
314.10	3422.0
314.90	4562.0
316.00	2877.0
317.10	464.0
321.00	1454.0
322.10	1047.0
323.00	15989.0
324.00	3000.0
324.90	453.0
325.90	318.0
327.00	3468.0
327.80	1873.0
331.90	1118.0
333.00	1632.0
334.00	10902.0
335.10	2982.0
336.10	420.0
338.80	366.0
340.00	371.0
341.00	2195.0
342.00	556.0
346.00	4098.0
347.00	747.0
350.90	372.0
352.00	4675.0
353.00	3485.0
354.00	6095.0
354.90	1251.0
359.00	406.0
365.00	24160.0
366.00	3151.0
369.90	592.0
370.90	1664.0
372.00	10315.0

Instrument :
BNA_N
ClientSampleId :
DFTPP82

373.00	2494.0
373.90	465.0
376.90	322.0
383.00	3068.0
384.00	694.0
385.00	481.0
389.90	1406.0
391.10	915.0
392.00	499.0
400.90	715.0
402.00	5541.0
403.00	6194.0
404.00	2433.0
405.10	520.0
420.90	5398.0
422.00	5131.0
423.00	37616.0
424.00	9995.0
425.00	1293.0
441.00	116072.0
442.00	742336.0
443.00	137664.0
444.00	11806.0
445.00	547.0

Instrument :
BNA_N
ClientSampleId :
DFTPP82

Scan 2468 (17.504 min): BN001902.D
DFTPP82

m/z	Abundance
40.00	1745.0
41.00	386.0
43.00	381.0
44.00	14476.0
73.00	359.0
207.00	2110.0
208.00	406.0
208.80	374.0
281.00	568.0

Instrument :
BNA_N
ClientSampleId :
DFTPP82