

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062524\
 Data File : BN032271.D
 Acq On : 26 Jun 2024 12:06
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 26 12:52:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN061024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 11:30:34 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	-0.02
2	1,4-Dioxane	0.490	0.438	10.6	93	0.00
3	n-Nitrosodimethylamine	0.465	0.374	19.6	85	0.00
4 S	2-Fluorophenol	1.137	0.970	14.7	88	-0.01
5 S	Phenol-d6	1.491	1.139	23.6	80	-0.01
6	bis(2-Chloroethyl)ether	1.284	0.920	28.3#	75	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	80	-0.02
8 S	Nitrobenzene-d5	0.332	0.281	15.4	78	-0.01
9	Naphthalene	1.077	0.972	9.7	83	-0.02
10	Hexachlorobutadiene	0.188	0.198	-5.3	96	-0.03
11 SURR	2-Methylnaphthalene-d10	0.637	0.527	17.3	76	-0.02
12	2-Methylnaphthalene	0.736	0.624	15.2	78	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	70	-0.02
14 S	2,4,6-Tribromophenol	0.192	0.150	21.9	68	-0.01
15 S	2-Fluorobiphenyl	1.517	1.420	6.4	76	-0.02
16	Acenaphthylene	1.844	1.497	18.8	69	-0.02
17	Acenaphthene	1.187	1.055	11.1	73	-0.02
18	Fluorene	1.586	1.327	16.3	68	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	63	0.00
20	4,6-Dinitro-2-methylphenol	0.069	0.028	59.4#	40#	0.01
21	4-Bromophenyl-phenylether	0.227	0.209	7.9	68	-0.01
22	Hexachlorobenzene	0.235	0.242	-3.0	75	-0.01
23	Atrazine	0.208	0.154	26.0#	55	-0.01
24	Pentachlorophenol	0.117	0.077	34.2#	59	0.00
25	Phenanthrene	1.110	1.032	7.0	68	-0.01
26	Anthracene	1.030	0.879	14.7	64	0.00
27 SURR	Fluoranthene-d10	1.089	0.960	11.8	65	0.00
28	Fluoranthene	1.346	1.228	8.8	67	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	63	0.00
30	Pyrene	1.589	1.418	10.8	67	0.00
31 S	Terphenyl-d14	0.978	0.803	17.9	62	0.00
32	Benzo(a)anthracene	1.507	1.314	12.8	66	0.00
33	Chrysene	1.424	1.348	5.3	71	0.00
34	Bis(2-ethylhexyl)phthalate	1.029	0.701	31.9#	51	0.00
35 I	Perylene-d12	1.000	1.000	0.0	68	0.00
36	Indeno(1,2,3-cd)pyrene	1.492	1.552	-4.0	88	0.00
37	Benzo(b)fluoranthene	1.458	1.361	6.7	75	0.00
38	Benzo(k)fluoranthene	1.378	1.352	1.9	79	0.00
39 C	Benzo(a)pyrene	1.258	1.184	5.9	77	0.00
40	Dibenzo(a,h)anthracene	1.176	1.237	-5.2	89	0.00
41	Benzo(g,h,i)perylene	1.263	1.317	-4.3	89	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4