

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062621\
 Data File : BN015094.D
 Acq On : 26 Jun 2021 00:54
 Operator : CG/JU
 Sample : M2680-14
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BGD46

Quant Time: Jun 26 01:26:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN061821MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 24 14:43:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.728	152	163448	20.000	ng/u1	0.00
20) Naphthalene-d8	10.498	136	673262	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.351	164	404433	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.098	188	803559	20.000	ng/u1	0.00
79) Chrysene-d12	21.292	240	772949	20.000	ng/u1	0.00
88) Perylene-d12	23.539	264	795231	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.287	96	26998	6.427	ng/uL	0.00
4) Pyridine-d5	3.675	84	283284	25.015	ng/u1	0.00
7) Phenol-d5	6.893	99	462502	32.602	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.063	67	275322	33.047	ng/u1	0.00
11) 2-Chlorophenol-d4	7.263	132	364295	33.866	ng/u1	0.00
15) 4-Methylphenol-d8	8.416	113	350419	32.005	ng/u1	0.00
21) Nitrobenzene-d5	8.869	128	178639	38.788	ng/u1	0.00
24) 2-Nitrophenol-d4	9.587	143	168919	41.175	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.116	165	330874	34.171	ng/u1	0.00
31) 4-Chloroaniline-d4	10.628	131	491831	32.775	ng/u1	0.00
46) Dimethylphthalate-d6	13.763	166	989441	35.628	ng/u1	0.00
49) Acenaphthylene-d8	14.045	160	1304103	36.173	ng/u1	0.00
54) 4-Nitrophenol-d4	14.539	143	121490	25.080	ng/u1	0.00
60) Fluorene-d10	15.345	176	865826	35.836	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.457	200	51950	15.653	ng/u1	0.00
73) Anthracene-d10	17.198	188	1288519	35.034	ng/u1	0.00
81) Pyrene-d10	19.498	212	1422849	35.120	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.398	264	1432747	36.496	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062621\
Data File : BN015094.D
Acq On : 26 Jun 2021 00:54
Operator : CG/JU
Sample : M2680-14
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BGD46

Quant Time: Jun 26 01:26:12 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN061821MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 24 14:43:13 2021
Response via : Initial Calibration

