

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062721\
 Data File : BN015140.D
 Acq On : 27 Jun 2021 06:10
 Operator : CG/JU
 Sample : M2680-14
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BGD46

Manual Integrations
 APPROVED

mohammad
 6/28/2021 1:28:13 PM

Quant Time: Jun 28 01:26:30 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 22 11:53:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.728	152	3812	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.496	136	13201	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.354	164	7233	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.099	188	14252m	0.400	ng/ul	-0.01
17) Chrysene-d12	21.293	240	12654	0.400	ng/ul	-0.02
23) Perylene-d12	23.535	264	12230	0.400	ng/ul	-0.04
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.285	96	29263	6.811	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.091	152	7866	0.380	ng/ul	-0.01
18) Fluoranthene-d10	19.132	212	15445	0.371	ng/ul	-0.02
Target Compounds						
					Qvalue	
19) Fluoranthene	19.160	202	2326	0.040	ng/ul#	86
20) Pyrene	19.522	202	1741	0.030	ng/ul#	72
21) Benzo(a)anthracene	21.276	228	1837	0.039	ng/ul#	77
22) Chrysene	21.328	228	1923	0.034	ng/ul#	88
24) Benzo(b)fluoranthene	22.865	252	2151	0.038	ng/ul#	8

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062721\
Data File : BN015140.D
Acq On : 27 Jun 2021 06:10
Operator : CG/JU
Sample : M2680-14
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BGD46

Manual Integrations
APPROVED

mohammad
6/28/2021 1:28:13 PM

Quant Time: Jun 28 01:26:30 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 22 11:53:52 2021
Response via : Initial Calibration

