

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062721\
 Data File : BN015161.D
 Acq On : 27 Jun 2021 20:26
 Operator : CG/JU
 Sample : PB137315BL
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK315

Manual Integrations
 APPROVED

mohammad
 6/28/2021 1:29:03 PM

Quant Time: Jun 28 01:29:28 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 22 11:53:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.728	152	4484	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.496	136	15676	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.349	164	8815	0.400	ng/ul	-0.02
13) Phenanthrene-d10	17.098	188	17369m	0.400	ng/ul	-0.01
17) Chrysene-d12	21.299	240	14064	0.400	ng/ul	-0.01
23) Perylene-d12	23.544	264	14887	0.400	ng/ul	-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.285	96	40669	8.047	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.091	152	10484	0.426	ng/ul	-0.01
18) Fluoranthene-d10	19.132	212	20910	0.452	ng/ul	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062721\
Data File : BN015161.D
Acq On : 27 Jun 2021 20:26
Operator : CG/JU
Sample : PB137315BL
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK315

Manual Integrations
APPROVED

mohammad
6/28/2021 1:29:03 PM

Quant Time: Jun 28 01:29:28 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 22 11:53:52 2021
Response via : Initial Calibration

