


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062721\
 Data File : BN015119.D
 Acq On : 26 Jun 2021 16:24
 Operator : CG/JU
 Sample : DFTPP153
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP153

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062121.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jun 22 11:53:52 2021



AutoFind: Scans 2458, 2459, 2460; Background Corrected with Scan 2452

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.7	95995	PASS
68	69	0.00	2	1.9	1935	PASS
69	198	0.00	100	36.4	100784	PASS
70	69	0.00	2	0.5	535	PASS
127	198	10	80	46.0	127208	PASS
197	198	0.00	2	0.3	955	PASS
198	198	100	100	100.0	276672	PASS
199	198	5	9	6.5	17995	PASS
275	198	10	60	23.3	64581	PASS
365	198	1	100	2.5	6830	PASS
441	198	0.01	100	13.2	36651	PASS
442	198	50	100	83.4	230749	PASS
443	442	15	24	18.4	42483	PASS

m/z	Abundance
39.90	901.0
41.00	337.0
43.00	558.0
44.00	1211.0
49.00	565.0
71.00	400.0
73.10	509.0
83.90	400.0
207.10	2067.0
208.00	342.0

Instrument :
BNA_N
ClientSampleId :
DFTPP153

m/z	Abundance
37.10	347.0
38.00	924.0
39.00	6130.0
40.00	1102.0
41.10	444.0
43.00	589.0
44.00	1264.0
49.00	901.0
50.00	20824.0
51.00	85704.0
52.00	4039.0
55.00	661.0
56.00	2475.0
57.00	5828.0
61.00	1326.0
62.00	1101.0
63.00	3445.0
63.90	601.0
65.00	2005.0
67.00	381.0
68.00	1597.0
69.00	87560.0
70.00	608.0
71.00	496.0
73.00	1056.0
74.00	8459.0
75.00	14047.0
76.00	4859.0
77.00	101112.0
78.00	7100.0
79.00	6545.0
80.00	4947.0
81.00	7395.0
82.00	1734.0
82.90	1759.0
83.90	523.0
84.90	1393.0
86.00	2184.0
87.00	999.0
87.90	413.0
91.00	1735.0
92.00	1963.0
93.00	10551.0
94.00	896.0
96.00	888.0
98.00	8695.0
99.00	7974.0
99.90	716.0
101.00	3874.0
103.00	1269.0
104.00	2085.0
105.00	2302.0
105.90	886.0
107.00	29600.0
108.00	5029.0
109.00	1187.0
110.00	57344.0
111.00	9213.0
112.00	1011.0
116.00	1804.0
117.00	22352.0
118.00	2040.0
118.90	422.0
119.90	376.0
122.00	1674.0
123.00	2930.0
124.00	1072.0
125.00	1014.0
126.20	736.0
127.00	106160.0
128.00	8234.0
129.00	37192.0
130.00	3102.0
131.00	619.0
131.90	469.0
133.00	314.0
134.00	1019.0
135.00	3006.0
136.00	1274.0
137.00	1927.0
139.90	503.0
141.00	5149.0
142.00	1584.0
143.00	1209.0
144.90	325.0
145.90	775.0
147.00	2296.0
147.90	5529.0
149.00	1422.0
150.00	451.0
151.10	491.0
153.00	1575.0

Instrument :
BNA_N
ClientSampleId :
DFTPP153

154.00	1248.0
155.00	2994.0
156.00	4157.0
157.10	774.0
157.80	881.0
159.00	552.0
160.00	1483.0
161.00	2190.0
162.00	724.0
164.90	1746.0
166.00	1434.0
167.00	10199.0
168.00	7028.0
169.00	1224.0
170.80	476.0
172.00	876.0
173.00	1191.0
174.00	2325.0
175.00	3968.0
176.10	1024.0
177.00	1641.0
177.90	619.0
179.00	7167.0
180.00	4526.0
181.00	2452.0
182.00	418.0
184.00	540.0
185.00	3042.0
186.00	25936.0
187.00	8099.0
188.00	859.0
188.90	1481.0
191.00	924.0
192.00	2136.0
193.00	2597.0
193.90	570.0
195.00	346.0
196.00	5694.0
197.10	1119.0
197.90	200768.0
199.00	13872.0
200.00	1004.0
201.40	840.0
203.00	1123.0
204.00	6718.0
205.00	11304.0
206.00	50624.0
207.00	8468.0
208.00	1961.0
209.00	728.0
210.10	859.0
211.00	1889.0
211.90	319.0
215.00	566.0
216.00	1059.0
217.00	11656.0
217.90	1799.0
221.00	7873.0
223.00	2851.0
224.00	25808.0
225.00	6702.0
226.10	953.0
227.00	10070.0
228.00	1723.0
229.00	2399.0
231.00	1035.0
234.00	648.0
234.90	787.0
236.00	554.0
237.00	1217.0
238.90	354.0
240.80	493.0
242.00	1286.0
243.00	1633.0
244.00	20384.0
245.00	2764.0
245.90	3927.0
246.90	742.0
249.00	649.0
253.00	461.0
254.20	1053.0
255.00	104424.0
256.00	16160.0
257.00	1433.0
258.00	5331.0
259.00	760.0
265.00	2089.0
270.90	367.0
273.00	2781.0
274.00	7555.0
275.00	39536.0
276.00	6038.0
277.00	3169.0
278.00	636.0
283.00	314.0
285.00	794.0

Instrument :
BNA_N
ClientSampleId :
DFTPP153

293.00	593.0
296.00	10982.0
296.90	1518.0
303.00	1316.0
304.00	406.0
314.00	645.0
314.90	1160.0
316.10	689.0
323.00	4004.0
324.00	654.0
326.90	639.0
327.90	399.0
333.00	363.0
334.00	2334.0
335.10	601.0
341.10	379.0
345.90	856.0
352.00	1013.0
353.00	576.0
354.00	1121.0
365.00	4547.0
365.90	677.0
372.00	1868.0
373.00	686.0
382.90	543.0
402.00	600.0
403.00	1040.0
403.90	357.0
421.00	1071.0
422.00	789.0
423.00	6589.0
424.00	2360.0
441.00	17888.0
442.00	90520.0
443.10	21552.0
444.00	2326.0

Instrument :
BNA_N
ClientSampleId :
DFTPP153

m/z	Abundance
37.00	455.0
38.00	1366.0
39.00	8575.0
39.90	1032.0
41.00	315.0
43.00	584.0
44.00	1312.0
49.10	1200.0
50.00	28360.0
51.00	112928.0
52.00	6220.0
53.00	328.0
54.90	585.0
56.00	3482.0
57.00	7836.0
57.90	327.0
61.00	1633.0
62.00	1813.0
63.00	4863.0
64.00	594.0
65.00	2335.0
68.00	2232.0
69.00	118416.0
70.00	635.0
71.00	387.0
73.00	1230.0
74.00	11853.0
75.00	17832.0
76.00	6419.0
77.00	141696.0
78.00	9175.0
79.00	9042.0
80.00	7132.0
81.00	9891.0
82.00	2544.0
83.00	2370.0
83.90	721.0
85.00	2637.0
86.00	3129.0
87.00	1196.0
87.90	597.0
91.00	2279.0
92.00	2474.0
93.00	14925.0
94.00	931.0
95.00	418.0
96.10	829.0
98.00	12163.0
99.00	9819.0
100.00	929.0
101.00	5650.0
103.00	1951.0
103.90	3536.0
105.00	3084.0
106.00	1066.0
107.00	43256.0
108.00	6742.0
109.00	1277.0
110.00	81736.0
111.00	11660.0
112.00	1344.0
112.90	613.0
116.00	2485.0
117.00	32456.0
118.00	2522.0
118.90	379.0
120.00	670.0
121.90	3061.0
123.00	4072.0
124.00	1696.0
125.00	1754.0
126.20	1071.0
127.00	152768.0
128.00	11874.0
129.00	56208.0
130.00	4636.0
131.00	887.0
132.00	600.0
133.90	1385.0
135.00	4580.0
135.90	1826.0
137.00	2852.0
137.90	769.0
139.00	301.0
139.90	605.0
141.00	6739.0
141.90	2394.0
142.90	1550.0
144.10	391.0
144.90	471.0
145.90	1357.0
147.00	4121.0

Instrument :
BNA_N
ClientSampleId :
DFTPP153

148.00	8460.0
149.00	1754.0
150.00	408.0
151.00	739.0
151.90	619.0
153.00	2382.0
154.00	1886.0
155.00	3695.0
156.00	5910.0
157.10	1344.0
158.00	1743.0
159.00	1100.0
160.00	2257.0
160.90	3511.0
162.00	891.0
163.00	433.0
163.90	339.0
165.00	2473.0
166.00	2157.0
167.00	14826.0
168.00	8883.0
169.00	1467.0
169.90	577.0
171.00	704.0
171.90	1229.0
173.00	2165.0
174.00	2911.0
175.00	5806.0
176.00	1588.0
177.00	2604.0
178.00	905.0
179.00	11453.0
180.00	8174.0
181.00	3828.0
182.00	529.0
182.90	437.0
184.00	742.0
185.00	6321.0
186.00	44272.0
187.00	12807.0
188.00	1349.0
188.90	2325.0
189.80	361.0
191.00	1386.0
192.00	3637.0
193.00	4125.0
194.00	863.0
195.00	455.0
196.00	8787.0
198.00	341696.0
199.00	21920.0
200.00	1814.0
201.50	1417.0
202.90	2246.0
204.00	11117.0
205.00	17704.0
206.00	81640.0
207.00	11745.0
207.90	2965.0
209.00	1016.0
210.10	1241.0
211.00	3122.0
215.00	839.0
216.00	1689.0
217.00	18824.0
218.00	2677.0
218.90	322.0
221.00	12926.0
221.70	3207.0
223.00	4577.0
224.00	41456.0
225.00	11065.0
226.00	1466.0
227.00	15634.0
228.00	2455.0
229.00	3613.0
230.00	693.0
231.00	1654.0
234.00	1117.0
234.90	1256.0
235.90	947.0
237.00	1796.0
238.90	703.0
239.90	530.0
241.00	872.0
242.00	2461.0
243.10	2588.0
244.00	33440.0
245.00	4712.0
246.00	6443.0
247.00	904.0
247.90	348.0
249.00	1276.0
252.00	504.0
252.90	696.0
254.10	1464.0

Instrument :
BNA_N
ClientSampleId :
DFTPP153

255.00	177920.0
256.00	26416.0
257.00	1804.0
258.00	9318.0
259.00	1320.0
264.10	316.0
265.00	3907.0
266.00	551.0
271.00	340.0
271.90	408.0
272.90	4935.0
274.00	14373.0
275.00	80336.0
276.00	9628.0
277.00	6478.0
278.00	1079.0
283.00	780.0
284.00	559.0
285.10	1162.0
293.00	1585.0
293.90	342.0
295.00	515.0
296.00	20824.0
297.00	2779.0
301.90	402.0
303.00	2488.0
304.00	869.0
314.00	813.0
314.90	1940.0
316.10	1323.0
321.00	749.0
321.90	306.0
323.10	6308.0
324.00	1205.0
327.00	1170.0
327.90	793.0
332.00	375.0
333.00	697.0
334.00	3881.0
335.00	1123.0
341.00	785.0
345.90	1499.0
346.90	428.0
352.00	1977.0
353.00	1202.0
354.00	2487.0
365.00	8061.0
366.00	1113.0
371.00	610.0
372.00	3498.0
373.00	757.0
382.90	765.0
383.90	412.0
390.00	373.0
391.10	406.0
402.00	1587.0
402.90	1691.0
403.90	791.0
405.00	315.0
420.90	1672.0
422.00	2091.0
423.00	13443.0
424.00	3302.0
425.00	352.0
441.00	41416.0
442.00	284736.0
443.00	53360.0
444.00	4775.0
445.00	444.0

Instrument :
BNA_N
ClientSampleId :
DFTPP153

m/z	Abundance
37.00	380.0
38.00	1165.0
39.10	6713.0
40.00	953.0
41.10	508.0
43.00	517.0
44.00	1229.0
49.00	1027.0
50.00	22072.0
51.00	89352.0
52.00	4541.0
54.90	599.0
56.00	2387.0
57.00	6430.0
60.90	1547.0
62.00	1368.0
63.00	3503.0
64.00	539.0
65.00	1894.0
68.10	1976.0
69.00	96376.0
70.00	361.0
71.00	469.0
73.00	974.0
74.00	9290.0
75.00	14758.0
76.00	4807.0
77.00	107504.0
78.00	7895.0
79.00	6719.0
80.00	5221.0
81.00	7687.0
82.00	2046.0
83.00	2125.0
83.90	691.0
84.90	2211.0
86.00	2622.0
87.00	1014.0
87.90	507.0
91.00	1918.0
92.00	2112.0
93.00	11876.0
94.00	893.0
96.00	681.0
98.00	9424.0
99.00	8207.0
100.00	744.0
101.00	4331.0
103.00	1690.0
104.00	2658.0
105.00	2683.0
106.00	866.0
107.00	34000.0
108.00	5712.0
109.00	1023.0
110.00	66480.0
111.00	9518.0
112.00	1166.0
113.00	444.0
116.00	1600.0
117.00	26328.0
118.00	1854.0
119.90	375.0
122.00	2261.0
123.00	3108.0
123.90	1643.0
124.90	1372.0
126.10	883.0
127.00	122696.0
128.00	9580.0
129.00	45088.0
129.90	3828.0
131.00	883.0
131.90	489.0
134.00	1099.0
135.00	4209.0
136.00	1415.0
137.00	2122.0
137.90	436.0
139.00	311.0
139.90	652.0
141.00	6206.0
142.00	1876.0
143.00	1257.0
144.00	305.0
145.90	1074.0
147.00	3518.0
148.00	7480.0
148.90	1712.0
150.00	373.0
150.90	880.0
151.90	584.0

Instrument :
BNA_N
ClientSampleId :
DFTPP153

152.90	1949.0
153.90	1524.0
155.00	3773.0
156.00	5232.0
157.00	1132.0
158.00	1157.0
158.90	984.0
160.00	1977.0
160.90	3462.0
162.00	989.0
164.00	446.0
165.00	2500.0
165.90	2250.0
167.00	13168.0
168.00	7562.0
169.00	1369.0
170.00	416.0
170.90	575.0
171.90	1028.0
173.00	1694.0
174.00	2866.0
175.00	5122.0
176.00	1544.0
177.00	2352.0
178.00	800.0
179.00	9593.0
180.00	6427.0
181.00	3316.0
182.00	496.0
183.90	819.0
185.10	4940.0
186.00	37936.0
187.00	10478.0
188.00	1106.0
188.90	2539.0
189.80	399.0
191.10	1368.0
192.00	2736.0
193.00	3789.0
193.90	716.0
194.80	506.0
196.00	7827.0
196.80	1659.0
197.10	1745.0
198.00	287552.0
199.00	18192.0
200.00	1634.0
201.50	1466.0
203.00	1935.0
204.00	10821.0
205.00	17400.0
206.00	70744.0
207.00	10926.0
208.00	2672.0
208.90	984.0
210.20	1325.0
211.00	2447.0
211.90	499.0
215.00	973.0
216.00	1780.0
217.00	18864.0
217.90	2237.0
221.00	12154.0
221.70	2510.0
223.00	4230.0
224.00	42176.0
225.00	10091.0
226.00	1239.0
227.00	15477.0
228.00	2497.0
228.90	3536.0
229.90	392.0
231.10	1305.0
232.90	377.0
233.90	949.0
235.00	1368.0
236.10	852.0
237.00	1359.0
238.80	536.0
239.90	546.0
241.10	985.0
242.00	2289.0
243.00	2136.0
244.00	33160.0
245.00	4335.0
246.00	5787.0
247.00	1242.0
249.00	1276.0
250.90	305.0
253.00	672.0
254.10	1254.0
255.00	169856.0
256.00	24152.0
257.00	1886.0
258.00	9025.0
258.90	1364.0

Instrument :
BNA_N
ClientSampleId :
DFTPP153

259.90	390.0
264.90	3291.0
266.00	752.0
270.90	319.0
271.90	460.0
273.00	4686.0
274.00	11747.0
275.00	73872.0
276.00	8987.0
277.00	6329.0
278.00	1014.0
283.10	801.0
283.90	467.0
285.10	1170.0
292.10	371.0
293.00	1359.0
294.00	477.0
296.00	19760.0
296.90	2521.0
301.90	452.0
303.00	2770.0
304.00	525.0
308.00	369.0
310.00	361.0
313.90	867.0
314.90	2175.0
316.00	1129.0
317.00	311.0
321.10	545.0
323.00	7629.0
324.00	1101.0
326.90	1000.0
328.00	548.0
332.10	490.0
332.90	578.0
334.00	3526.0
335.00	1224.0
341.00	733.0
346.00	1698.0
346.90	450.0
352.00	1848.0
353.00	1803.0
354.00	2380.0
355.00	481.0
364.90	7883.0
366.00	1449.0
371.00	573.0
372.10	4044.0
373.00	672.0
383.00	927.0
384.00	508.0
390.10	549.0
391.00	340.0
391.90	409.0
401.90	1405.0
403.00	2319.0
404.00	773.0
420.90	2123.0
421.90	2255.0
423.10	14625.0
424.00	3338.0
425.00	393.0
440.10	371.0
441.00	50648.0
442.00	316992.0
443.00	52536.0
444.00	5716.0
444.90	376.0

Instrument :
BNA_N
ClientSampleId :
DFTPP153