

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062722\
 Data File : BN020507.D
 Acq On : 27 Jun 2022 20:03
 Operator : CG/JU
 Sample : N3368-07
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-102

Quant Time: Jun 28 01:29:03 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 02:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.746	152	2810	0.400	ng	0.00
7) Naphthalene-d8	10.530	136	8486	0.400	ng	0.01
13) Acenaphthene-d10	14.367	164	4979	0.400	ng	0.00
19) Phenanthrene-d10	17.113	188	9802	0.400	ng	# 0.00
29) Chrysene-d12	21.306	240	7952	0.400	ng	# 0.00
35) Perylene-d12	23.583	264	6246	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.355	112	1075	0.138	ng	0.00
5) Phenol-d6	6.966	99	592	0.064	ng	0.02
8) Nitrobenzene-d5	8.907	82	1904	0.277	ng	0.01
11) 2-Methylnaphthalene-d10	12.121	152	5094	0.350	ng	0.00
14) 2,4,6-Tribromophenol	15.872	330	508	0.271	ng	0.01
15) 2-Fluorobiphenyl	12.998	172	6725	0.327	ng	0.00
27) Fluoranthene-d10	19.147	212	11191	0.384	ng	0.00
31) Terphenyl-d14	19.746	244	8132	0.428	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.225	149	890	0.056	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062722\
Data File : BN020507.D
Acq On : 27 Jun 2022 20:03
Operator : CG/JU
Sample : N3368-07
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-102

Quant Time: Jun 28 01:29:03 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 02:03:21 2022
Response via : Initial Calibration

