

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062722\
 Data File : BN020525.D
 Acq On : 28 Jun 2022 08:43
 Operator : CG/JU
 Sample : N3386-04
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW181S-20220614

Quant Time: Jun 28 09:22:25 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 02:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.739	152	2985	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	9527	0.400	ng	# 0.00
13) Acenaphthene-d10	14.367	164	5541	0.400	ng	0.00
19) Phenanthrene-d10	17.113	188	11529	0.400	ng	0.00
29) Chrysene-d12	21.306	240	10216	0.400	ng	# 0.00
35) Perylene-d12	23.583	264	8168	0.400	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	921	0.111	ng	0.00
5) Phenol-d6	6.959	99	530	0.054	ng	0.01
8) Nitrobenzene-d5	8.897	82	1598	0.207	ng	0.00
11) 2-Methylnaphthalene-d10	12.117	152	4258	0.261	ng	0.00
14) 2,4,6-Tribromophenol	15.861	330	511	0.245	ng	0.00
15) 2-Fluorobiphenyl	12.999	172	5662	0.248	ng	0.00
27) Fluoranthene-d10	19.143	212	11846	0.346	ng	0.00
31) Terphenyl-d14	19.742	244	8165	0.334	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.232	88	537	0.137	ng	# 87
34) Bis(2-ethylhexyl)phtha...	21.225	149	1571	0.076	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062722\
Data File : BN020525.D
Acq On : 28 Jun 2022 08:43
Operator : CG/JU
Sample : N3386-04
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW181S-20220614

Quant Time: Jun 28 09:22:25 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 02:03:21 2022
Response via : Initial Calibration

