

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062722\
 Data File : BN020528.D
 Acq On : 28 Jun 2022 10:30
 Operator : CG/JU
 Sample : N3354-15RE
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-59RE

Quant Time: Jun 28 23:49:54 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 02:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.739	152	3319	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	11060	0.400	ng	# 0.00
13) Acenaphthene-d10	14.367	164	14404	0.400	ng	# 0.00
19) Phenanthrene-d10	17.102	188	13451	0.400	ng	#-0.01
29) Chrysene-d12	21.306	240	11222	0.400	ng	# 0.00
35) Perylene-d12	23.586	264	9115	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	1464	0.159	ng	0.00
5) Phenol-d6	6.944	99	1199	0.109	ng	0.00
8) Nitrobenzene-d5	8.886	82	2973	0.332	ng	-0.01
11) 2-Methylnaphthalene-d10	12.110	152	6719	0.354	ng	-0.01
14) 2,4,6-Tribromophenol	15.861	330	1166	0.215	ng	0.00
15) 2-Fluorobiphenyl	12.988	172	9328	0.157	ng	-0.01
27) Fluoranthene-d10	19.143	212	15190	0.380	ng	0.00
31) Terphenyl-d14	19.746	244	12154	0.453	ng	0.00
Target Compounds						
30) Pyrene	19.535	202	121881	2.565	ng	99
34) Bis(2-ethylhexyl)phtha...	21.225	149	2856	0.126	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062722\
Data File : BN020528.D
Acq On : 28 Jun 2022 10:30
Operator : CG/JU
Sample : N3354-15RE
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-59RE

Quant Time: Jun 28 23:49:54 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 02:03:21 2022
Response via : Initial Calibration

