




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062722\
Data File : BN020500.D
Acq On : 27 Jun 2022 14:49
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 28 08:02:05 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 28 08:02:00 2022
Response via : Initial Calibration



TIC: BN020500.D\data.ms

(70) Pentachlorophenol (C)




16.775min (0.000) 20457.99 ng

response 192659

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.92
264.00	61.60	57.47
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062722\
 Data File : BN020500.D
 Acq On : 27 Jun 2022 14:49
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 28 08:02:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 28 08:02:00 2022
 Response via : Initial Calibration



TIC: BN020500.D\data.ms

(77) Benzidine

19.363min (0.000) 0.00 ng

response 1345564

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.03
183.00	13.20	11.96
0.00	0.00	0.00