

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062722\
 Data File : BN020519.D
 Acq On : 28 Jun 2022 05:07
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 28 06:56:32 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 02:03:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	-0.09
2	1,4-Dioxane	0.524	0.507	3.2	116	-0.03
3	n-Nitrosodimethylamine	0.478	0.422	11.7	109	-0.04
4 S	2-Fluorophenol	1.110	1.110	0.0	118	-0.07
5 S	Phenol-d6	1.327	1.383	-4.2	126	-0.09
6	bis(2-Chloroethyl)ether	1.142	1.157	-1.3	123	-0.09
7 I	Naphthalene-d8	1.000	1.000	0.0	122	-0.11
8 S	Nitrobenzene-d5	0.324	0.312	3.7	129	-0.11
9	Naphthalene	1.195	1.182	1.1	126	-0.11
10	Hexachlorobutadiene	0.218	0.213	2.3	123	-0.11
11 SURR	2-Methylnaphthalene-d10	0.686	0.669	2.5	126	-0.11
12	2-Methylnaphthalene	0.794	0.796	-0.3	128	-0.11
13 I	Acenaphthene-d10	1.000	1.000	0.0	130	-0.11
14 S	2,4,6-Tribromophenol	0.151	0.139	7.9	139	-0.11
15 S	2-Fluorobiphenyl	1.651	1.585	4.0	127	-0.12
16	Acenaphthylene	1.923	1.810	5.9	134	-0.11
17	Acenaphthene	1.308	1.268	3.1	131	-0.11
18	Fluorene	1.707	1.640	3.9	132	-0.12
19 I	Phenanthrene-d10	1.000	1.000	0.0	128	-0.11
20	4,6-Dinitro-2-methylphenol	0.051	0.025	51.0#	83	-0.12
21	4-Bromophenyl-phenylether	0.233	0.223	4.3	132	-0.12
22	Hexachlorobenzene	0.257	0.250	2.7	133	-0.11
23	Atrazine	0.183	0.157	14.2	121	-0.11
24	Pentachlorophenol	0.061	0.061	0.0	173#	-0.12
25	Phenanthrene	1.349	1.309	3.0	133	-0.11
26	Anthracene	1.163	1.085	6.7	131	-0.11
27 SURR	Fluoranthene-d10	1.188	1.118	5.9	130	-0.13
28	Fluoranthene	1.471	1.403	4.6	132	-0.13
29 I	Chrysene-d12	1.000	1.000	0.0	125	-0.13
30	Pyrene	1.694	1.715	-1.2	129	-0.13
31 S	Terphenyl-d14	0.956	0.975	-2.0	130	-0.13
32	Benzo(a)anthracene	1.429	1.356	5.1	126	-0.13
33	Chrysene	1.582	1.501	5.1	123	-0.13
34	Bis(2-ethylhexyl)phthalate	0.805	0.693	13.9	121	-0.13
35 I	Perylene-d12	1.000	1.000	0.0	101	-0.21
36	Indeno(1,2,3-cd)pyrene	1.782	1.492	16.3	89	-0.31
37	Benzo(b)fluoranthene	1.541	1.623	-5.3	112	-0.18
38	Benzo(k)fluoranthene	1.641	1.626	0.9	107	-0.18
39 C	Benzo(a)pyrene	1.337	1.324	1.0	105	-0.20
40	Dibenzo(a,h)anthracene	1.429	1.183	17.2	89	-0.32
41	Benzo(g,h,i)perylene	1.561	1.271	18.6	85	-0.34

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4