

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062723\
 Data File : BN026038.D
 Acq On : 27 Jun 2023 15:44
 Operator : MA/JU
 Sample : PB153778BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 PB153778BL

Quant Time: Jun 27 18:01:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 27 15:13:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.976	152	5110	0.400	ng	0.00
7) Naphthalene-d8	10.793	136	14555	0.400	ng	0.01
13) Acenaphthene-d10	14.619	164	9249	0.400	ng	0.01
19) Phenanthrene-d10	17.368	188	18010	0.400	ng	0.01
29) Chrysene-d12	21.551	240	9587	0.400	ng	# 0.00
35) Perylene-d12	24.025	264	7705	0.400	ng	# 0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.528	112	4512	0.364	ng	0.00
5) Phenol-d6	7.153	99	4247	0.269	ng	0.01
8) Nitrobenzene-d5	9.170	82	3494	0.281	ng	0.02
11) 2-Methylnaphthalene-d10	12.389	152	7169	0.337	ng	0.02
14) 2,4,6-Tribromophenol	16.115	330	930	0.276	ng	0.01
15) 2-Fluorobiphenyl	13.251	172	9733	0.273	ng	0.01
27) Fluoranthene-d10	19.393	212	13563	0.355	ng	0.00
31) Terphenyl-d14	19.983	244	9273	0.378	ng	0.00
Target Compounds						
3) n-Nitrosodimethylamine	3.708	42	45	0.008	ng	# 1
6) bis(2-Chloroethyl)ether	7.406	93	2	0.000	ng	# 1
9) Naphthalene	10.835	128	26	0.001	ng	# 1
10) Hexachlorobutadiene	11.156	225	7	0.001	ng	# 56
12) 2-Methylnaphthalene	12.434	142	4	0.000	ng	# 22
16) Acenaphthylene	14.352	152	38	0.001	ng	# 74
17) Acenaphthene	14.737	154	12	0.000	ng	# 17
18) Fluorene	15.680	166	20	0.001	ng	# 33
20) 4,6-Dinitro-2-methylph...	15.780	198	7	Below Cal		# 1
21) 4-Bromophenyl-phenylether	16.549	248	2	0.000	ng	# 44
22) Hexachlorobenzene	16.636	284	22	0.002	ng	# 1
23) Atrazine	16.835	200	6	0.001	ng	# 1
24) Pentachlorophenol	17.021	266	22	0.005	ng	# 26
25) Phenanthrene	17.406	178	59	0.001	ng	# 60
26) Anthracene	17.517	178	39	0.001	ng	# 1
28) Fluoranthene	19.421	202	116	0.002	ng	# 82
30) Pyrene	19.793	202	97	0.002	ng	# 92
32) Benzo(a)anthracene	21.542	228	209	0.006	ng	# 22
33) Chrysene	21.596	228	228	0.006	ng	# 31
34) Bis(2-ethylhexyl)phtha...	21.444	149	655	0.021	ng	# 93
36) Indeno(1,2,3-cd)pyrene	26.624	276	328	0.010	ng	# 70
37) Benzo(b)fluoranthene	23.279	252	263	0.009	ng	# 1
38) Benzo(k)fluoranthene	23.315	252	23	0.001	ng	# 1
39) Benzo(a)pyrene	23.911	252	104	0.004	ng	# 1
40) Dibenzo(a,h)anthracene	26.656	278	296	0.012	ng	# 1
41) Benzo(g,h,i)perylene	27.402	276	456	0.016	ng	# 19

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Jun 27 18:01:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 27 15:13:37 2023
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.976 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44

Instrument :
BNA_N
ClientSampleId :
PB153778BL







Tgt Ion:152 Resp: 5110
Ion Ratio Lower Upper
152 100
150 154.6 124.7 187.1
115 62.0 48.2 72.4



#3
n-Nitrosodimethylamine
Concen: 0.008 ng
RT: 3.708 min Scan# 73
Delta R.T. -0.014 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44

Tgt Ion: 42 Resp: 45
Ion Ratio Lower Upper
42 100
74 0.0 93.8 140.8#
44 320.0 3.0 4.4#









#4
2-Fluorophenol
Concen: 0.364 ng
RT: 5.528 min Scan# 31
Delta R.T. 0.000 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44




Instrument :
BNA_N
ClientSampleId :
PB153778BL


Tgt Ion:112 Resp: 4512
Ion Ratio Lower Upper
112 100
64 58.4 45.8 68.6
63 34.1 27.4 41.2



#5
Phenol-d6
Concen: 0.269 ng
RT: 7.153 min Scan# 550
Delta R.T. 0.014 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44

Tgt Ion: 99 Resp: 4247
Ion Ratio Lower Upper
99 100
42 21.1 18.4 27.6
71 37.3 30.6 46.0









#6
bis(2-Chloroethyl)ether
Concen: 0.000 ng
RT: 7.406 min Scan# 584
Delta R.T. 0.007 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44




Instrument :
BNA_N
ClientSampleId :
PB153778BL


Tgt Ion: 93 Resp: 2
Ion Ratio Lower Upper
93 100
63 0.0 58.8 88.2#
95 500.0 29.2 43.8#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.793 min Scan# 943
Delta R.T. 0.011 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44

Tgt Ion: 136 Resp: 14555
Ion Ratio Lower Upper
136 100
137 12.2 9.0 13.4
54 9.8 7.8 11.6
68 7.4 5.8 8.8









#8
Nitrobenzene-d5
Concen: 0.281 ng
RT: 9.170 min Scan# 791
Delta R.T. 0.021 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44




Instrument :
BNA_N
ClientSampleId :
PB153778BL


Tgt Ion: 82 Resp: 3494
Ion Ratio Lower Upper
82 100
128 43.0 33.4 50.2
54 58.4 41.7 62.5



#9
Naphthalene
Concen: 0.001 ng
RT: 10.835 min Scan# 947
Delta R.T. 0.000 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44

Tgt Ion: 128 Resp: 26
Ion Ratio Lower Upper
128 100
129 83.8 9.1 13.7#
127 72.5 10.9 16.3#





#10

Hexachlorobutadiene

Concen: 0.001 ng

RT: 11.156 min Scan# 91

Delta R.T. 0.043 min

Lab File: BN026038.D

Acq: 27 Jun 2023 15:44

Instrument :

BNA_N

ClientSampleId :

PB153778BL





Tgt Ion:225 Resp: 7

Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 28.6 50.3 75.5#



#11

2-Methylnaphthalene-d10

Concen: 0.337 ng

RT: 12.389 min Scan# 1243

Delta R.T. 0.015 min

Lab File: BN026038.D




Acq: 27 Jun 2023 15:44


Tgt Ion:152 Resp: 7169

Ion Ratio Lower Upper

152 100

151 19.6 16.2 24.2









#12
2-Methylnaphthalene
Concen: 0.000 ng
RT: 12.434 min Scan# 11
Delta R.T. -0.011 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44




Instrument :
BNA_N
ClientSampleId :
PB153778BL


Tgt Ion:142 Resp: 4
Ion Ratio Lower Upper
142 100
141 101.6 71.7 107.5
115 179.0 33.4 50.0#



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.619 min Scan# 1557
Delta R.T. 0.011 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44


Tgt Ion:164 Resp: 9249
Ion Ratio Lower Upper
164 100
162 105.1 85.3 127.9
160 50.2 40.0 60.0








#14
2,4,6-Tribromophenol
Concen: 0.276 ng
RT: 16.115 min Scan# 1689
Delta R.T. 0.012 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44

Instrument :
BNA_N
ClientSampleId :
PB153778BL







Tgt Ion: 330 Resp: 930
Ion Ratio Lower Upper
330 100
332 92.4 76.0 114.0
141 54.5 42.8 64.2



#15
2-Fluorobiphenyl
Concen: 0.273 ng
RT: 13.251 min Scan# 1429
Delta R.T. 0.011 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44

Tgt Ion: 172 Resp: 9733
Ion Ratio Lower Upper
172 100
171 35.3 27.8 41.8
170 24.0 19.3 28.9









#16
Acenaphthylene
Concen: 0.001 ng
RT: 14.352 min Scan# 1532
Delta R.T. 0.021 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44




Instrument :
BNA_N
ClientSampleId :
PB153778BL


Tgt Ion:152 Resp: 38
Ion Ratio Lower Upper
152 100
151 0.0 15.9 23.9#
153 13.2 10.6 15.8



#17
Acenaphthene
Concen: 0.000 ng
RT: 14.737 min Scan# 1568
Delta R.T. 0.064 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44

Tgt Ion:154 Resp: 12
Ion Ratio Lower Upper
154 100
153 0.0 94.4 141.6#
152 33.3 46.1 69.1#





#18

Fluorene

Concen: 0.001 ng

RT: 15.680 min Scan# 1

Delta R.T. 0.025 min

Lab File: BN026038.D

Acq: 27 Jun 2023 15:44

Instrument :

BNA_N

ClientSampleId :

PB153778BL





Tgt Ion:166 Resp: 20

Ion Ratio Lower Upper

166 100

165 160.0 79.7 119.5#

167 60.0 10.6 16.0#



#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.368 min Scan# 1790

Delta R.T. 0.012 min

Lab File: BN026038.D

Acq: 27 Jun 2023 15:44




Tgt Ion:188 Resp: 18010


Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.3 8.6 13.0









#20
4,6-Dinitro-2-methylphenol
Concen: Below Cal
RT: 15.780 min Scan# 1662
Delta R.T. 0.025 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44




Instrument :
BNA_N
ClientSampleId :
PB153778BL


Tgt Ion:198 Resp: 7
Ion Ratio Lower Upper
198 100
51 368.0 85.8 128.8#
105 186.0 48.2 72.2#



#21
4-Bromophenyl-phenylether
Concen: 0.000 ng
RT: 16.549 min Scan# 1724
Delta R.T. 0.000 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44

Tgt Ion:248 Resp: 2
Ion Ratio Lower Upper
248 100
250 122.9 79.9 119.9#
141 135.4 41.6 62.4#





#22

Hexachlorobenzene

Concen: 0.002 ng

RT: 16.636 min Scan# 1731

Delta R.T. -0.025 min

Lab File: BN026038.D

Acq: 27 Jun 2023 15:44

Instrument :

BNA_N

ClientSampleId :

PB153778BL





Tgt Ion:284 Resp: 22

Ion Ratio Lower Upper

284 100

142 0.0 38.7 58.1#

249 190.9 28.5 42.7#



#23

Atrazine

Concen: 0.001 ng

RT: 16.835 min Scan# 1747

Delta R.T. 0.025 min

Lab File: BN026038.D

Acq: 27 Jun 2023 15:44




Tgt Ion:200 Resp: 6


Ion Ratio Lower Upper

200 100

173 105.9 22.9 34.3#

215 102.0 35.5 53.3#





#24

Pentachlorophenol

Concen: 0.005 ng

RT: 17.021 min Scan# 1762

Delta R.T. 0.000 min

Lab File: BN026038.D

Acq: 27 Jun 2023 15:44

Instrument :

BNA_N

ClientSampleId :

PB153778BL





Tgt Ion:266 Resp: 22

Ion Ratio Lower Upper

266 100

264 113.6 49.3 73.9#

268 0.0 48.3 72.5#



#25

Phenanthrene

Concen: 0.001 ng

RT: 17.406 min Scan# 1793

Delta R.T. 0.012 min

Lab File: BN026038.D

Acq: 27 Jun 2023 15:44




Tgt Ion:178 Resp: 59


Ion Ratio Lower Upper

178 100

176 0.0 15.6 23.4#

179 0.0 12.3 18.5#





#26

Anthracene

Concen: 0.001 ng

RT: 17.517 min Scan# 1802

Delta R.T. 0.025 min

Lab File: BN026038.D

Acq: 27 Jun 2023 15:44

Instrument :

BNA_N

ClientSampleId :

PB153778BL





Tgt Ion:178 Resp: 39

Ion Ratio Lower Upper

178 100

176 0.0 15.0 22.4#

179 123.1 12.4 18.6#



#27

Fluoranthene-d10

Concen: 0.355 ng

RT: 19.393 min Scan# 2072

Delta R.T. 0.009 min

Lab File: BN026038.D

Acq: 27 Jun 2023 15:44




Tgt Ion:212 Resp: 13563


Ion Ratio Lower Upper

212 100

106 14.2 11.6 17.4

104 8.0 7.2 10.8









#28
Fluoranthene
Concen: 0.002 ng
RT: 19.421 min Scan# 2078
Delta R.T. 0.005 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44




Instrument :
BNA_N
ClientSampleId :
PB153778BL


Tgt Ion:202 Resp: 116
Ion Ratio Lower Upper
202 100
101 16.4 8.3 12.5#
203 8.6 13.5 20.3#



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.551 min Scan# 2480
Delta R.T. 0.009 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44

Tgt Ion:240 Resp: 9587
Ion Ratio Lower Upper
240 100
120 23.9 14.6 22.0#
236 31.3 24.2 36.2





#30

Pyrene

Concen: 0.002 ng

RT: 19.793 min Scan# 2155

Delta R.T. 0.014 min

Lab File: BN026038.D

Acq: 27 Jun 2023 15:44

Instrument :

BNA_N

ClientSampleId :

PB153778BL





Tgt Ion:202 Resp: 97

Ion Ratio Lower Upper

202 100

200 25.8 17.0 25.4#

203 20.6 14.2 21.4



#31

Terphenyl-d14

Concen: 0.378 ng

RT: 19.983 min Scan# 2199

Delta R.T. 0.005 min

Lab File: BN026038.D

Acq: 27 Jun 2023 15:44




Tgt Ion:244 Resp: 9273


Ion Ratio Lower Upper

244 100

212 11.0 8.7 13.1

122 16.7 13.7 20.5









#32
Benzo(a)anthracene
Concen: 0.006 ng
RT: 21.542 min Scan# 2479
Delta R.T. 0.018 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44




Instrument :
BNA_N
ClientSampleId :
PB153778BL


Tgt Ion:228 Resp: 209
Ion Ratio Lower Upper
228 100
226 69.4 23.0 34.6#
229 57.1 15.9 23.9#



#33
Chrysene
Concen: 0.006 ng
RT: 21.596 min Scan# 2485
Delta R.T. 0.018 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44

Tgt Ion:228 Resp: 228
Ion Ratio Lower Upper
228 100
226 66.7 24.8 37.2#
229 53.9 15.6 23.4#









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.021 ng
 RT: 21.444 min Scan# 2468
 Delta R.T. 0.009 min
 Lab File: BN026038.D
 Acq: 27 Jun 2023 15:44




Instrument :
 BNA_N
 ClientSampleId :
 PB153778BL


Tgt Ion:149 Resp: 655
 Ion Ratio Lower Upper
 149 100
 167 28.2 20.2 30.4
 279 10.1 2.4 3.6#



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 24.025 min Scan# 3128
 Delta R.T. 0.012 min
 Lab File: BN026038.D
 Acq: 27 Jun 2023 15:44

Tgt Ion:264 Resp: 7705
 Ion Ratio Lower Upper
 264 100
 260 27.6 21.9 32.9
 265 136.1 80.3 120.5#









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.010 ng
RT: 26.624 min Scan# 4008
Delta R.T. 0.026 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44




Instrument :
BNA_N
ClientSampleId :
PB153778BL


Tgt Ion:276 Resp: 328
Ion Ratio Lower Upper
276 100
138 11.9 21.4 32.2#
277 8.8 19.6 29.4#



#37
Benzo(b)fluoranthene
Concen: 0.009 ng
RT: 23.279 min Scan# 2873
Delta R.T. 0.021 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44

Tgt Ion:252 Resp: 263
Ion Ratio Lower Upper
252 100
253 128.2 24.1 36.1#
125 118.0 15.4 23.0#





#38

Benzo(k)fluoranthene

Concen: 0.001 ng

RT: 23.315 min Scan# 21

Delta R.T. 0.012 min

Lab File: BN026038.D


Acq: 27 Jun 2023 15:44

Instrument :

BNA_N

ClientSampleId :

PB153778BL






Tgt Ion:252 Resp: 23

Ion Ratio Lower Upper

252 100

253 127.0 23.6 35.4#

125 116.9 15.8 23.8#



#39

Benzo(a)pyrene

Concen: 0.004 ng

RT: 23.911 min Scan# 3089

Delta R.T. 0.000 min

Lab File: BN026038.D

Acq: 27 Jun 2023 15:44




Tgt Ion:252 Resp: 104


Ion Ratio Lower Upper

252 100

253 136.7 26.2 39.4#

125 119.9 18.7 28.1#









#40
Dibenzo(a,h)anthracene
Concen: 0.012 ng
RT: 26.656 min Scan# 4016
Delta R.T. 0.035 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44

Instrument :
BNA_N
ClientSampleId :
PB153778BL

Tgt Ion:278 Resp: 296
Ion Ratio Lower Upper
278 100
139 63.2 18.1 27.1#
279 165.9 25.4 38.0#



#41
Benzo(g,h,i)perylene
Concen: 0.016 ng
RT: 27.402 min Scan# 4283
Delta R.T. 0.006 min
Lab File: BN026038.D
Acq: 27 Jun 2023 15:44

Tgt Ion:276 Resp: 456
Ion Ratio Lower Upper
276 100
277 61.0 20.3 30.5#
138 73.7 21.5 32.3#

