

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN062823\  
 Data File : BN026052.D  
 Acq On : 28 Jun 2023 10:54  
 Operator : MA/JU  
 Sample : 03408-02DL 10X  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 GWOW-BR-05-390-410-062323DL

Quant Time: Jun 29 00:10:41 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN062723.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Jun 28 05:23:31 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units | Dev(Min)  |
|-----------------------------|--------|------|----------|-------|-------|-----------|
| -----                       |        |      |          |       |       |           |
| Internal Standards          |        |      |          |       |       |           |
| 1) 1,4-Dichlorobenzene-d4   | 7.976  | 152  | 5691     | 0.400 | ng    | 0.00      |
| 7) Naphthalene-d8           | 10.782 | 136  | 17587    | 0.400 | ng    | 0.00      |
| 13) Acenaphthene-d10        | 14.609 | 164  | 11401    | 0.400 | ng    | 0.00      |
| 19) Phenanthrene-d10        | 17.356 | 188  | 24359    | 0.400 | ng    | 0.00      |
| 29) Chrysene-d12            | 21.551 | 240  | 13936    | 0.400 | ng    | 0.00      |
| 35) Perylene-d12            | 24.010 | 264  | 13248    | 0.400 | ng    | # 0.00    |
| System Monitoring Compounds |        |      |          |       |       |           |
| 4) 2-Fluorophenol           | 5.535  | 112  | 209      | 0.015 | ng    | 0.00      |
| 5) Phenol-d6                | 7.153  | 99   | 145      | 0.008 | ng    | 0.01      |
| 8) Nitrobenzene-d5          | 9.170  | 82   | 380      | 0.025 | ng    | 0.02      |
| 11) 2-Methylnaphthalene-d10 | 12.385 | 152  | 991      | 0.039 | ng    | 0.01      |
| 14) 2,4,6-Tribromophenol    | 16.102 | 330  | 177      | 0.043 | ng    | 0.00      |
| 15) 2-Fluorobiphenyl        | 13.240 | 172  | 1958     | 0.044 | ng    | 0.00      |
| 27) Fluoranthene-d10        | 19.388 | 212  | 2617     | 0.051 | ng    | 0.00      |
| 31) Terphenyl-d14           | 19.978 | 244  | 2095     | 0.059 | ng    | 0.00      |
| Target Compounds            |        |      |          |       |       |           |
| 2) 1,4-Dioxane              | 3.361  | 88   | 21547    | 3.579 | ng    | Qvalue 97 |
| -----                       |        |      |          |       |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN062823\  
Data File : BN026052.D  
Acq On : 28 Jun 2023 10:54  
Operator : MA/JU  
Sample : 03408-02DL 10X  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
GWOW-BR-05-390-410-062323DL

Quant Time: Jun 29 00:10:41 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN062723.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Jun 28 05:23:31 2023  
Response via : Initial Calibration

