

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062824\
 Data File : BN032312.D
 Acq On : 28 Jun 2024 14:01
 Operator : MA/JU
 Sample : PB161558BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS558

Quant Time: Jun 28 15:13:33 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN062724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 27 13:23:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.616	152	8524	0.400	ng/ul	0.00
4) Naphthalene-d8	10.393	136	29093	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.262	164	17892	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.016	188	35736	0.400	ng/ul	0.00
17) Chrysene-d12	21.220	240	19975	0.400	ng/ul	0.00
23) Perylene-d12	23.441	264	21517	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.152	96	5401	0.570	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.993	152	13577	0.311	ng/ul	0.00
18) Fluoranthene-d10	19.055	212	23802	0.396	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.181	88	13586	1.260	ng/ul#	88
5) Naphthalene	10.442	128	24529	0.319	ng/ul	100
7) 2-Methylnaphthalene	12.065	142	15847	0.311	ng/ul	100
8) 1-Methylnaphthalene	12.285	142	15574	0.294	ng/ul	99
10) Acenaphthylene	13.980	152	24710	0.332	ng/ul	100
11) Acenaphthene	14.322	153	18676	0.323	ng/ul	98
12) Fluorene	15.322	166	20296	0.295	ng/ul	100
14) Pentachlorophenol	16.682	266	3858	0.477	ng/ul	100
15) Phenanthrene	17.058	178	30543	0.312	ng/ul	100
16) Anthracene	17.155	178	27353	0.313	ng/ul	99
19) Fluoranthene	19.082	202	30606	0.402	ng/ul	97
20) Pyrene	19.450	202	30466	0.383	ng/ul	100
21) Benzo(a)anthracene	21.208	228	23310	0.323	ng/ul	99
22) Chrysene	21.258	228	25486	0.322	ng/ul	99
24) Benzo(b)fluoranthene	22.772	252	24054	0.257	ng/ul	93
25) Benzo(k)fluoranthene	22.819	252	26112	0.287	ng/ul#	93
26) Benzo(a)pyrene	23.345	252	23596	0.363	ng/ul	94
27) Indeno(1,2,3-cd)pyrene	25.680	276	31881	0.313	ng/ul	99
28) Dibenzo(a,h)anthracene	25.690	278	24864	0.329	ng/ul	97
29) Benzo(g,h,i)perylene	26.360	276	26565	0.327	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062824\
Data File : BN032312.D
Acq On : 28 Jun 2024 14:01
Operator : MA/JU
Sample : PB161558BS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS558

Quant Time: Jun 28 15:13:33 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN062724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 27 13:23:38 2024
Response via : Initial Calibration

