

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062824\
 Data File : BN032313.D
 Acq On : 28 Jun 2024 14:37
 Operator : MA/JU
 Sample : P2898-07
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E1G85

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/29/2024
 Supervised By :mohammad ahmed 07/04/2024

Quant Time: Jun 28 15:14:07 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN062724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 27 13:23:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.616	152	9563	0.400	ng/ul	0.00
4) Naphthalene-d8	10.387	136	32143	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.258	164	19995	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.012	188	39164m	0.400	ng/ul	-0.01
17) Chrysene-d12	21.220	240	21564m	0.400	ng/ul	0.00
23) Perylene-d12	23.438	264	20910m	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.148	96	42682	4.015	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.988	152	14219	0.295	ng/ul	-0.01
18) Fluoranthene-d10	19.050	212	26573	0.410	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062824\
Data File : BN032313.D
Acq On : 28 Jun 2024 14:37
Operator : MA/JU
Sample : P2898-07
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
E1G85

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/29/2024
Supervised By :mohammad ahmed 07/04/2024

Quant Time: Jun 28 15:14:07 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN062724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 27 13:23:38 2024
Response via : Initial Calibration

