

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062824\
 Data File : BN032321.D
 Acq On : 28 Jun 2024 20:03
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4384

Quant Time: Jun 28 23:06:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN062724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 27 13:23:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.616	152	6416	0.400	ng/ul	0.00
4) Naphthalene-d8	10.393	136	19117	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.262	164	10832	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.020	188	22433	0.400	ng/ul	0.00
17) Chrysene-d12	21.220	240	20835	0.400	ng/ul	0.00
23) Perylene-d12	23.439	264	23805	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.152	96	3125	0.438	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.993	152	11013	0.384	ng/ul	0.00
18) Fluoranthene-d10	19.055	212	23180	0.370	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.186	88	3380	0.416	ng/ul	99
5) Naphthalene	10.442	128	21038	0.416	ng/ul	99
7) 2-Methylnaphthalene	12.070	142	12960	0.387	ng/ul	99
8) 1-Methylnaphthalene	12.285	142	13765	0.396	ng/ul	99
10) Acenaphthylene	13.980	152	19535	0.434	ng/ul	98
11) Acenaphthene	14.323	153	15066	0.431	ng/ul	98
12) Fluorene	15.322	166	16253	0.390	ng/ul	99
14) Pentachlorophenol	16.687	266	2085	0.411	ng/ul	99
15) Phenanthrene	17.062	178	24966	0.406	ng/ul	100
16) Anthracene	17.160	178	22776	0.415	ng/ul	99
19) Fluoranthene	19.083	202	30025	0.378	ng/ul	97
20) Pyrene	19.450	202	31043	0.374	ng/ul	100
21) Benzo(a)anthracene	21.206	228	29002	0.386	ng/ul	99
22) Chrysene	21.255	228	33120	0.401	ng/ul	99
24) Benzo(b)fluoranthene	22.769	252	33265	0.322	ng/ul	97
25) Benzo(k)fluoranthene	22.813	252	34903	0.347	ng/ul	98
26) Benzo(a)pyrene	23.342	252	32463	0.451	ng/ul	100
27) Indeno(1,2,3-cd)pyrene	25.676	276	42450	0.377	ng/ul#	98
28) Dibenzo(a,h)anthracene	25.686	278	33146	0.396	ng/ul	99
29) Benzo(g,h,i)perylene	26.361	276	35489	0.395	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062824\
Data File : BN032321.D
Acq On : 28 Jun 2024 20:03
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4384

Quant Time: Jun 28 23:06:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN062724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 27 13:23:38 2024
Response via : Initial Calibration

