

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070119\
 Data File : BN006670.D
 Acq On : 01 Jul 2019 21:40
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 02 01:54:47 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 13:47:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	120	0.00
2	1,4-Dioxane	0.400	0.388	3.0	124	0.00
3	n-Nitrosodimethylamine	0.400	0.358	10.5	121	0.00
4 S	2-Fluorophenol	0.400	0.411	-2.7	128	0.00
5 S	Phenol-d6	0.400	0.419	-4.7	132	0.00
6	bis(2-Chloroethyl)ether	0.400	0.408	-2.0	117	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	111	0.00
8 S	Nitrobenzene-d5	0.400	0.412	-3.0	123	0.00
9	Naphthalene	0.400	0.421	-5.2	115	-0.01
10	Hexachlorobutadiene	0.400	0.428	-7.0	115	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.393	1.8	107	0.00
12	2-Methylnaphthalene	0.400	0.397	0.8	107	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	94	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.379	5.3	99	0.00
15 S	2-Fluorobiphenyl	0.400	0.429	-7.2	100	-0.01
16	Acenaphthylene	0.400	0.431	-7.7	103	0.00
17	Acenaphthene	0.400	0.425	-6.2	97	0.00
18	Fluorene	0.400	0.404	-1.0	93	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	83	-0.01
20	4-Bromophenyl-phenylether	0.400	0.428	-7.0	90	-0.01
21	Hexachlorobenzene	0.400	0.433	-8.2	87	0.00
22	Pentachlorophenol	0.400	0.491	-22.7	112	-0.01
23	Phenanthrene	0.400	0.417	-4.2	86	0.00
24	Anthracene	0.400	0.412	-3.0	88	0.00
25 SURR	Fluoranthene-d10	0.400	0.396	1.0	82	-0.01
26	Fluoranthene	0.400	0.399	0.3	80	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	72	-0.02
28	Pvrene	0.400	0.459	-14.8	81	-0.01
29 S	Terphenyl-d14	0.400	0.441	-10.2	79	0.00
30	Benzo(a)anthracene	0.400	0.424	-6.0	78	-0.02
31	Chrysene	0.400	0.428	-7.0	73	-0.02
32	Bis(2-ethylhexyl)phthalate	0.400	0.502	-25.5#	101	-0.02
33	Indeno(1,2,3-cd)pyrene	0.400	0.519	-29.8#	81	-0.05
34 I	Perylene-d12	0.400	0.400	0.0	73	-0.04
35	Benzo(b)fluoranthene	0.400	0.404	-1.0	73	-0.04
36	Benzo(k)fluoranthene	0.400	0.399	0.3	71	-0.04
37 C	Benzo(a)pyrene	0.400	0.421	-5.2	74	-0.04
38	Dibenzo(a,h)anthracene	0.400	0.467	-16.8	80	-0.05
39	Benzo(a,h,i)perylene	0.400	0.480	-20.0	82	-0.05

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070119\
Data File : BN006670.D
Acq On : 01 Jul 2019 21:40
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 02 01:54:47 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 13:47:56 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				