

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070119\
 Data File : BN006686.D
 Acq On : 02 Jul 2019 07:28
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 02 08:48:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 13:47:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.554	0.554	0.0	120	0.00
3	n-Nitrosodimethylamine	0.435	0.416	4.4	121	0.00
4 S	2-Fluorophenol	0.981	1.110	-13.1	132	0.00
5 S	Phenol-d6	1.299	1.470	-13.2	134	0.00
6	bis(2-Chloroethyl)ether	1.158	1.109	4.2	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	106	-0.01
8 S	Nitrobenzene-d5	0.269	0.294	-9.3	124	-0.01
9	Naphthalene	1.051	1.111	-5.7	110	-0.01
10	Hexachlorobutadiene	0.216	0.232	-7.4	110	-0.01
11 SURR	2-Methylnaphthalene-d10	0.630	0.632	-0.3	103	0.00
12	2-Methylnaphthalene	0.726	0.740	-1.9	104	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	89	-0.01
14 S	2,4,6-Tribromophenol	0.188	0.215	-14.4	113	-0.01
15 S	2-Fluorobiphenyl	1.482	1.642	-10.8	99	-0.01
16	Acenaphthylene	1.907	2.487	-30.4#	118	0.00
17	Acenaphthene	1.256	1.427	-13.6	99	0.00
18	Fluorene	1.585	1.772	-11.8	98	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	82	-0.01
20	4-Bromophenyl-phenylether	0.222	0.246	-10.8	92	-0.01
21	Hexachlorobenzene	0.280	0.299	-6.8	85	0.00
22	Pentachlorophenol	0.045	0.068	-51.1#	157#	-0.02
23	Phenanthrene	1.137	1.222	-7.5	87	0.00
24	Anthracene	0.987	1.106	-12.1	95	0.00
25 SURR	Fluoranthene-d10	1.175	1.237	-5.3	86	0.00
26	Fluoranthene	1.382	1.456	-5.4	84	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	72	-0.01
28	Pvrene	1.400	1.680	-20.0	85	0.00
29 S	Terphenyl-d14	0.838	0.982	-17.2	84	0.00
30	Benzo(a)anthracene	1.308	1.526	-16.7	86	-0.02
31	Chrysene	1.410	1.488	-5.5	72	-0.02
32	Bis(2-ethylhexyl)phthalate	0.658	1.157	-75.8#	142	-0.01
33	Indeno(1,2,3-cd)pyrene	1.427	1.865	-30.7#	81	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	72	-0.03
35	Benzo(b)fluoranthene	1.361	1.404	-3.2	74	-0.03
36	Benzo(k)fluoranthene	1.450	1.396	3.7	68	-0.03
37 C	Benzo(a)pyrene	1.272	1.367	-7.5	75	-0.03
38	Dibenzo(a,h)anthracene	1.111	1.319	-18.7	81	-0.03
39	Benzo(a,h,i)perylene	1.109	1.363	-22.9	83	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070119\
Data File : BN006686.D
Acq On : 02 Jul 2019 07:28
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 02 08:48:16 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 13:47:56 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			