

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070222\
 Data File : BN020607.D
 Acq On : 01 Jul 2022 23:54
 Operator : CG/JU
 Sample : N3457-20
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW178S-20220621

Quant Time: Jul 02 03:13:33 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 02 03:10:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.825	152	3811	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	11544	0.400	ng	# 0.00
13) Acenaphthene-d10	14.474	164	6408	0.400	ng	0.00
19) Phenanthrene-d10	17.225	188	13152	0.400	ng	# 0.00
29) Chrysene-d12	21.431	240	10365	0.400	ng	# 0.00
35) Perylene-d12	23.782	264	8328	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.420	112	955	0.090	ng	0.00
5) Phenol-d6	7.053	99	562	0.044	ng	0.02
8) Nitrobenzene-d5	9.003	82	1960	0.210	ng	0.01
11) 2-Methylnaphthalene-d10	12.231	152	4969	0.251	ng	0.00
14) 2,4,6-Tribromophenol	15.984	330	474	0.196	ng	0.01
15) 2-Fluorobiphenyl	13.105	172	7528	0.285	ng	0.00
27) Fluoranthene-d10	19.265	212	12107	0.310	ng	0.00
31) Terphenyl-d14	19.868	244	9224	0.372	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	578	0.116	ng	# 69
34) Bis(2-ethylhexyl)phtha...	21.351	149	1218	0.058	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070222\
Data File : BN020607.D
Acq On : 01 Jul 2022 23:54
Operator : CG/JU
Sample : N3457-20
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW178S-20220621

Quant Time: Jul 02 03:13:33 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 02 03:10:12 2022
Response via : Initial Calibration

