

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070222\
 Data File : BN020615.D
 Acq On : 02 Jul 2022 04:47
 Operator : CG/JU
 Sample : N3458-11
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE120D1-20220622

Quant Time: Jul 02 05:22:41 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 02 03:10:12 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 07/05/2022
 Supervised By :Jagrut Upadhyay 07/05/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.826	152	4254	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	12492	0.400	ng	# 0.00
13) Acenaphthene-d10	14.474	164	6891	0.400	ng	0.00
19) Phenanthrene-d10	17.226	188	13339	0.400	ng	# 0.00
29) Chrysene-d12	21.431	240	9718	0.400	ng	0.00
35) Perylene-d12	23.788	264	8036	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.421	112	1006	0.085	ng	0.00
5) Phenol-d6	7.060	99	520	0.037	ng	0.03
8) Nitrobenzene-d5	9.003	82	1922m	0.190	ng	0.01
11) 2-Methylnaphthalene-d10	12.227	152	5037	0.235	ng	0.00
14) 2,4,6-Tribromophenol	15.985	330	463	0.178	ng	0.01
15) 2-Fluorobiphenyl	13.105	172	6986	0.246	ng	0.00
27) Fluoranthene-d10	19.265	212	11351	0.286	ng	0.00
31) Terphenyl-d14	19.868	244	8505	0.366	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	30035	5.387	ng	94
34) Bis(2-ethylhexyl)phtha...	21.351	149	1167	0.060	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070222\
Data File : BN020615.D
Acq On : 02 Jul 2022 04:47
Operator : CG/JU
Sample : N3458-11
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE120D1-20220622

Quant Time: Jul 02 05:22:41 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 02 03:10:12 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 07/05/2022
Supervised By : Jagrut Upadhyay 07/05/2022

