

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070222\
 Data File : BN020629.D
 Acq On : 02 Jul 2022 15:06
 Operator : CG/JU
 Sample : N3485-03
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE120D3-20220623

Quant Time: Jul 04 02:39:08 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 02 03:10:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.825	152	3940	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	11746	0.400	ng	# 0.00
13) Acenaphthene-d10	14.474	164	6440	0.400	ng	0.00
19) Phenanthrene-d10	17.225	188	12871	0.400	ng	# 0.00
29) Chrysene-d12	21.440	240	10291	0.400	ng	# 0.00
35) Perylene-d12	23.793	264	8234	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.420	112	958	0.088	ng	0.00
5) Phenol-d6	7.053	99	568	0.043	ng	0.02
8) Nitrobenzene-d5	9.003	82	2281	0.240	ng	0.01
11) 2-Methylnaphthalene-d10	12.227	152	5662	0.281	ng	0.00
14) 2,4,6-Tribromophenol	15.974	330	491	0.202	ng	0.00
15) 2-Fluorobiphenyl	13.105	172	8029	0.302	ng	0.00
27) Fluoranthene-d10	19.265	212	12381	0.324	ng	0.00
31) Terphenyl-d14	19.872	244	9281	0.377	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	2224	0.431	ng	90
34) Bis(2-ethylhexyl)phtha...	21.360	149	1334	0.064	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070222\
Data File : BN020629.D
Acq On : 02 Jul 2022 15:06
Operator : CG/JU
Sample : N3485-03
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE120D3-20220623

Quant Time: Jul 04 02:39:08 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 02 03:10:12 2022
Response via : Initial Calibration

