




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070222\
Data File : BN020601.D
Acq On : 01 Jul 2022 20:12
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 02 04:34:08 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 02 04:34:03 2022
Response via : Initial Calibration



TIC: BN020601.D\data.ms

(70) Pentachlorophenol (C)




16.886min (0.000) 15636.11 ng

response 226274

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	65.67
264.00	61.60	65.58
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070222\
 Data File : BN020601.D
 Acq On : 01 Jul 2022 20:12
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 02 04:34:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 02 04:34:03 2022
 Response via : Initial Calibration



TIC: BN020601.D\data.ms

(77) Benzidine

19.480min (0.000) 0.00 ng

response 1556957

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.44
183.00	13.20	11.87
0.00	0.00	0.00