




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070222\
Data File : BN020620.D
Acq On : 02 Jul 2022 09:38
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 20 Sample Multiplier: 1

Quant Time: Jul 04 05:02:36 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 04 05:02:31 2022
Response via : Initial Calibration



TIC: BN020620.D\data.ms

(70) Pentachlorophenol (C)




16.881min (0.000) 17321.79 ng

response 194390

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	70.35
264.00	61.60	64.54
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070222\
Data File : BN020620.D
Acq On : 02 Jul 2022 09:38
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 20 Sample Multiplier: 1

Quant Time: Jul 04 05:02:36 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 04 05:02:31 2022
Response via : Initial Calibration



TIC: BN020620.D\data.ms

(77) Benzidine

19.481min (0.000) 0.00 ng

response 1339931

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.31
183.00	13.20	11.73
0.00	0.00	0.00