

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070223\
 Data File : BN026263.D
 Acq On : 04 Jul 2023 15:36
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 121 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4332

Quant Time: Jul 05 00:03:44 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 05 00:03:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	8500	0.400	ng/ul	0.00
4) Naphthalene-d8	10.759	136	27748	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.589	164	15055	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.343	188	25487	0.400	ng/ul	0.00
17) Chrysene-d12	21.533	240	15322	0.400	ng/ul	0.00
23) Perylene-d12	23.997	264	17682	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.314	96	3252	0.340	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.348	152	15419	0.361	ng/ul	0.00
18) Fluoranthene-d10	19.373	212	22278	0.382	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.352	88	3342	0.329	ng/ul#	92
5) Naphthalene	10.809	128	29285	0.368	ng/ul	100
7) 2-Methylnaphthalene	12.420	142	17899	0.353	ng/ul	100
8) 1-Methylnaphthalene	12.640	142	18921	0.361	ng/ul	100
10) Acenaphthylene	14.312	152	28261	0.401	ng/ul	100
11) Acenaphthene	14.654	153	21394	0.371	ng/ul	99
12) Fluorene	15.644	166	22489	0.340	ng/ul	99
14) Pentachlorophenol	17.006	266	1990	0.218	ng/ul	100
15) Phenanthrene	17.386	178	30263	0.359	ng/ul	100
16) Anthracene	17.479	178	27671	0.370	ng/ul	99
19) Fluoranthene	19.401	202	30591	0.374	ng/ul	99
20) Pyrene	19.764	202	31628	0.372	ng/ul	98
21) Benzo(a)anthracene	21.515	228	24031	0.380	ng/ul	98
22) Chrysene	21.571	228	24284	0.369	ng/ul	99
24) Benzo(b)fluoranthene	23.245	252	26037	0.345	ng/ul	94
25) Benzo(k)fluoranthene	23.292	252	26360	0.348	ng/ul	97
26) Benzo(a)pyrene	23.889	252	23952	0.365	ng/ul	96
27) Indeno(1,2,3-cd)pyrene	26.582	276	31018	0.433	ng/ul#	93
28) Dibenzo(a,h)anthracene	26.595	278	23726	0.438	ng/ul	98
29) Benzo(g,h,i)perylene	27.380	276	26364	0.414	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070223\
Data File : BN026263.D
Acq On : 04 Jul 2023 15:36
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 121 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4332

Quant Time: Jul 05 00:03:44 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 05 00:03:30 2023
Response via : Initial Calibration

