

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070223\
 Data File : BN026213.D
 Acq On : 03 Jul 2023 07:00
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 70 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4329

Quant Time: Jul 03 07:51:41 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 01 01:46:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.955	152	6610	0.400	ng/ul	0.00
4) Naphthalene-d8	10.765	136	21825	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.598	164	12154	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.347	188	22055	0.400	ng/ul	0.00
17) Chrysene-d12	21.543	240	12779	0.400	ng/ul	0.00
23) Perylene-d12	23.999	264	14549	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.318	96	2501	0.336	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.354	152	12659	0.377	ng/ul	-0.01
18) Fluoranthene-d10	19.377	212	20288	0.417	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.352	88	2687	0.340	ng/ul#	87
5) Naphthalene	10.814	128	24005	0.384	ng/ul	100
7) 2-Methylnaphthalene	12.425	142	14767	0.371	ng/ul	100
8) 1-Methylnaphthalene	12.645	142	15621	0.379	ng/ul	100
10) Acenaphthylene	14.320	152	24424	0.429	ng/ul	99
11) Acenaphthene	14.658	153	18124	0.389	ng/ul	99
12) Fluorene	15.648	166	19374	0.363	ng/ul	100
14) Pentachlorophenol	17.009	266	3768	0.477	ng/ul	99
15) Phenanthrene	17.389	178	27635	0.379	ng/ul	99
16) Anthracene	17.482	178	25218	0.389	ng/ul	98
19) Fluoranthene	19.405	202	28521	0.418	ng/ul	99
20) Pyrene	19.772	202	28931	0.408	ng/ul	96
21) Benzo(a)anthracene	21.526	228	21027	0.399	ng/ul	97
22) Chrysene	21.578	228	21394	0.390	ng/ul	96
24) Benzo(b)fluoranthene	23.247	252	23026	0.371	ng/ul	92
25) Benzo(k)fluoranthene	23.294	252	22723	0.364	ng/ul	93
26) Benzo(a)pyrene	23.888	252	20540	0.381	ng/ul#	93
27) Indeno(1,2,3-cd)pyrene	26.564	276	26749	0.454	ng/ul#	91
28) Dibenzo(a,h)anthracene	26.574	278	20584	0.462	ng/ul	98
29) Benzo(g,h,i)perylene	27.362	276	22779	0.435	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070223\
Data File : BN026213.D
Acq On : 03 Jul 2023 07:00
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 70 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4329

Quant Time: Jul 03 07:51:41 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 01 01:46:24 2023
Response via : Initial Calibration

