

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070223\
 Data File : BN026214.D
 Acq On : 03 Jul 2023 07:36
 Operator : MA/JU
 Sample : PB153767BL
 Misc :
 ALS Vial : 71 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK767

Quant Time: Jul 03 08:22:46 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 01 01:46:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.951	152	7805	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.763	136	25249	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.593	164	14806	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.346	188	27780	0.400	ng/ul	0.00
17) Chrysene-d12	21.540	240	14393	0.400	ng/ul	0.00
23) Perylene-d12	24.004	264	17916	0.400	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.314	96	41141	4.681	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.358	152	10633	0.274	ng/ul	0.00
18) Fluoranthene-d10	19.376	212	17627	0.322	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070223\
Data File : BN026214.D
Acq On : 03 Jul 2023 07:36
Operator : MA/JU
Sample : PB153767BL
Misc :
ALS Vial : 71 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK767

Quant Time: Jul 03 08:22:46 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 01 01:46:24 2023
Response via : Initial Calibration

