

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN070318\
 Data File : BN001943.D
 Acq On : 03 Jul 2018 14:55
 Operator : JU/SJ
 Sample : J3721-08
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F1H19

Quant Time: Jul 04 00:41:23 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN061918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 03 23:50:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	500732	20.00	ng/ul	0.00
18) Naphthalene-d8	10.48	136	2337251	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.34	164	1429140	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.08	188	3301625	20.00	ng/ul	0.00
75) Chrysene-d12	21.28	240	3344667	20.00	ng/ul	0.00
83) Perylene-d12	23.51	264	3013070	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.27	96	61624	5.34	ng/uL	0.00
5) Phenol-d5	6.88	99	1241251	27.39	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.05	67	828791	29.64	ng/ul	0.00
9) 2-Chlorophenol-d4	7.25	132	1004416	27.96	ng/ul	0.00
13) 4-Methylphenol-d8	8.41	113	1017609	28.03	ng/ul	0.00
19) Nitrobenzene-d5	8.86	128	506677	29.19	ng/ul	0.00
22) 2-Nitrophenol-d4	9.57	143	552931	30.78	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.10	165	1032573	26.92	ng/ul	0.00
29) 4-Chloroaniline-d4	10.62	131	1398243	35.21	ng/ul	0.00
43) Dimethylphthalate-d6	13.75	166	3499948	29.00	ng/ul	0.00
46) Acenaphthylene-d8	14.03	160	4217307	28.31	ng/ul	0.00
51) 4-Nitrophenol-d4	14.53	143	437110	19.66	ng/ul	0.00
57) Fluorene-d10	15.33	176	3018957	28.54	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.45	200	376453	20.93	ng/ul	0.00
70) Anthracene-d10	17.18	188	4567918	28.64	ng/ul	0.00
76) Pyrene-d10	19.48	212	5254115	31.16	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.37	264	4909440	29.96	ng/ul	0.00

Target Compounds

					Qvalue	
6) Phenol	6.91	94	60685	1.328	ng/ul	95
44) Dimethylphthalate	13.80	163	1590564	13.409	ng/ul	99
77) Pyrene	19.51	202	284188	1.358	ng/ul#	87
80) Benzo(a)anthracene	21.26	228	219935	1.040	ng/ul	99
82) Chrysene	21.32	228	329396	1.675	ng/ul	96
85) Benzo(b)fluoranthene	22.84	252	441266	2.319	ng/ul#	95
88) Benzo(a)pyrene	23.41	252	218230	1.227	ng/ul#	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN070318\
Data File : BN001943.D
Acq On : 03 Jul 2018 14:55
Operator : JU/SJ
Sample : J3721-08
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F1H19

Quant Time: Jul 04 00:41:23 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN061918.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 03 23:50:21 2018
Response via : Initial Calibration

