


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070318\
 Data File : BN001934.D
 Acq On : 03 Jul 2018 08:46
 Operator : JU/SJ
 Sample : DFTPP85
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP85

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN061918.M
 Title : SVOA CALIBRATION
 Last Update : Wed Jul 04 00:59:54 2018



AutoFind: Scans 2473, 2474, 2475; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.2	318400	PASS
68	69	0.00	2	1.7	5295	PASS
69	198	0.00	100	37.4	311629	PASS
70	69	0.00	2	0.5	1705	PASS
127	198	10	80	46.8	390378	PASS
197	198	0.00	2	0.5	3886	PASS
198	198	100	100	100.0	833706	PASS
199	198	5	9	6.7	56069	PASS
275	198	10	60	25.7	214165	PASS
365	198	1	100	3.0	25330	PASS
441	443	0.01	100	76.7	107472	PASS
442	198	50	100	84.9	707882	PASS
443	442	15	24	19.8	140114	PASS

m/z	Abundance
36.00	348.0
37.00	1106.0
38.00	3417.0
39.00	22424.0
40.00	2848.0
41.00	954.0
43.10	760.0
44.00	15656.0
44.90	883.0
49.00	1850.0
50.00	76624.0
51.00	312896.0
52.00	16392.0
53.00	745.0
55.00	1945.0
56.00	8904.0
57.00	21608.0
58.00	990.0
58.90	418.0
61.00	3969.0
62.00	4107.0
63.00	12541.0
63.90	1879.0
65.00	7262.0
66.00	510.0
67.00	552.0
68.00	5195.0
69.00	303424.0
70.00	1632.0
71.10	391.0
73.00	2330.0
74.00	28864.0
75.00	50176.0
76.00	15131.0
77.00	338432.0
78.00	24936.0
79.00	21832.0
80.00	16544.0
81.00	23456.0
81.90	6384.0
83.00	4915.0
84.10	946.0
85.00	3728.0
86.00	5726.0
86.90	2860.0
87.90	1138.0
88.90	489.0
90.90	4725.0
92.00	6988.0
93.00	35592.0
94.00	2354.0
95.00	886.0
96.00	1884.0
96.80	736.0
98.00	27128.0
99.00	24984.0
100.00	2430.0
101.00	13354.0
101.90	635.0
102.90	4013.0
104.00	8164.0
105.00	7763.0
106.00	2724.0
107.00	99096.0
108.00	15795.0
109.00	2817.0
110.00	181952.0
111.00	27880.0
112.00	3535.0
112.90	1213.0
113.70	444.0
115.20	454.0
116.00	5110.0
116.90	73096.0
118.00	5566.0
118.90	934.0
120.00	1212.0
121.00	368.0
121.90	6343.0
122.90	9008.0
124.00	4929.0
124.90	3917.0
126.10	1760.0
127.00	354624.0
128.00	27952.0
129.00	130176.0
130.00	11863.0
130.90	1847.0
131.90	1315.0
132.90	443.0
134.00	3386.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

135.00	9469.0
136.00	3418.0
137.00	6525.0
137.90	1149.0
138.90	1123.0
140.00	1408.0
141.00	17200.0
142.00	6041.0
142.90	4184.0
143.90	728.0
144.90	777.0
146.00	2621.0
147.00	9101.0
147.90	18112.0
149.00	4867.0
149.90	1204.0
151.00	1835.0
151.70	1547.0
153.00	5120.0
153.90	4324.0
155.00	9713.0
156.00	13861.0
157.00	2444.0
158.00	3619.0
159.00	2043.0
160.00	5286.0
161.00	7884.0
161.90	2263.0
162.90	895.0
163.90	593.0
165.00	6087.0
166.00	5146.0
167.00	31240.0
167.90	22024.0
168.90	3324.0
170.00	1114.0
170.90	1825.0
171.90	3113.0
172.90	3751.0
174.00	7289.0
175.00	12219.0
176.00	3805.0
177.00	5890.0
177.90	1875.0
179.00	25264.0
180.00	19256.0
180.90	8371.0
181.90	1133.0
183.00	651.0
183.90	1853.0
185.00	12205.0
186.00	92360.0
187.00	25528.0
188.00	3235.0
188.90	5547.0
190.00	996.0
191.00	3773.0
192.00	7654.0
193.00	9197.0
194.00	1916.0
195.00	713.0
196.00	18912.0
196.80	3337.0
197.90	699200.0
198.90	48608.0
200.00	3782.0
201.40	3486.0
202.90	4050.0
204.00	23296.0
205.00	40088.0
206.00	171776.0
207.00	25872.0
208.00	6307.0
209.00	2214.0
210.00	3182.0
211.00	6520.0
212.90	340.0
215.00	1690.0
216.00	3488.0
216.90	45040.0
218.00	5994.0
218.90	502.0
219.70	489.0
221.00	29888.0
221.80	5660.0
223.00	10435.0
224.00	93472.0
225.00	25496.0
226.10	3228.0
226.90	38840.0
228.00	5622.0
228.90	8414.0
230.00	971.0
231.00	3220.0
231.90	642.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

232.90	517.0
233.90	2517.0
234.90	3224.0
236.00	1880.0
237.00	5122.0
237.90	705.0
238.90	1713.0
239.90	1260.0
240.90	2235.0
242.00	4672.0
243.00	5370.0
244.00	72552.0
245.00	9897.0
245.90	13295.0
247.00	3462.0
248.00	653.0
248.90	3032.0
250.00	565.0
250.90	435.0
252.00	599.0
252.90	1262.0
254.00	2000.0
255.00	375424.0
256.00	55024.0
256.90	4275.0
257.90	20880.0
259.00	3781.0
260.00	652.0
260.90	578.0
263.00	340.0
264.00	770.0
265.00	8993.0
266.00	1349.0
270.10	349.0
271.00	874.0
272.00	848.0
272.90	9930.0
274.00	27976.0
275.00	161472.0
276.00	21392.0
276.90	13618.0
277.90	2697.0
279.00	509.0
281.00	635.0
282.00	714.0
282.90	1440.0
284.00	1076.0
285.00	2976.0
286.00	484.0
289.00	750.0
289.90	377.0
290.70	317.0
291.90	802.0
292.90	3651.0
294.00	962.0
294.90	623.0
296.00	42200.0
297.00	6458.0
297.90	453.0
301.00	834.0
302.00	699.0
302.90	4848.0
303.90	1378.0
307.90	635.0
309.00	741.0
309.90	727.0
312.90	448.0
313.90	1996.0
314.90	4055.0
316.00	2819.0
317.00	697.0
321.00	1621.0
321.90	949.0
323.00	13334.0
324.00	2564.0
325.90	329.0
326.90	2728.0
328.00	1371.0
329.00	533.0
332.00	1022.0
332.90	1724.0
334.00	9865.0
335.00	2495.0
335.90	656.0
341.00	1560.0
341.90	495.0
345.90	2402.0
346.80	419.0
352.00	3525.0
353.00	2589.0
354.00	4507.0
354.80	612.0
359.00	324.0
365.00	18152.0
365.90	3308.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

369.90	412.0
370.90	814.0
372.00	6959.0
373.00	1661.0
373.90	342.0
382.90	1549.0
384.00	509.0
390.00	730.0
390.90	968.0
392.00	684.0
400.90	575.0
401.90	2400.0
403.00	3462.0
404.00	1881.0
404.90	515.0
414.90	306.0
421.00	3698.0
422.00	3596.0
423.00	23144.0
424.00	9602.0
425.00	1402.0
441.00	68544.0
442.00	434240.0
443.00	85112.0
444.00	8097.0
444.90	528.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

m/z	Abundance
36.00	314.0
37.00	1278.0
38.00	3533.0
39.00	26032.0
40.00	2796.0
41.10	1425.0
41.90	361.0
43.10	800.0
44.00	14736.0
45.00	1186.0
49.00	2189.0
50.00	92848.0
51.00	379776.0
52.00	17680.0
53.00	955.0
55.00	2778.0
56.00	11243.0
57.00	26416.0
57.90	1391.0
58.90	464.0
60.10	464.0
61.00	4484.0
61.90	5636.0
63.00	15897.0
64.00	1973.0
65.00	7787.0
65.90	643.0
67.00	640.0
67.90	5769.0
69.00	365824.0
70.00	2028.0
71.00	512.0
73.00	3751.0
74.00	36528.0
75.00	58160.0
76.00	19152.0
77.00	419584.0
78.00	29928.0
79.00	25992.0
80.00	21096.0
81.00	28976.0
82.00	7030.0
82.90	6802.0
83.90	951.0
84.90	5443.0
86.00	8224.0
87.00	3978.0
88.00	1491.0
89.00	800.0
90.90	5872.0
91.90	7687.0
93.00	45712.0
94.00	3620.0
95.00	718.0
96.10	2204.0
97.00	1101.0
98.00	33000.0
99.00	29248.0
100.00	2590.0
100.90	16864.0
102.10	1133.0
103.00	5383.0
104.00	11216.0
105.00	8928.0
105.90	2356.0
107.00	127000.0
108.00	19360.0
109.00	2758.0
110.00	235520.0
111.00	36040.0
111.90	4626.0
112.90	1301.0
114.90	570.0
116.00	7704.0
117.00	97288.0
117.90	7916.0
119.00	1001.0
120.00	1780.0
120.90	684.0
122.00	8863.0
123.00	13041.0
123.90	6218.0
124.90	4876.0
126.00	1316.0
127.00	466240.0
128.00	32624.0
129.00	175040.0
130.00	14943.0
131.00	2896.0
132.00	1260.0
132.70	708.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

134.00	4925.0
135.00	13721.0
136.00	5869.0
137.00	8776.0
137.90	1694.0
138.90	1141.0
139.90	1982.0
141.00	20768.0
142.00	6613.0
142.90	4744.0
143.90	1492.0
144.90	1227.0
145.90	3667.0
147.00	11052.0
148.00	26976.0
148.90	5741.0
149.90	1613.0
151.00	3158.0
151.80	1714.0
153.00	6553.0
153.90	5338.0
155.00	13205.0
156.00	17592.0
157.00	3383.0
157.90	4834.0
159.00	3148.0
160.00	6436.0
161.00	9681.0
162.00	3093.0
163.00	711.0
163.90	1054.0
165.00	8164.0
166.00	7490.0
167.00	42168.0
168.00	26240.0
169.00	4530.0
169.90	1229.0
170.90	2052.0
171.90	3813.0
173.00	4950.0
173.90	8907.0
175.00	18256.0
176.00	5403.0
176.90	7881.0
178.00	2956.0
178.90	34080.0
179.90	24496.0
181.00	10414.0
181.90	1925.0
182.90	870.0
184.00	3108.0
185.00	17480.0
186.00	133248.0
187.00	35480.0
188.00	3986.0
188.90	7942.0
189.90	1593.0
191.00	3635.0
192.00	9829.0
193.00	12253.0
193.90	2482.0
194.90	1425.0
196.00	27392.0
196.90	4454.0
197.90	1000000.0
198.90	67096.0
200.00	5522.0
201.50	3899.0
202.90	5368.0
204.00	35376.0
205.00	59152.0
206.00	254528.0
207.00	35808.0
208.00	8669.0
208.90	3037.0
211.00	9714.0
212.90	794.0
214.90	2671.0
215.90	5680.0
216.90	65952.0
218.00	7958.0
219.00	992.0
221.00	43664.0
221.80	11040.0
223.00	14469.0
224.00	135936.0
225.00	35568.0
226.00	4333.0
227.00	56952.0
228.00	7077.0
229.00	12408.0
229.90	2034.0
231.00	5827.0
231.90	816.0
232.90	959.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

233.90	4078.0
235.00	4178.0
236.00	2663.0
237.00	5825.0
238.00	755.0
239.00	2125.0
239.90	1721.0
241.00	2833.0
241.90	6169.0
243.00	7396.0
244.00	111640.0
245.00	12911.0
245.90	19504.0
247.00	4452.0
248.00	880.0
248.90	4013.0
250.00	882.0
250.80	580.0
251.90	911.0
252.90	2156.0
253.90	2931.0
255.00	573568.0
256.00	79840.0
257.00	6524.0
258.00	30304.0
259.00	5088.0
259.90	1042.0
261.00	910.0
262.90	423.0
263.90	989.0
264.90	12129.0
265.90	2003.0
269.90	894.0
270.90	1326.0
272.00	1461.0
273.00	16146.0
274.00	43232.0
275.00	255168.0
276.00	33624.0
277.00	21616.0
278.00	4059.0
279.00	713.0
281.10	428.0
282.00	688.0
282.90	2589.0
284.00	1423.0
285.00	3680.0
286.00	666.0
288.90	911.0
290.00	590.0
290.90	576.0
291.90	1141.0
293.00	4805.0
293.90	1794.0
295.00	1519.0
296.00	66512.0
297.00	9519.0
297.90	584.0
300.90	1191.0
301.90	899.0
303.00	7805.0
304.00	2563.0
304.80	497.0
307.90	1249.0
308.90	564.0
309.90	967.0
313.00	552.0
314.00	3483.0
314.90	7508.0
316.00	4513.0
316.90	796.0
319.80	355.0
321.00	2518.0
322.00	1429.0
323.00	22008.0
324.00	4056.0
325.00	393.0
325.80	393.0
326.90	4319.0
328.00	2211.0
329.00	306.0
331.90	1465.0
333.00	2079.0
334.00	13468.0
335.00	3732.0
335.80	428.0
339.00	355.0
339.90	354.0
341.00	2898.0
342.00	468.0
345.90	4778.0
347.00	1015.0
350.90	572.0
352.00	6295.0
353.00	4107.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

354.00	7416.0
355.00	1791.0
359.00	450.0
364.90	29448.0
365.90	4620.0
370.00	616.0
371.00	1630.0
372.00	11723.0
373.00	3421.0
373.90	320.0
382.90	3865.0
383.90	625.0
384.20	673.0
384.90	364.0
390.00	1812.0
390.90	1089.0
391.90	680.0
396.90	319.0
401.00	692.0
402.00	4971.0
403.00	6490.0
404.00	2821.0
404.80	334.0
405.10	354.0
410.00	376.0
415.00	358.0
421.00	6373.0
422.00	5650.0
423.00	42744.0
424.00	14834.0
425.10	2205.0
441.00	123408.0
442.00	814080.0
443.00	161664.0
444.00	15443.0
445.00	784.0

Instrument :

BNA_N

ClientSampleId :

DFTPP85

m/z	Abundance
37.00	871.0
38.00	2964.0
39.00	18104.0
39.90	2743.0
41.00	1755.0
43.10	984.0
44.00	15430.0
45.00	553.0
49.00	2075.0
50.00	61904.0
51.00	262528.0
52.00	13249.0
53.00	583.0
54.90	1864.0
56.00	7962.0
57.00	18472.0
57.90	1142.0
59.00	365.0
59.90	362.0
61.00	3911.0
62.10	4044.0
63.00	10626.0
64.00	1540.0
65.00	5032.0
65.90	440.0
67.10	488.0
68.00	4921.0
69.00	266816.0
70.00	1457.0
72.90	2978.0
74.00	26480.0
75.00	42824.0
76.00	15178.0
77.00	303232.0
78.00	21704.0
79.00	19656.0
80.00	14022.0
81.00	20976.0
81.90	5141.0
83.00	4857.0
84.00	798.0
85.00	3891.0
86.00	5017.0
87.00	3150.0
88.00	1290.0
88.90	548.0
91.00	4515.0
92.00	5050.0
93.00	34488.0
94.00	2168.0
95.00	800.0
96.00	1648.0
97.00	564.0
98.00	26144.0
99.00	20392.0
99.90	1803.0
101.00	13108.0
101.90	468.0
102.90	3862.0
103.90	8003.0
104.90	6436.0
106.00	2166.0
107.00	95672.0
108.00	14875.0
109.00	3013.0
109.90	178240.0
111.00	26560.0
112.00	3456.0
112.90	1142.0
114.90	532.0
116.00	4707.0
117.00	71312.0
118.00	5658.0
118.90	979.0
120.00	1382.0
120.80	594.0
121.90	6578.0
122.90	9402.0
123.90	4621.0
124.90	4130.0
126.10	1012.0
127.00	350272.0
128.00	26048.0
129.00	130544.0
130.00	11457.0
131.00	2382.0
131.90	1134.0
133.00	422.0
134.00	4210.0
134.90	11077.0
135.90	4012.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

137.00	5753.0
138.00	1362.0
138.90	695.0
140.00	1886.0
141.00	16480.0
141.90	5378.0
142.90	4007.0
144.00	806.0
144.90	1249.0
146.00	2640.0
147.00	8739.0
147.90	18528.0
149.00	4052.0
149.90	1349.0
151.00	1988.0
151.90	1415.0
152.90	5803.0
154.00	3661.0
155.00	9386.0
156.00	14737.0
157.00	2715.0
157.90	3329.0
159.00	2697.0
159.90	5630.0
161.00	7464.0
162.00	2183.0
163.00	562.0
164.00	786.0
164.90	6032.0
166.00	5431.0
167.00	35008.0
168.00	19024.0
169.00	3421.0
169.80	1472.0
170.90	1753.0
172.00	3568.0
172.90	3969.0
174.00	7582.0
175.00	13829.0
176.00	4939.0
176.90	5803.0
177.90	1956.0
178.90	26624.0
180.00	19784.0
181.00	8952.0
182.00	1666.0
182.90	921.0
183.90	2268.0
185.00	14250.0
186.00	101736.0
187.00	29856.0
188.00	3528.0
188.90	6516.0
190.00	816.0
191.10	2568.0
191.90	9105.0
193.00	10228.0
194.00	2218.0
195.90	23208.0
196.90	3868.0
197.90	801920.0
198.90	52504.0
200.00	4133.0
201.40	3692.0
203.00	5477.0
204.00	28544.0
205.00	46256.0
206.00	207552.0
207.00	28928.0
208.00	7224.0
209.00	2151.0
210.30	3103.0
211.00	7819.0
212.10	1095.0
212.90	781.0
215.00	2060.0
215.90	4269.0
216.90	55480.0
217.90	7052.0
218.90	486.0
219.40	420.0
221.00	32688.0
223.00	11600.0
224.00	113968.0
225.00	28088.0
226.10	3740.0
227.00	47664.0
227.90	6640.0
229.00	9971.0
229.90	1399.0
231.00	4497.0
232.00	764.0
232.90	837.0
234.00	2803.0
235.00	3330.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

235.90	2435.0
237.00	4026.0
238.10	490.0
239.00	2128.0
239.90	1506.0
240.90	2917.0
242.00	6211.0
243.00	7220.0
244.00	91760.0
245.00	13217.0
245.90	15364.0
246.90	3532.0
248.00	701.0
248.90	3300.0
249.90	751.0
250.90	934.0
252.00	981.0
252.90	1942.0
254.00	3042.0
255.00	481152.0
256.00	69960.0
257.00	5416.0
257.90	25144.0
259.00	3756.0
260.00	804.0
261.00	860.0
263.80	869.0
264.90	11153.0
265.90	1674.0
269.90	501.0
270.90	1033.0
272.00	1167.0
273.00	14545.0
274.00	38016.0
275.00	225856.0
276.00	29304.0
277.00	17184.0
278.00	2955.0
278.90	524.0
281.00	736.0
282.00	413.0
283.00	1953.0
283.90	1403.0
285.00	2851.0
286.00	625.0
288.90	830.0
290.00	396.0
290.90	423.0
291.90	840.0
292.90	4243.0
294.00	1084.0
295.00	1069.0
296.00	59416.0
297.00	8423.0
297.90	552.0
300.90	681.0
302.00	1203.0
303.00	7386.0
303.90	2033.0
304.90	386.0
307.90	1217.0
308.90	791.0
309.90	847.0
313.00	691.0
313.90	2793.0
314.90	6700.0
316.00	3838.0
316.90	683.0
320.90	1798.0
321.90	1189.0
323.00	19256.0
324.00	4286.0
325.90	775.0
326.90	3708.0
328.00	1926.0
332.00	1566.0
333.00	1915.0
334.00	13153.0
335.00	3768.0
335.90	444.0
339.00	447.0
340.00	499.0
341.00	2046.0
342.00	1049.0
346.00	4723.0
346.90	985.0
352.00	5729.0
353.00	4555.0
354.00	6559.0
355.00	1333.0
359.00	607.0
364.90	28392.0
365.90	4205.0
366.90	365.0
369.90	862.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

370.90	1629.0
372.00	10391.0
372.90	2495.0
373.90	352.0
377.00	347.0
383.00	2932.0
384.00	822.0
385.00	392.0
389.90	1179.0
391.00	1155.0
392.00	965.0
400.90	1134.0
401.90	4519.0
403.00	7331.0
404.00	2994.0
405.00	579.0
414.90	332.0
421.00	6403.0
421.90	6398.0
423.00	44152.0
424.00	12083.0
425.00	1633.0
440.00	543.0
441.00	130464.0
442.00	875328.0
443.00	173568.0
444.00	16079.0
445.00	622.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

m/z	Abundance
39.90	2406.0
41.00	788.0
43.10	663.0
44.00	17304.0
48.90	609.0
55.00	915.0
57.10	564.0
69.00	392.0
72.90	587.0
78.10	334.0
83.00	491.0
83.90	340.0
95.90	449.0
146.90	310.0
148.90	310.0
207.00	2160.0
207.90	567.0
281.00	561.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85