

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070324\
 Data File : BN032431.D
 Acq On : 03 Jul 2024 10:44
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4393

Quant Time: Jul 03 11:10:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN062724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 03 03:26:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.603	152	7205	0.400	ng/ul	0.00
4) Naphthalene-d8	10.382	136	21945	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.249	164	13101	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.008	188	25683	0.400	ng/ul	0.00
17) Chrysene-d12	21.212	240	20373	0.400	ng/ul	0.00
23) Perylene-d12	23.430	264	23991	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.148	96	3801	0.475	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.977	152	13190	0.401	ng/ul	0.00
18) Fluoranthene-d10	19.045	212	26129	0.426	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.177	88	3742	0.410	ng/ul#	75
5) Naphthalene	10.426	128	24040	0.414	ng/ul	100
7) 2-Methylnaphthalene	12.054	142	15445	0.401	ng/ul	99
8) 1-Methylnaphthalene	12.274	142	16530	0.414	ng/ul	99
10) Acenaphthylene	13.971	152	24810	0.455	ng/ul	99
11) Acenaphthene	14.313	153	17652	0.417	ng/ul	99
12) Fluorene	15.308	166	19474	0.386	ng/ul	99
14) Pentachlorophenol	16.674	266	2896	0.498	ng/ul	98
15) Phenanthrene	17.050	178	29435	0.418	ng/ul	99
16) Anthracene	17.143	178	28450	0.453	ng/ul	99
19) Fluoranthene	19.073	202	33323	0.430	ng/ul	96
20) Pyrene	19.441	202	34317	0.423	ng/ul	99
21) Benzo(a)anthracene	21.197	228	31184	0.424	ng/ul	98
22) Chrysene	21.247	228	31250	0.387	ng/ul	97
24) Benzo(b)fluoranthene	22.763	252	32212	0.309	ng/ul	92
25) Benzo(k)fluoranthene	22.807	252	34018	0.336	ng/ul#	93
26) Benzo(a)pyrene	23.333	252	31879	0.440	ng/ul	96
27) Indeno(1,2,3-cd)pyrene	25.663	276	41306	0.364	ng/ul	100
28) Dibenzo(a,h)anthracene	25.673	278	32995	0.391	ng/ul#	94
29) Benzo(g,h,i)perylene	26.354	276	33738	0.373	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070324\
Data File : BN032431.D
Acq On : 03 Jul 2024 10:44
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4393

Quant Time: Jul 03 11:10:55 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN062724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 03 03:26:42 2024
Response via : Initial Calibration

