


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN070518\
 Data File : BN001958.D
 Acq On : 05 Jul 2018 09:20
 Operator : JU/SJ
 Sample : DFTPP86
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP86

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN061918.M
 Title : SVOA CALIBRATION
 Last Update : Fri Jul 06 12:27:03 2018



AutoFind: Scans 2473, 2474, 2475; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.2	286058	PASS
68	69	0.00	2	1.7	4852	PASS
69	198	0.00	100	35.8	282624	PASS
70	69	0.00	2	0.5	1425	PASS
127	198	10	80	45.4	358485	PASS
197	198	0.00	2	0.1	1089	PASS
198	198	100	100	100.0	789994	PASS
199	198	5	9	6.6	52250	PASS
275	198	10	60	26.0	205610	PASS
365	198	1	100	2.9	23165	PASS
441	443	0.01	100	78.6	108986	PASS
442	198	50	100	89.0	702869	PASS
443	442	15	24	19.7	138640	PASS

m/z	Abundance
36.00	342.0
37.00	730.0
38.00	3270.0
39.00	20776.0
40.00	2963.0
41.00	693.0
43.00	515.0
44.00	15216.0
44.90	969.0
49.00	1899.0
50.00	70952.0
51.00	297344.0
52.00	15135.0
52.90	833.0
55.00	1437.0
56.00	9059.0
57.00	21288.0
58.00	1189.0
59.00	403.0
61.00	3816.0
62.00	4389.0
63.00	12287.0
64.00	1537.0
65.00	6046.0
66.10	606.0
66.90	451.0
68.00	5181.0
69.00	287616.0
70.00	1550.0
70.90	435.0
73.00	2410.0
74.00	27232.0
75.00	45960.0
76.00	15338.0
77.00	330048.0
78.00	22840.0
79.00	22232.0
79.90	15206.0
81.00	22408.0
82.00	5669.0
83.00	4938.0
83.80	742.0
85.00	4164.0
86.00	6873.0
87.00	2650.0
88.00	1211.0
89.00	617.0
90.90	5786.0
91.90	5554.0
93.00	35080.0
94.00	2431.0
95.00	673.0
96.00	2068.0
97.10	646.0
98.00	25952.0
99.00	24272.0
100.00	2420.0
101.00	12342.0
101.90	434.0
103.00	4053.0
103.90	7896.0
105.00	6963.0
105.90	2525.0
107.00	95896.0
108.00	14573.0
109.10	2634.0
110.00	184832.0
111.00	27520.0
112.00	2804.0
112.80	837.0
115.00	542.0
116.00	5321.0
117.00	66992.0
118.00	5668.0
119.00	829.0
120.00	1172.0
121.90	5360.0
123.00	9659.0
123.90	4193.0
125.00	4523.0
126.10	1775.0
127.00	344000.0
128.00	26880.0
129.00	129632.0
130.00	10811.0
131.00	1847.0
131.90	1209.0
132.90	532.0
133.90	3584.0
135.00	9144.0
136.00	4152.0

Instrument :
BNA_N
ClientSampleId :
DFTPP86

137.00	5674.0
138.00	1558.0
138.90	609.0
140.00	2012.0
141.00	16216.0
142.00	5272.0
142.90	3282.0
144.00	908.0
145.00	747.0
146.00	3220.0
146.90	7777.0
148.00	17424.0
149.00	4227.0
149.90	1330.0
151.30	2274.0
152.90	5189.0
154.00	3526.0
155.00	9166.0
156.00	13888.0
156.90	2656.0
158.00	3789.0
158.90	2049.0
160.00	5080.0
161.00	7422.0
162.00	2153.0
163.00	729.0
163.90	614.0
165.00	6017.0
165.90	4105.0
167.00	30432.0
168.00	18248.0
169.00	3179.0
170.10	1050.0
171.00	1549.0
171.90	2765.0
173.00	3733.0
174.00	6549.0
175.00	13255.0
176.00	3758.0
176.90	6084.0
177.90	2235.0
179.00	25192.0
180.00	16832.0
180.90	8005.0
181.90	1360.0
182.70	645.0
184.00	1750.0
185.00	11890.0
186.00	93840.0
187.00	29104.0
188.00	3051.0
188.90	5406.0
189.90	1055.0
190.90	2819.0
191.90	7034.0
193.00	8474.0
194.00	1737.0
196.00	18552.0
196.70	3849.0
197.90	707456.0
198.90	50704.0
200.00	3448.0
201.40	3867.0
203.00	4986.0
204.00	24392.0
205.00	41352.0
206.00	176128.0
207.00	27240.0
208.00	6266.0
209.00	2179.0
210.00	2448.0
210.90	6780.0
212.80	494.0
214.90	1881.0
215.90	3914.0
216.90	45800.0
218.00	5360.0
219.00	608.0
221.00	26384.0
223.00	10357.0
224.00	96008.0
225.00	24400.0
226.00	2809.0
227.00	36872.0
228.00	5568.0
228.90	7736.0
229.90	1201.0
231.00	3537.0
231.90	703.0
232.90	768.0
234.00	2780.0
235.00	3020.0
235.90	1441.0
237.00	3922.0
238.00	634.0

Instrument :
BNA_N
ClientSampleId :
DFTPP86

238.90	1246.0
239.90	1534.0
241.00	1893.0
242.00	4220.0
243.00	4690.0
244.00	77552.0
245.00	9804.0
245.90	13422.0
246.90	2880.0
248.00	828.0
248.90	2774.0
250.00	476.0
250.90	568.0
251.90	1052.0
253.00	1801.0
254.00	2415.0
255.00	377984.0
256.00	54448.0
257.00	4537.0
257.90	18760.0
259.00	3553.0
260.00	614.0
260.90	1038.0
263.90	691.0
264.90	8085.0
265.90	1746.0
269.80	328.0
270.90	926.0
272.00	800.0
272.90	10101.0
274.00	28128.0
275.00	163648.0
276.00	23144.0
277.00	13033.0
278.00	2106.0
279.00	573.0
281.00	705.0
282.00	528.0
283.00	1593.0
284.00	1090.0
285.00	1818.0
286.10	633.0
288.90	644.0
290.00	489.0
291.90	610.0
292.90	3115.0
294.00	852.0
295.00	1017.0
296.00	41856.0
297.00	5631.0
298.00	453.0
301.10	627.0
302.00	841.0
303.00	4526.0
304.00	1392.0
304.90	305.0
307.90	588.0
308.90	641.0
309.80	528.0
313.00	443.0
313.90	2396.0
314.90	5151.0
316.00	2942.0
317.00	628.0
320.90	1215.0
321.90	840.0
323.00	14174.0
324.10	2705.0
327.00	2124.0
327.90	1347.0
332.00	896.0
332.90	1348.0
334.00	9664.0
335.00	2065.0
336.10	377.0
341.00	1498.0
342.00	494.0
345.90	2870.0
346.80	504.0
352.00	3949.0
353.00	3092.0
354.00	4694.0
354.90	844.0
358.90	457.0
365.00	16784.0
365.90	2375.0
370.90	871.0
372.00	7693.0
373.00	1890.0
383.00	1848.0
383.90	791.0
390.00	794.0
391.00	587.0
391.90	500.0
402.00	2669.0

Instrument :
BNA_N
ClientSampleId :
DFTPP86

403.00	4200.0
403.90	1746.0
404.90	541.0
421.00	3586.0
422.00	3532.0
423.00	23544.0
424.00	6959.0
425.00	1081.0
441.00	73632.0
442.00	467008.0
443.00	98224.0
444.00	9609.0
445.00	590.0

Instrument :
BNA_N
ClientSampleId :
DFTPP86

m/z	Abundance
37.00	1172.0
38.00	3829.0
39.00	24128.0
40.00	2844.0
41.00	896.0
43.00	443.0
44.00	13941.0
45.00	881.0
49.00	2123.0
50.00	82488.0
51.00	340992.0
52.00	17360.0
53.00	704.0
55.00	2066.0
56.00	10151.0
57.00	23856.0
57.90	1372.0
58.90	310.0
59.90	317.0
60.90	4536.0
62.00	4751.0
63.00	14187.0
64.00	1779.0
65.00	7399.0
65.90	749.0
66.90	659.0
68.00	5471.0
69.00	341760.0
70.00	1559.0
73.00	2920.0
74.00	34632.0
75.00	54912.0
76.00	17616.0
77.00	403136.0
78.00	27744.0
79.00	24264.0
80.00	18824.0
81.00	27128.0
82.00	6764.0
83.00	6727.0
83.90	835.0
85.00	4544.0
86.00	6899.0
87.00	2894.0
87.90	1410.0
89.00	775.0
90.90	6566.0
92.00	6417.0
93.00	41392.0
94.10	2885.0
95.00	970.0
96.00	2239.0
97.10	614.0
98.00	30912.0
99.00	26376.0
99.90	1963.0
101.00	16292.0
101.90	848.0
102.90	5174.0
104.00	9840.0
105.00	8617.0
106.00	3107.0
107.00	118816.0
108.00	18112.0
109.00	3797.0
110.00	224448.0
111.00	33288.0
112.00	4302.0
113.10	1269.0
114.90	511.0
116.00	5534.0
117.00	86056.0
118.00	6210.0
118.90	895.0
119.90	1534.0
121.00	352.0
122.00	7400.0
123.00	10911.0
123.90	5353.0
124.90	5166.0
126.00	1771.0
127.00	435712.0
128.00	34272.0
129.00	164864.0
130.00	12946.0
130.90	2747.0
132.00	1017.0
133.00	712.0
133.90	4607.0
135.00	12765.0
136.00	5954.0

Instrument :
BNA_N
ClientSampleId :
DFTPP86

137.00	7025.0
137.80	1504.0
138.90	904.0
140.00	1883.0
141.00	20776.0
141.90	6987.0
143.00	5250.0
143.90	1464.0
144.90	1271.0
146.00	2759.0
147.00	11164.0
148.00	22152.0
149.00	5426.0
149.90	1038.0
151.10	2965.0
152.90	6771.0
154.00	6144.0
155.00	10790.0
156.00	18128.0
157.00	3452.0
158.00	3795.0
159.00	3361.0
160.00	6748.0
161.00	10116.0
162.00	3162.0
163.00	752.0
164.10	1292.0
165.00	7663.0
166.00	6757.0
167.00	39504.0
168.00	19744.0
169.00	3608.0
169.90	1482.0
171.00	1822.0
171.90	3709.0
173.00	5388.0
174.00	8240.0
175.00	17048.0
176.00	4981.0
176.90	8346.0
177.90	2472.0
179.00	30840.0
180.00	21408.0
181.00	10531.0
182.00	1467.0
183.00	968.0
183.90	2675.0
185.00	15039.0
186.00	124176.0
187.00	33408.0
188.00	3592.0
188.90	7513.0
189.90	1025.0
191.00	3810.0
192.00	10141.0
193.00	11393.0
194.00	2375.0
195.00	1153.0
196.00	26360.0
196.60	4406.0
197.90	952384.0
198.90	62592.0
200.00	5043.0
201.40	4241.0
202.90	6022.0
204.00	30416.0
205.00	55152.0
206.00	241920.0
207.00	32992.0
208.00	8273.0
209.00	2525.0
209.90	3345.0
211.00	9036.0
212.90	534.0
215.00	2299.0
215.90	4810.0
216.90	61224.0
218.00	7963.0
218.90	808.0
221.00	40416.0
221.80	9161.0
223.00	13359.0
224.00	130472.0
225.00	33680.0
226.10	3882.0
226.90	54960.0
228.00	7202.0
229.00	10358.0
229.90	1925.0
231.00	4152.0
231.90	736.0
233.00	1250.0
234.00	3071.0
235.00	4253.0
236.00	2224.0

Instrument :
BNA_N
ClientSampleId :
DFTPP86

236.90	4372.0
237.90	763.0
238.90	2310.0
240.00	1908.0
240.90	2998.0
242.00	6722.0
243.00	6822.0
244.00	103072.0
245.00	13708.0
246.00	19448.0
246.90	3438.0
247.90	1449.0
249.00	3714.0
249.90	1038.0
250.90	947.0
252.00	1107.0
252.90	2027.0
254.00	4129.0
255.00	541440.0
256.00	81728.0
257.00	6194.0
258.00	29064.0
259.00	4446.0
259.80	856.0
260.90	1067.0
262.90	358.0
263.90	741.0
264.90	11985.0
265.90	1888.0
269.90	647.0
270.90	869.0
272.00	1294.0
273.00	17112.0
274.00	42792.0
275.00	252672.0
276.00	32496.0
277.00	19400.0
277.90	3223.0
278.90	775.0
281.00	838.0
281.90	556.0
282.90	2147.0
284.00	1293.0
285.00	3403.0
285.90	624.0
288.00	344.0
289.00	697.0
289.80	705.0
290.90	567.0
292.00	969.0
293.00	4163.0
294.00	1039.0
295.00	1398.0
296.00	63056.0
297.00	8619.0
298.00	534.0
300.90	886.0
302.00	1311.0
303.00	7379.0
304.10	2136.0
305.00	399.0
308.00	863.0
308.90	732.0
310.00	880.0
311.00	401.0
313.00	465.0
313.90	3464.0
314.90	6636.0
316.00	3792.0
316.90	1106.0
321.00	2030.0
322.00	1253.0
323.00	23736.0
324.00	4360.0
325.90	525.0
326.90	4079.0
328.00	1891.0
328.90	536.0
332.00	2097.0
333.00	1907.0
334.00	14694.0
335.00	3925.0
336.00	462.0
341.00	2307.0
341.90	667.0
345.90	4273.0
347.00	864.0
352.00	7013.0
353.00	5151.0
354.00	7453.0
355.10	1156.0
358.90	451.0
364.90	28640.0
365.90	4636.0
370.00	597.0

Instrument :
BNA_N
ClientSampleId :
DFTPP86

371.00	1796.0
372.00	11609.0
373.00	3451.0
373.90	316.0
383.00	3143.0
383.90	1207.0
390.00	1706.0
391.00	1099.0
392.00	646.0
401.00	865.0
402.00	5153.0
403.00	7512.0
404.00	2474.0
405.00	598.0
414.80	366.0
421.00	6146.0
422.00	5790.0
423.00	40512.0
424.00	10414.0
425.00	880.0
441.00	129920.0
442.00	818752.0
443.00	164800.0
444.00	14916.0
445.00	917.0

Instrument :
BNA_N
ClientSampleId :
DFTPP86

m/z	Abundance
37.00	841.0
38.00	1869.0
39.00	14940.0
40.00	2353.0
41.00	877.0
43.10	324.0
44.00	15107.0
44.90	707.0
49.00	1386.0
50.00	52608.0
51.00	219840.0
52.00	10780.0
53.00	516.0
55.00	1292.0
56.00	5901.0
57.00	14937.0
57.90	824.0
59.80	345.0
61.00	2949.0
62.00	2735.0
63.00	8748.0
64.00	1477.0
65.00	4777.0
68.00	3904.0
69.00	218496.0
70.00	1168.0
72.90	1821.0
74.00	20200.0
75.00	34408.0
76.00	11806.0
77.00	258880.0
78.00	16816.0
79.00	15463.0
80.00	11539.0
81.00	17104.0
82.00	4575.0
83.00	3674.0
84.00	644.0
85.00	3658.0
85.90	5153.0
86.90	2261.0
87.90	1033.0
89.10	478.0
91.10	3833.0
92.00	4086.0
93.00	28224.0
94.00	1528.0
94.90	628.0
96.00	1531.0
98.00	21048.0
99.00	16968.0
99.90	1639.0
101.00	9760.0
101.90	826.0
102.90	3047.0
103.90	6620.0
104.90	6358.0
106.00	2064.0
107.00	78936.0
108.00	11635.0
108.90	2674.0
110.00	153216.0
111.00	21576.0
112.00	2611.0
112.90	873.0
114.90	464.0
116.00	4108.0
117.00	56736.0
118.00	4567.0
118.80	626.0
119.90	870.0
120.90	331.0
122.00	5287.0
123.00	9055.0
123.90	3555.0
125.00	3978.0
126.10	1545.0
127.00	295744.0
128.00	22784.0
129.00	109600.0
129.90	8648.0
130.90	1757.0
132.10	865.0
132.90	739.0
133.90	3042.0
135.00	8230.0
135.90	3905.0
136.90	4727.0
138.10	890.0
138.90	976.0
139.90	1456.0

Instrument :
BNA_N
ClientSampleId :
DFTPP86

140.90	14640.0
142.00	4473.0
143.00	2389.0
143.80	825.0
145.00	779.0
146.00	2250.0
146.90	7958.0
148.00	15482.0
148.90	3192.0
149.90	1008.0
151.10	2131.0
151.70	1385.0
153.00	4699.0
154.00	3394.0
155.00	7941.0
156.00	12289.0
157.00	2526.0
157.90	2747.0
158.90	1761.0
159.90	4723.0
161.00	7816.0
162.00	1700.0
162.90	565.0
163.90	866.0
164.90	5504.0
166.00	3994.0
167.00	26968.0
168.00	16077.0
169.00	2447.0
170.00	1089.0
170.80	1234.0
171.90	2421.0
173.00	3723.0
174.00	6547.0
175.00	12605.0
176.00	4084.0
176.90	6261.0
178.00	1796.0
178.90	21928.0
180.00	16207.0
181.00	7155.0
182.00	919.0
182.70	651.0
183.90	1976.0
185.00	11485.0
186.00	87824.0
187.00	25536.0
188.00	2797.0
189.00	5352.0
189.90	890.0
191.00	2590.0
192.00	7072.0
193.00	8188.0
193.90	2025.0
195.00	926.0
196.00	18752.0
196.90	3269.0
197.90	710144.0
198.90	43456.0
200.00	3835.0
201.30	2760.0
203.00	4328.0
204.00	24112.0
205.00	40880.0
206.00	181376.0
207.00	24808.0
208.00	5845.0
209.00	2485.0
209.90	2590.0
210.90	6554.0
212.90	435.0
214.00	356.0
215.00	1698.0
216.00	3335.0
216.90	45248.0
217.90	4964.0
219.00	728.0
220.00	769.0
221.00	30824.0
221.80	6342.0
223.00	9292.0
224.00	97640.0
225.00	25936.0
226.00	2925.0
227.00	40128.0
228.00	5184.0
228.90	7915.0
230.00	937.0
231.00	2928.0
231.90	709.0
232.90	714.0
233.90	2345.0
234.90	3224.0
235.90	1403.0
237.00	3486.0

Instrument :
BNA_N
ClientSampleId :
DFTPP86

238.00	507.0
238.90	1884.0
239.90	1257.0
241.00	2248.0
242.00	5879.0
243.00	5529.0
244.00	77736.0
245.00	10154.0
246.00	14786.0
246.90	2697.0
248.00	830.0
248.90	2586.0
250.00	547.0
250.80	537.0
251.90	637.0
252.90	1414.0
254.00	2852.0
255.00	422784.0
256.00	59320.0
257.00	4036.0
257.90	22792.0
258.90	3178.0
259.90	599.0
261.10	714.0
263.90	521.0
264.90	9091.0
265.90	1234.0
269.90	766.0
270.90	993.0
272.00	1159.0
272.90	12991.0
274.00	33448.0
275.00	200512.0
276.00	25456.0
277.00	15211.0
277.90	2586.0
279.00	549.0
280.90	536.0
282.10	328.0
283.00	1535.0
284.00	973.0
285.00	3620.0
285.90	395.0
288.80	706.0
290.00	638.0
291.90	966.0
293.00	3471.0
294.00	769.0
295.00	876.0
296.00	52568.0
296.90	7159.0
297.90	566.0
300.90	787.0
302.10	684.0
303.00	5558.0
304.00	1999.0
305.00	437.0
307.90	796.0
309.00	706.0
309.90	952.0
313.10	464.0
314.00	2675.0
315.00	6285.0
316.00	3591.0
316.90	891.0
321.10	2036.0
322.00	1075.0
323.00	18152.0
324.00	3981.0
325.90	366.0
326.90	2959.0
327.90	1666.0
332.00	1158.0
332.90	1477.0
334.00	11205.0
335.00	3214.0
336.00	476.0
340.00	337.0
341.00	2238.0
341.80	451.0
346.00	4117.0
347.00	788.0
352.00	5298.0
353.00	3742.0
354.00	5716.0
355.00	1163.0
358.80	423.0
365.00	24072.0
366.00	3474.0
370.10	461.0
370.90	1819.0
372.00	11148.0
373.00	2244.0
373.90	390.0
382.90	2804.0

Instrument :
BNA_N
ClientSampleId :
DFTPP86

383.90	1012.0
384.90	351.0
390.00	1558.0
390.90	914.0
392.10	637.0
401.00	899.0
402.00	4136.0
403.00	6174.0
404.00	2520.0
404.90	478.0
415.10	359.0
421.00	5335.0
422.00	6963.0
423.00	37848.0
424.00	7973.0
425.00	999.0
441.00	123408.0
442.00	822848.0
443.00	152896.0
444.00	14428.0
444.90	726.0

Instrument :
BNA_N
ClientSampleId :
DFTPP86

Scan 2467 (17.498 min): BN001958.D
DFTPP86

m/z	Abundance
38.90	390.0
39.90	2230.0
40.90	311.0
43.10	404.0
44.00	15685.0
55.00	424.0
72.90	464.0
191.00	336.0
207.00	2030.0
207.90	527.0
209.00	607.0
281.00	706.0

Instrument :
BNA_N
ClientSampleId :
DFTPP86