



Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN070519\
Data File : BN006704.D
Acq On : 05 Jul 2019 08:29
Operator : HP/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 08 01:04:41 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN032619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 27 05:27:09 2019
Response via : Initial Calibration



TIC: BN006704.D

(77) Benzidine



19.280min (-0.053) 0.00ng

response 1755808

Ion	Exp%	Act%
184.00	100	100
185.00	14.80	13.85
183.00	12.40	11.81
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN070519\
Data File : BN006704.D
Acq On : 05 Jul 2019 08:29
Operator : HP/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 08 01:04:41 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN032619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 27 05:27:09 2019
Response via : Initial Calibration



TIC: BN006704.D

(70) Pentachlorophenol (C)

16.669min (-0.059) 27609.67ng

response 275053

Ion	Exp%	Act%
265.70	100	100
268.00	63.50	65.35
264.00	62.50	62.18
0.00	0.00	0.00