

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN070519\
 Data File : BN006705.D
 Acq On : 05 Jul 2019 09:03
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 05 10:44:37 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 05 05:30:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	-0.03
2	1,4-Dioxane	0.625	0.667	-6.7	82	-0.04
3	n-Nitrosodimethylamine	0.471	0.446	5.3	78	-0.02
4 S	2-Fluorophenol	1.086	0.985	9.3	72	-0.02
5 S	Phenol-d6	1.375	1.258	8.5	74	-0.02
6	bis(2-Chloroethyl)ether	1.162	1.296	-11.5	96	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	82	-0.03
8 S	Nitrobenzene-d5	0.307	0.272	11.4	76	-0.03
9	Naphthalene	1.094	1.077	1.6	77	-0.03
10	Hexachlorobutadiene	0.222	0.228	-2.7	77	-0.03
11 SURR	2-Methylnaphthalene-d10	0.601	0.610	-1.5	80	-0.02
12	2-Methylnaphthalene	0.812	0.714	12.1	76	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	-0.02
14 S	2,4,6-Tribromophenol	0.197	0.211	-7.1	110	-0.02
15 S	2-Fluorobiphenyl	1.565	1.398	10.7	83	-0.02
16	Acenaphthylene	1.994	1.908	4.3	90	-0.02
17	Acenaphthene	1.271	1.231	3.1	87	-0.02
18	Fluorene	1.578	1.579	-0.1	91	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	107	-0.02
20	4-Bromophenyl-phenylether	0.236	0.216	8.5	97	-0.02
21	Hexachlorobenzene	0.280	0.276	1.4	99	-0.02
22	Pentachlorophenol	0.040	0.042	-5.0	164#	-0.04
23	Phenanthrene	1.153	1.068	7.4	97	-0.02
24	Anthracene	1.042	0.956	8.3	97	-0.02
25 SURR	Fluoranthene-d10	1.144	1.198	-4.7	106	-0.02
26	Fluoranthene	1.354	1.382	-2.1	103	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	121	-0.02
28	Pvrene	1.601	1.433	10.5	103	-0.02
29 S	Terphenyl-d14	0.921	0.878	4.7	110	-0.02
30	Benzo(a)anthracene	1.380	1.279	7.3	104	-0.03
31	Chrysene	1.422	1.358	4.5	110	-0.02
32	Bis(2-ethylhexyl)phthalate	4.115	0.860	79.1#	40#	-0.02
33	Indeno(1,2,3-cd)pyrene	1.701	1.768	-3.9	115	-0.05
34 I	Perylene-d12	1.000	1.000	0.0	126	-0.04
35	Benzo(b)fluoranthene	1.284	1.174	8.6	108	-0.03
36	Benzo(k)fluoranthene	1.392	1.323	5.0	113	-0.03
37 C	Benzo(a)pyrene	1.244	1.180	5.1	110	-0.04
38	Dibenzo(a,h)anthracene	1.236	1.185	4.1	113	-0.05
39	Benzo(a,h,i)perylene	1.272	1.246	2.0	116	-0.06

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN070519\
Data File : BN006705.D
Acq On : 05 Jul 2019 09:03
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 05 10:44:37 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 05 05:30:28 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			