

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN070519\
 Data File : BN006729.D
 Acq On : 05 Jul 2019 22:54
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 06 00:01:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 05 05:30:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	79	-0.02
2	1,4-Dioxane	0.400	0.427	-6.7	84	-0.04
3	n-Nitrosodimethylamine	0.400	0.362	9.5	77	-0.02
4 S	2-Fluorophenol	0.400	0.361	9.8	73	-0.02
5 S	Phenol-d6	0.400	0.371	7.3	77	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.427	-6.7	94	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	85	-0.03
8 S	Nitrobenzene-d5	0.400	0.351	12.3	78	-0.03
9	Naphthalene	0.400	0.390	2.5	80	-0.03
10	Hexachlorobutadiene	0.400	0.404	-1.0	79	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.398	0.5	82	-0.02
12	2-Methylnaphthalene	0.400	0.356	11.0	77	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	93	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.382	4.5	95	-0.02
15 S	2-Fluorobiphenyl	0.400	0.370	7.5	83	-0.02
16	Acenaphthylene	0.400	0.382	4.5	87	-0.02
17	Acenaphthene	0.400	0.390	2.5	85	-0.03
18	Fluorene	0.400	0.391	2.3	87	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	100	-0.02
20	4-Bromophenyl-phenylether	0.400	0.362	9.5	90	-0.02
21	Hexachlorobenzene	0.400	0.395	1.3	93	-0.02
22	Pentachlorophenol	0.400	0.542	-35.5#	152	-0.05
23	Phenanthrene	0.400	0.372	7.0	91	-0.02
24	Anthracene	0.400	0.355	11.3	88	-0.02
25 SURR	Fluoranthene-d10	0.400	0.406	-1.5	96	-0.02
26	Fluoranthene	0.400	0.395	1.3	93	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	112	-0.03
28	Pvrene	0.400	0.348	13.0	93	-0.02
29 S	Terphenyl-d14	0.400	0.374	6.5	101	-0.02
30	Benzo(a)anthracene	0.400	0.362	9.5	94	-0.03
31	Chrysene	0.400	0.388	3.0	104	-0.03
32	Bis(2-ethylhexyl)phthalate	0.400	0.067	83.3#	30	-0.03
33	Indeno(1,2,3-cd)pyrene	0.400	0.420	-5.0	108	-0.08
34 I	Perylene-d12	0.400	0.400	0.0	119	-0.06
35	Benzo(b)fluoranthene	0.400	0.375	6.3	104	-0.05
36	Benzo(k)fluoranthene	0.400	0.370	7.5	104	-0.05
37 C	Benzo(a)pyrene	0.400	0.383	4.3	104	-0.05
38	Dibenzo(a,h)anthracene	0.400	0.390	2.5	108	-0.07
39	Benzo(a,h,i)perylene	0.400	0.388	3.0	108	-0.08

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN070519\
Data File : BN006729.D
Acq On : 05 Jul 2019 22:54
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 06 00:01:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 05 05:30:28 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				