

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN070519\
 Data File : BN006729.D
 Acq On : 05 Jul 2019 22:54
 Operator : HP/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Jul 06 00:01:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 05 05:30:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	-0.02
2	1,4-Dioxane	0.625	0.668	-6.9	84	-0.04
3	n-Nitrosodimethylamine	0.471	0.427	9.3	77	-0.02
4 S	2-Fluorophenol	1.086	0.981	9.7	73	-0.02
5 S	Phenol-d6	1.375	1.274	7.3	77	-0.02
6	bis(2-Chloroethyl)ether	1.162	1.241	-6.8	94	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	85	-0.03
8 S	Nitrobenzene-d5	0.307	0.269	12.4	78	-0.03
9	Naphthalene	1.094	1.068	2.4	80	-0.03
10	Hexachlorobutadiene	0.222	0.224	-0.9	79	-0.03
11 SURR	2-Methylnaphthalene-d10	0.601	0.598	0.5	82	-0.02
12	2-Methylnaphthalene	0.812	0.694	14.5	77	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	93	-0.02
14 S	2,4,6-Tribromophenol	0.197	0.188	4.6	95	-0.02
15 S	2-Fluorobiphenyl	1.565	1.449	7.4	83	-0.02
16	Acenaphthylene	1.994	1.903	4.6	87	-0.02
17	Acenaphthene	1.271	1.238	2.6	85	-0.03
18	Fluorene	1.578	1.544	2.2	87	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	-0.02
20	4-Bromophenyl-phenylether	0.236	0.213	9.7	90	-0.02
21	Hexachlorobenzene	0.280	0.277	1.1	93	-0.02
22	Pentachlorophenol	0.040	0.042	-5.0	152#	-0.05
23	Phenanthrene	1.153	1.073	6.9	91	-0.02
24	Anthracene	1.042	0.925	11.2	88	-0.02
25 SURR	Fluoranthene-d10	1.144	1.162	-1.6	96	-0.02
26	Fluoranthene	1.354	1.338	1.2	93	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	112	-0.03
28	Pvrene	1.601	1.393	13.0	93	-0.02
29 S	Terphenyl-d14	0.921	0.861	6.5	101	-0.02
30	Benzo(a)anthracene	1.380	1.247	9.6	94	-0.03
31	Chrysene	1.422	1.379	3.0	104	-0.03
32	Bis(2-ethylhexyl)phthalate	4.115	0.687	83.3#	30#	-0.03
33	Indeno(1,2,3-cd)pyrene	1.701	1.787	-5.1	108	-0.08
34 I	Perylene-d12	1.000	1.000	0.0	119	-0.06
35	Benzo(b)fluoranthene	1.284	1.204	6.2	104	-0.05
36	Benzo(k)fluoranthene	1.392	1.289	7.4	104	-0.05
37 C	Benzo(a)pyrene	1.244	1.192	4.2	104	-0.05
38	Dibenzo(a,h)anthracene	1.236	1.204	2.6	108	-0.07
39	Benzo(a,h,i)perylene	1.272	1.234	3.0	108	-0.08

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN070519\
Data File : BN006729.D
Acq On : 05 Jul 2019 22:54
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 06 00:01:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 05 05:30:28 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			