

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070519\
 Data File : BN006731.D
 Acq On : 06 Jul 2019 00:38
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 06 01:23:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 05 05:30:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	-0.02
2	1,4-Dioxane	0.625	0.682	-9.1	84	-0.04
3	n-Nitrosodimethylamine	0.471	0.404	14.2	71	-0.02
4 S	2-Fluorophenol	1.086	0.936	13.8	69	0.00
5 S	Phenol-d6	1.375	1.258	8.5	74	-0.02
6	bis(2-Chloroethyl)ether	1.162	1.213	-4.4	90	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	83	-0.03
8 S	Nitrobenzene-d5	0.307	0.266	13.4	75	-0.01
9	Naphthalene	1.094	1.082	1.1	79	-0.03
10	Hexachlorobutadiene	0.222	0.228	-2.7	78	-0.03
11 SURR	2-Methylnaphthalene-d10	0.601	0.592	1.5	79	-0.02
12	2-Methylnaphthalene	0.812	0.703	13.4	76	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	91	-0.02
14 S	2,4,6-Tribromophenol	0.197	0.178	9.6	88	-0.02
15 S	2-Fluorobiphenyl	1.565	1.473	5.9	83	-0.02
16	Acenaphthylene	1.994	1.865	6.5	84	-0.02
17	Acenaphthene	1.271	1.248	1.8	84	-0.02
18	Fluorene	1.578	1.518	3.8	84	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	-0.02
20	4-Bromophenyl-phenylether	0.236	0.222	5.9	89	-0.02
21	Hexachlorobenzene	0.280	0.286	-2.1	92	-0.02
22	Pentachlorophenol	0.040	0.038	5.0	132	-0.05
23	Phenanthrene	1.153	1.093	5.2	88	-0.02
24	Anthracene	1.042	0.933	10.5	85	-0.02
25 SURR	Fluoranthene-d10	1.144	1.150	-0.5	91	-0.02
26	Fluoranthene	1.354	1.350	0.3	90	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	107	-0.02
28	Pvrene	1.601	1.400	12.6	90	-0.02
29 S	Terphenyl-d14	0.921	0.884	4.0	99	-0.02
30	Benzo(a)anthracene	1.380	1.287	6.7	93	-0.03
31	Chrysene	1.422	1.347	5.3	97	-0.03
32	Bis(2-ethylhexyl)phthalate	4.115	0.704	82.9#	29#	-0.02
33	Indeno(1,2,3-cd)pyrene	1.701	1.798	-5.7	104	-0.06
34 I	Perylene-d12	1.000	1.000	0.0	113	-0.05
35	Benzo(b)fluoranthene	1.284	1.204	6.2	99	-0.04
36	Benzo(k)fluoranthene	1.392	1.344	3.4	102	-0.04
37 C	Benzo(a)pyrene	1.244	1.213	2.5	101	-0.05
38	Dibenzo(a,h)anthracene	1.236	1.200	2.9	102	-0.06
39	Benzo(a,h,i)perylene	1.272	1.255	1.3	104	-0.07

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070519\
Data File : BN006731.D
Acq On : 06 Jul 2019 00:38
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 06 01:23:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 05 05:30:28 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			