

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070519\
 Data File : BN006731.D
 Acq On : 06 Jul 2019 00:38
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 06 01:23:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 05 05:30:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	77	-0.02
2	1,4-Dioxane	0.400	0.437	-9.2	84	-0.04
3	n-Nitrosodimethylamine	0.400	0.343	14.2	71	-0.02
4 S	2-Fluorophenol	0.400	0.345	13.8	69	0.00
5 S	Phenol-d6	0.400	0.366	8.5	74	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.417	-4.2	90	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	83	-0.03
8 S	Nitrobenzene-d5	0.400	0.347	13.3	75	-0.01
9	Naphthalene	0.400	0.395	1.3	79	-0.03
10	Hexachlorobutadiene	0.400	0.411	-2.7	78	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.394	1.5	79	-0.02
12	2-Methylnaphthalene	0.400	0.361	9.8	76	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	91	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.362	9.5	88	-0.02
15 S	2-Fluorobiphenyl	0.400	0.377	5.8	83	-0.02
16	Acenaphthylene	0.400	0.374	6.5	84	-0.02
17	Acenaphthene	0.400	0.393	1.8	84	-0.02
18	Fluorene	0.400	0.385	3.8	84	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	95	-0.02
20	4-Bromophenyl-phenylether	0.400	0.376	6.0	89	-0.02
21	Hexachlorobenzene	0.400	0.408	-2.0	92	-0.02
22	Pentachlorophenol	0.400	0.507	-26.7#	132	-0.05
23	Phenanthrene	0.400	0.379	5.3	88	-0.02
24	Anthracene	0.400	0.358	10.5	85	-0.02
25 SURR	Fluoranthene-d10	0.400	0.402	-0.5	91	-0.02
26	Fluoranthene	0.400	0.399	0.3	90	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	107	-0.02
28	Pvrene	0.400	0.350	12.5	90	-0.02
29 S	Terphenyl-d14	0.400	0.384	4.0	99	-0.02
30	Benzo(a)anthracene	0.400	0.373	6.8	93	-0.03
31	Chrysene	0.400	0.379	5.3	97	-0.03
32	Bis(2-ethylhexyl)phthalate	0.400	0.068	83.0#	29	-0.02
33	Indeno(1,2,3-cd)pyrene	0.400	0.423	-5.7	104	-0.06
34 I	Perylene-d12	0.400	0.400	0.0	113	-0.05
35	Benzo(b)fluoranthene	0.400	0.375	6.3	99	-0.04
36	Benzo(k)fluoranthene	0.400	0.386	3.5	102	-0.04
37 C	Benzo(a)pyrene	0.400	0.390	2.5	101	-0.05
38	Dibenzo(a,h)anthracene	0.400	0.389	2.8	102	-0.06
39	Benzo(a,h,i)perylene	0.400	0.395	1.3	104	-0.07

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070519\
Data File : BN006731.D
Acq On : 06 Jul 2019 00:38
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 06 01:23:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 05 05:30:28 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				