

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070519\
 Data File : BN006746.D
 Acq On : 06 Jul 2019 09:21
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 08 00:25:26 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 05 05:30:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	77	-0.02
2	1,4-Dioxane	0.400	0.447	-11.7	86	-0.04
3	n-Nitrosodimethylamine	0.400	0.390	2.5	81	-0.02
4 S	2-Fluorophenol	0.400	0.347	13.3	69	0.00
5 S	Phenol-d6	0.400	0.358	10.5	72	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.435	-8.7	94	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	80	-0.03
8 S	Nitrobenzene-d5	0.400	0.351	12.3	74	-0.01
9	Naphthalene	0.400	0.412	-3.0	80	-0.03
10	Hexachlorobutadiene	0.400	0.435	-8.7	80	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.407	-1.7	79	-0.02
12	2-Methylnaphthalene	0.400	0.367	8.3	75	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	86	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.367	8.3	85	-0.02
15 S	2-Fluorobiphenyl	0.400	0.390	2.5	81	-0.02
16	Acenaphthylene	0.400	0.380	5.0	80	-0.02
17	Acenaphthene	0.400	0.409	-2.2	83	-0.02
18	Fluorene	0.400	0.399	0.3	82	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	90	-0.02
20	4-Bromophenyl-phenylether	0.400	0.383	4.3	86	-0.02
21	Hexachlorobenzene	0.400	0.430	-7.5	91	-0.02
22	Pentachlorophenol	0.400	0.504	-26.0#	124	-0.05
23	Phenanthrene	0.400	0.389	2.8	85	-0.02
24	Anthracene	0.400	0.363	9.3	81	-0.02
25 SURR	Fluoranthene-d10	0.400	0.424	-6.0	90	-0.02
26	Fluoranthene	0.400	0.417	-4.2	88	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	102	-0.02
28	Pvrene	0.400	0.362	9.5	88	-0.02
29 S	Terphenyl-d14	0.400	0.394	1.5	96	-0.02
30	Benzo(a)anthracene	0.400	0.378	5.5	89	-0.03
31	Chrysene	0.400	0.397	0.8	97	-0.03
32	Bis(2-ethylhexyl)phthalate	0.400	0.064	84.0#	26	-0.03
33	Indeno(1,2,3-cd)pyrene	0.400	0.433	-8.2	101	-0.06
34 I	Perylene-d12	0.400	0.400	0.0	105	-0.05
35	Benzo(b)fluoranthene	0.400	0.395	1.3	97	-0.04
36	Benzo(k)fluoranthene	0.400	0.413	-3.2	102	-0.04
37 C	Benzo(a)pyrene	0.400	0.406	-1.5	98	-0.05
38	Dibenzo(a,h)anthracene	0.400	0.408	-2.0	100	-0.06
39	Benzo(a,h,i)perylene	0.400	0.416	-4.0	102	-0.06

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070519\
Data File : BN006746.D
Acq On : 06 Jul 2019 09:21
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 08 00:25:26 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 05 05:30:28 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			