

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070523\
 Data File : BN026315.D
 Acq On : 06 Jul 2023 05:11
 Operator : MA/JU
 Sample : 03248-14DL 5X
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DCGP2DL

Quant Time: Jul 06 05:49:26 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 05 00:03:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	4	0.400	ng/ul	#-0.01
4) Naphthalene-d8	10.748	136	45	0.400	ng/ul	#-0.01
9) Acenaphthene-d10	14.580	164	118	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.335	188	186	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.533	240	127	0.400	ng/ul	# 0.00
23) Perylene-d12	24.041	264	17	0.400	ng/ul	# 0.04
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.310	96	4273	948.572	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.343	152	1490	21.527	ng/ul	0.00
18) Fluoranthene-d10	19.364	212	2925	6.051	ng/ul	0.00
Target Compounds						
						Qvalue
5) Naphthalene	10.798	128	901	6.986	ng/ul#	68
7) 2-Methylnaphthalene	12.415	142	600	7.306	ng/ul	96
8) 1-Methylnaphthalene	12.629	142	459	5.394	ng/ul	99
10) Acenaphthylene	14.303	152	1555	2.814	ng/ul#	75
11) Acenaphthene	14.645	153	375	0.830	ng/ul#	89
12) Fluorene	15.635	166	384	0.741	ng/ul#	79
14) Pentachlorophenol	17.006	266	10	0.150	ng/ul#	56
15) Phenanthrene	17.373	178	7799	12.694	ng/ul#	93
16) Anthracene	17.466	178	1122	2.055	ng/ul#	64
19) Fluoranthene	19.392	202	22651	33.392	ng/ul	99
20) Pyrene	19.755	202	19477	27.606	ng/ul	99
21) Benzo(a)anthracene	21.509	228	7679	14.657	ng/ul#	71
22) Chrysene	21.565	228	9620	17.647	ng/ul#	79
24) Benzo(b)fluoranthene	23.231	252	23132	318.803	ng/ul	68
25) Benzo(k)fluoranthene	23.418	252	1986	27.257	ng/ul#	1
26) Benzo(a)pyrene	23.871	252	9995	158.600	ng/ul#	60
27) Indeno(1,2,3-cd)pyrene	26.535	276	8787	127.654	ng/ul#	91
28) Dibenzo(a,h)anthracene	26.666	278	561	10.769	ng/ul#	1
29) Benzo(g,h,i)perylene	27.333	276	7690	125.586	ng/ul#	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070523\
Data File : BN026315.D
Acq On : 06 Jul 2023 05:11
Operator : MA/JU
Sample : 03248-14DL 5X
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DCGP2DL

Quant Time: Jul 06 05:49:26 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 05 00:03:30 2023
Response via : Initial Calibration

