

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070523\
 Data File : BN026336.D
 Acq On : 06 Jul 2023 18:49
 Operator : MA/JU
 Sample : 03257-07
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EXYY3

Quant Time: Jul 07 00:58:00 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 06 17:50:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.938	152	5202	0.400	ng/ul	0.00
4) Naphthalene-d8	10.743	136	17216	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.579	164	13361	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.330	188	25621	0.400	ng/ul	0.00
17) Chrysene-d12	21.532	240	13523	0.400	ng/ul	0.00
23) Perylene-d12	23.976	264	18765	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.306	96	32016	5.465	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.338	152	9088	0.343	ng/ul	0.00
18) Fluoranthene-d10	19.363	212	23722	0.461	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070523\
Data File : BN026336.D
Acq On : 06 Jul 2023 18:49
Operator : MA/JU
Sample : 03257-07
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EXYY3

Quant Time: Jul 07 00:58:00 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 06 17:50:03 2023
Response via : Initial Calibration

