

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070523\
 Data File : BN026348.D
 Acq On : 07 Jul 2023 02:17
 Operator : MA/JU
 Sample : PB153517BS
 Misc :
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS517

Quant Time: Jul 07 02:51:31 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 06 17:50:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.938	152	5978	0.400	ng/ul	0.00
4) Naphthalene-d8	10.748	136	20077	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.580	164	11030	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.334	188	17806	0.400	ng/ul	0.00
17) Chrysene-d12	21.523	240	11165	0.400	ng/ul	0.00
23) Perylene-d12	23.973	264	12165	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.306	96	4454	0.662	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.338	152	9599	0.311	ng/ul	0.00
18) Fluoranthene-d10	19.363	212	14434	0.340	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.344	88	10692	1.496	ng/ul#	85
5) Naphthalene	10.798	128	19159	0.333	ng/ul	100
7) 2-Methylnaphthalene	12.409	142	11096	0.303	ng/ul	100
8) 1-Methylnaphthalene	12.629	142	11904	0.314	ng/ul	100
10) Acenaphthylene	14.302	152	17682	0.342	ng/ul	99
11) Acenaphthene	14.640	153	13799	0.327	ng/ul	99
12) Fluorene	15.634	166	14184	0.293	ng/ul	98
14) Pentachlorophenol	16.997	266	2477	0.388	ng/ul	99
15) Phenanthrene	17.377	178	18861	0.321	ng/ul	98
16) Anthracene	17.469	178	16392	0.314	ng/ul	98
19) Fluoranthene	19.391	202	19492	0.327	ng/ul	98
20) Pyrene	19.754	202	20521	0.331	ng/ul	96
21) Benzo(a)anthracene	21.509	228	14877	0.323	ng/ul	99
22) Chrysene	21.561	228	15928	0.332	ng/ul	99
24) Benzo(b)fluoranthene	23.221	252	16780	0.323	ng/ul	95
25) Benzo(k)fluoranthene	23.271	252	17196	0.330	ng/ul	97
26) Benzo(a)pyrene	23.862	252	15169	0.336	ng/ul	94
27) Indeno(1,2,3-cd)pyrene	26.527	276	20026	0.407	ng/ul#	95
28) Dibenzo(a,h)anthracene	26.544	278	15445	0.414	ng/ul	98
29) Benzo(g,h,i)perylene	27.312	276	17759	0.405	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070523\
Data File : BN026348.D
Acq On : 07 Jul 2023 02:17
Operator : MA/JU
Sample : PB153517BS
Misc :
ALS Vial : 67 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS517

Quant Time: Jul 07 02:51:31 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 06 17:50:03 2023
Response via : Initial Calibration

