

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070523\
 Data File : BN026349.D
 Acq On : 07 Jul 2023 03:31
 Operator : MA/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 69 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4339

Quant Time: Jul 07 04:01:38 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 06 17:50:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.938	152	6442	0.400	ng/ul	0.00
4) Naphthalene-d8	10.743	136	21710	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.579	164	11619	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.330	188	19141	0.400	ng/ul	0.00
17) Chrysene-d12	21.517	240	12326	0.400	ng/ul	-0.01
23) Perylene-d12	23.967	264	13725	0.400	ng/ul	-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.310	96	2578	0.355	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.337	152	11504	0.345	ng/ul	0.00
18) Fluoranthene-d10	19.359	212	16717	0.356	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.343	88	2537	0.329	ng/ul	94
5) Naphthalene	10.792	128	22825	0.367	ng/ul	99
7) 2-Methylnaphthalene	12.409	142	13290	0.335	ng/ul	100
8) 1-Methylnaphthalene	12.623	142	14274	0.348	ng/ul	100
10) Acenaphthylene	14.302	152	20963	0.385	ng/ul	100
11) Acenaphthene	14.639	153	16423	0.369	ng/ul	99
12) Fluorene	15.634	166	16785	0.329	ng/ul	98
14) Pentachlorophenol	16.992	266	2478	0.361	ng/ul	99
15) Phenanthrene	17.372	178	22501	0.356	ng/ul	100
16) Anthracene	17.469	178	19504	0.347	ng/ul	98
19) Fluoranthene	19.391	202	23178	0.352	ng/ul	97
20) Pyrene	19.754	202	24311	0.355	ng/ul	97
21) Benzo(a)anthracene	21.500	228	17572	0.346	ng/ul	99
22) Chrysene	21.552	228	18616	0.352	ng/ul	99
24) Benzo(b)fluoranthene	23.221	252	19127	0.327	ng/ul	98
25) Benzo(k)fluoranthene	23.268	252	20466	0.348	ng/ul	99
26) Benzo(a)pyrene	23.861	252	17791	0.350	ng/ul	99
27) Indeno(1,2,3-cd)pyrene	26.534	276	23321	0.420	ng/ul#	94
28) Dibenzo(a,h)anthracene	26.551	278	17823	0.424	ng/ul	98
29) Benzo(g,h,i)perylene	27.312	276	20469	0.414	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070523\
Data File : BN026349.D
Acq On : 07 Jul 2023 03:31
Operator : MA/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 69 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4339

Quant Time: Jul 07 04:01:38 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 06 17:50:03 2023
Response via : Initial Calibration

