

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070821\
 Data File : BN015422.D
 Acq On : 08 Jul 2021 23:28
 Operator : CG/JU
 Sample : M2905-11
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DBK34

Manual Integrations
 APPROVED

mohammad
 7/9/2021 3:04:09 PM

Quant Time: Jul 09 01:16:33 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 08 11:17:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	2992	0.400	ng/ul	0.00
4) Naphthalene-d8	10.485	136	11436	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.340	164	9005	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.090	188	14116m	0.400	ng/ul	0.00
17) Chrysene-d12	21.284	240	10297	0.400	ng/ul	# 0.00
23) Perylene-d12	23.523	264	10371	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	16861	5.000	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.074	152	7634	0.425	ng/ul	0.00
18) Fluoranthene-d10	19.122	212	14049	0.415	ng/ul	0.00
Target Compounds				Qvalue		
2) 1,4-Dioxane	3.311	88	22387	6.143	ng/ul#	84
5) Naphthalene	10.534	128	27937	0.766	ng/ul#	94
7) 2-Methylnaphthalene	12.146	142	8242	0.345	ng/ul#	1
8) 1-Methylnaphthalene	12.371	142	7643	0.331	ng/ul	87
12) Fluorene	15.390	166	4222	0.107	ng/ul#	77
15) Phenanthrene	17.103	178	8386	0.164	ng/ul#	35
16) Anthracene	17.221	178	2140	0.047	ng/ul#	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070821\
Data File : BN015422.D
Acq On : 08 Jul 2021 23:28
Operator : CG/JU
Sample : M2905-11
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DBK34

Manual Integrations
APPROVED

mohammad
7/9/2021 3:04:09 PM

Quant Time: Jul 09 01:16:33 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 08 11:17:31 2021
Response via : Initial Calibration

