


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070821\
 Data File : BN015400.D
 Acq On : 08 Jul 2021 09:02
 Operator : CG/JU
 Sample : DFTPP154
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP154

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062121.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 08 11:17:31 2021



AutoFind: Scans 2456, 2457, 2458; Background Corrected with Scan 2449

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.4	92168	PASS
68	69	0.00	2	1.7	1642	PASS
69	198	0.00	100	35.5	97939	PASS
70	69	0.00	2	0.5	469	PASS
127	198	10	80	46.9	129251	PASS
197	198	0.00	2	0.6	1714	PASS
198	198	100	100	100.0	275541	PASS
199	198	5	9	7.0	19224	PASS
275	198	10	60	24.7	68099	PASS
365	198	1	100	2.9	7872	PASS
441	198	0.01	100	13.6	37360	PASS
442	198	50	100	90.7	249835	PASS
443	442	15	24	19.6	49091	PASS

m/z	Abundance
39.90	509.0
44.00	1624.0
48.90	396.0
207.00	1115.0
208.00	350.0

Instrument :
BNA_N
ClientSampleId :
DFTPP154

m/z	Abundance
37.10	476.0
38.00	1002.0
39.00	6856.0
39.90	822.0
44.00	1210.0
49.00	883.0
50.00	24376.0
51.00	94184.0
52.00	4951.0
55.00	609.0
56.00	2948.0
57.00	6920.0
61.00	1559.0
62.00	1318.0
63.00	3783.0
64.00	556.0
65.00	2108.0
67.90	1736.0
69.00	98016.0
70.00	471.0
73.10	761.0
74.00	10380.0
75.00	15683.0
76.00	5834.0
77.00	116816.0
78.00	9191.0
79.00	7718.0
79.90	4921.0
81.00	8143.0
82.00	1908.0
83.00	2337.0
83.90	511.0
84.90	1366.0
86.00	2254.0
87.00	979.0
88.00	472.0
91.00	1829.0
91.90	2068.0
93.00	12436.0
94.00	898.0
96.00	790.0
98.00	9162.0
99.00	9143.0
99.90	858.0
101.00	4298.0
102.90	1373.0
103.90	3213.0
105.00	2676.0
106.00	886.0
107.00	35048.0
108.00	5303.0
109.00	1168.0
110.00	65928.0
111.00	10343.0
112.00	1192.0
112.90	368.0
116.00	1825.0
117.00	27264.0
118.00	2233.0
119.00	354.0
120.00	441.0
121.90	2273.0
123.00	3245.0
124.00	1736.0
125.00	1374.0
126.20	749.0
127.00	124696.0
128.00	9280.0
129.00	44696.0
130.00	4242.0
131.00	788.0
132.10	457.0
133.90	1311.0
135.00	3451.0
136.00	1489.0
137.00	2332.0
137.90	408.0
139.90	577.0
141.00	5410.0
142.00	1919.0
142.90	1411.0
144.90	361.0
145.90	908.0
147.00	2900.0
148.00	7232.0
148.90	1680.0
149.90	566.0
151.00	568.0
151.90	501.0
152.90	2044.0
153.90	1417.0

Instrument :
BNA_N
ClientSampleId :
DFTPP154

155.00	3217.0
156.00	4910.0
157.10	1164.0
158.00	1059.0
158.90	678.0
159.90	1545.0
160.90	2905.0
161.80	774.0
163.00	358.0
164.90	2021.0
166.00	1434.0
167.00	11550.0
168.00	8723.0
168.90	1264.0
169.90	327.0
171.00	572.0
171.90	1140.0
173.00	1384.0
174.00	2583.0
175.00	4829.0
175.90	1297.0
177.00	1987.0
178.00	908.0
178.90	8383.0
180.00	6791.0
181.00	2853.0
182.00	634.0
182.90	309.0
184.00	632.0
185.00	3730.0
186.00	34064.0
187.00	10557.0
188.00	1253.0
188.90	2176.0
189.90	309.0
190.90	1228.0
192.00	2950.0
193.00	3045.0
194.00	641.0
195.00	397.0
196.00	7118.0
197.00	1580.0
197.90	244352.0
199.00	17520.0
200.00	1283.0
201.40	958.0
203.00	1701.0
204.00	8485.0
205.00	13900.0
206.00	61760.0
207.00	9969.0
208.00	2141.0
208.90	765.0
211.00	2265.0
214.90	692.0
216.00	1248.0
216.90	14957.0
218.00	1844.0
219.90	338.0
221.00	10621.0
221.80	2081.0
223.00	3332.0
224.00	32968.0
225.00	8708.0
226.00	1143.0
227.00	13281.0
227.90	2259.0
228.90	2711.0
229.90	415.0
231.00	990.0
234.00	991.0
235.00	1033.0
235.90	582.0
237.00	1631.0
238.90	636.0
240.00	503.0
240.90	732.0
242.00	1414.0
243.00	2008.0
244.00	27264.0
245.00	3790.0
246.00	4732.0
247.00	723.0
248.90	850.0
252.90	588.0
255.00	127216.0
256.00	19616.0
256.90	1627.0
258.00	6443.0
259.00	1242.0
261.00	354.0
265.00	2858.0
265.90	462.0
273.00	3822.0
274.00	10573.0

Instrument :

BNA_N

ClientSampleId :

DFTPP154

275.00	54040.0
276.00	8290.0
277.00	4668.0
277.90	668.0
282.90	476.0
283.90	352.0
285.00	753.0
292.90	978.0
294.00	413.0
295.00	395.0
296.00	14975.0
297.00	2370.0
302.00	397.0
303.00	1607.0
303.90	369.0
313.90	653.0
315.00	1555.0
315.90	911.0
321.00	669.0
323.00	4649.0
324.00	748.0
327.00	807.0
327.90	421.0
333.00	454.0
334.10	2768.0
335.00	688.0
341.00	498.0
346.00	1064.0
352.00	1353.0
353.00	1012.0
354.00	1721.0
355.00	376.0
364.90	5261.0
365.90	1207.0
372.00	2454.0
373.10	408.0
382.90	706.0
389.90	317.0
402.00	921.0
403.00	1310.0
403.90	558.0
421.00	1506.0
422.00	1180.0
423.00	9951.0
424.00	3558.0
425.00	656.0
441.00	24696.0
441.90	161024.0
443.00	32704.0
444.00	3387.0

Instrument :
BNA_N
ClientSampleId :
DFTPP154

m/z	Abundance
37.00	505.0
38.00	1167.0
39.00	7449.0
40.00	791.0
41.00	341.0
44.00	1492.0
49.10	964.0
50.00	26584.0
51.00	105328.0
52.10	5867.0
55.00	621.0
56.00	3004.0
57.00	7244.0
58.00	455.0
61.00	1441.0
62.00	1625.0
63.00	4391.0
64.00	524.0
65.00	2532.0
68.00	1778.0
69.00	112712.0
70.00	517.0
72.90	1073.0
74.00	11631.0
75.00	16912.0
76.00	6628.0
77.00	134400.0
78.00	9340.0
79.00	8857.0
80.00	6312.0
81.00	9555.0
82.00	2543.0
83.00	2211.0
83.90	514.0
85.00	1751.0
86.00	2466.0
87.00	1496.0
87.90	474.0
91.00	2099.0
92.00	2303.0
93.00	14062.0
94.00	1034.0
96.00	748.0
98.00	11016.0
99.00	10007.0
100.00	952.0
101.00	5364.0
102.00	345.0
103.00	2021.0
104.00	3404.0
105.00	3136.0
106.00	865.0
107.00	41160.0
108.00	5941.0
109.00	1488.0
110.00	80512.0
111.00	12065.0
112.00	1354.0
113.00	503.0
116.00	2183.0
117.00	31624.0
118.00	2664.0
118.90	348.0
120.00	445.0
122.00	2660.0
123.00	4517.0
124.00	2089.0
125.00	2026.0
126.20	848.0
127.00	147968.0
128.00	10986.0
129.00	56112.0
130.00	4998.0
130.90	1005.0
132.10	469.0
134.00	1559.0
135.00	4289.0
136.00	1750.0
137.00	2674.0
137.90	542.0
138.80	318.0
140.00	774.0
141.00	7341.0
141.90	2499.0
142.90	1422.0
144.00	340.0
144.90	447.0
146.00	1108.0
147.00	3976.0
148.00	8593.0
149.00	1883.0

Instrument :
BNA_N
ClientSampleId :
DFTPP154

149.90	530.0
151.10	1035.0
152.00	618.0
153.00	2125.0
154.00	1653.0
155.00	4380.0
156.00	6064.0
157.00	1304.0
157.90	1270.0
159.00	1154.0
160.00	2357.0
161.00	3702.0
161.90	1000.0
163.10	338.0
165.00	2472.0
166.00	2351.0
167.00	14953.0
168.00	9449.0
168.90	1547.0
170.00	379.0
170.90	618.0
171.90	1357.0
173.00	1684.0
174.00	2945.0
175.00	6045.0
176.00	1801.0
177.00	3045.0
177.90	793.0
179.00	10749.0
180.00	8582.0
181.00	3235.0
182.00	611.0
182.90	312.0
183.90	801.0
185.00	5242.0
186.00	42016.0
187.00	12331.0
188.00	1322.0
188.90	2338.0
189.90	564.0
191.00	1131.0
192.00	3538.0
193.00	4013.0
194.00	862.0
195.10	454.0
196.00	8795.0
196.70	1724.0
197.00	1922.0
197.90	322432.0
199.00	22136.0
200.00	1937.0
201.40	1750.0
203.00	2129.0
204.00	11371.0
205.00	18744.0
206.00	86456.0
207.00	10683.0
208.00	2711.0
209.00	1065.0
210.00	1254.0
210.90	2758.0
214.90	951.0
216.00	1902.0
217.00	22408.0
218.00	2662.0
221.00	13706.0
221.80	3405.0
223.00	4418.0
224.00	44648.0
225.00	12197.0
226.00	1352.0
227.00	17512.0
227.90	2433.0
229.00	4120.0
230.00	721.0
231.10	1849.0
231.90	361.0
233.00	332.0
234.00	1155.0
234.90	1457.0
235.90	877.0
236.90	1439.0
239.00	789.0
240.00	432.0
240.90	1008.0
242.00	2737.0
243.00	2408.0
244.00	35976.0
245.00	4829.0
245.90	6491.0
246.90	1247.0
248.90	1257.0
251.90	376.0
252.90	655.0
254.10	1441.0

Instrument :
BNA_N
ClientSampleId :
DFTPP154

255.00	167872.0
256.00	26680.0
257.00	2072.0
258.00	9596.0
259.00	1579.0
259.90	308.0
264.90	3685.0
266.00	644.0
270.90	437.0
272.00	460.0
273.00	5672.0
274.00	14171.0
275.00	81080.0
276.00	11416.0
277.00	6426.0
277.90	1035.0
281.90	305.0
283.00	612.0
283.90	387.0
285.00	1023.0
288.90	373.0
292.90	1685.0
294.00	496.0
294.90	343.0
296.00	20808.0
297.00	3194.0
302.00	431.0
303.00	2306.0
304.00	792.0
307.90	362.0
309.80	391.0
314.00	1075.0
314.90	2114.0
316.00	1328.0
321.00	593.0
322.00	479.0
323.00	6181.0
324.00	1433.0
327.00	1054.0
328.00	588.0
332.00	525.0
333.00	622.0
334.00	5051.0
334.90	1027.0
341.00	746.0
345.90	1844.0
346.90	357.0
352.00	2164.0
353.00	1505.0
354.00	2125.0
355.10	472.0
364.90	9929.0
365.90	1307.0
371.00	677.0
372.00	3946.0
373.00	963.0
382.90	1228.0
390.00	434.0
391.00	387.0
402.00	1740.0
403.00	2467.0
404.00	784.0
420.90	2163.0
422.00	2005.0
423.00	15478.0
424.00	4539.0
425.00	463.0
441.00	44944.0
442.00	272896.0
443.00	59976.0
444.00	5271.0

Instrument :
BNA_N
ClientSampleId :
DFTPP154

m/z	Abundance
37.00	364.0
38.00	867.0
39.10	5919.0
40.00	640.0
44.00	1350.0
48.90	797.0
50.00	18352.0
51.00	76992.0
52.00	4147.0
55.00	545.0
56.00	2311.0
57.00	5072.0
61.00	1077.0
62.00	1178.0
63.00	3121.0
63.90	427.0
65.00	1590.0
68.00	1413.0
69.00	83088.0
70.00	418.0
72.90	952.0
74.00	8054.0
75.00	13341.0
76.00	4147.0
77.00	98648.0
78.00	6721.0
79.00	5866.0
80.00	4932.0
81.00	6484.0
82.00	1837.0
83.00	1803.0
84.00	410.0
85.00	1412.0
86.00	2174.0
87.00	823.0
87.90	379.0
91.00	1643.0
92.00	1589.0
93.00	10874.0
93.90	627.0
96.00	684.0
98.00	8397.0
99.00	7177.0
100.00	549.0
101.00	3852.0
102.90	1282.0
103.90	2769.0
104.90	2343.0
106.00	858.0
107.00	31152.0
108.00	4907.0
109.00	850.0
110.00	59288.0
111.00	9263.0
111.90	1107.0
113.00	311.0
116.00	1658.0
117.00	24120.0
118.00	1873.0
120.00	469.0
122.00	2046.0
123.00	2970.0
123.90	1305.0
125.00	1303.0
127.00	115088.0
128.00	8134.0
129.00	38648.0
130.00	3301.0
130.90	608.0
134.00	1200.0
135.00	3529.0
135.90	1466.0
137.00	1843.0
138.00	380.0
139.90	560.0
141.00	5294.0
142.00	2050.0
143.00	1091.0
144.90	335.0
146.00	1011.0
147.00	3110.0
148.00	6583.0
149.00	1384.0
150.00	340.0
151.10	585.0
151.90	503.0
153.00	1692.0
153.90	1473.0
155.00	2933.0
156.00	4943.0
157.00	1003.0

Instrument :
BNA_N
ClientSampleId :
DFTPP154

158.00	1145.0
158.90	945.0
160.00	1838.0
161.00	2687.0
162.00	822.0
163.90	345.0
165.00	2144.0
166.00	1595.0
167.00	11470.0
168.00	7727.0
169.00	1160.0
170.00	426.0
171.00	561.0
172.00	774.0
172.90	1269.0
173.90	2761.0
175.00	4601.0
176.00	1658.0
177.00	2349.0
177.90	820.0
179.00	8825.0
180.00	6552.0
181.00	3104.0
181.90	342.0
183.90	681.0
185.00	4150.0
186.00	32592.0
187.00	9864.0
188.00	1072.0
189.00	1920.0
189.90	326.0
191.10	957.0
192.00	3142.0
193.00	3515.0
194.00	758.0
195.20	359.0
196.00	8132.0
197.00	1639.0
197.90	259840.0
199.00	18016.0
200.00	1258.0
201.50	1114.0
203.00	1773.0
204.00	9188.0
205.00	15262.0
206.00	69136.0
207.00	10407.0
207.90	2367.0
208.90	1103.0
209.90	868.0
211.00	3125.0
212.90	347.0
214.90	717.0
216.00	1349.0
216.90	17880.0
218.00	2424.0
221.00	11403.0
221.80	2661.0
223.00	3801.0
224.00	37944.0
225.00	9333.0
226.00	1211.0
227.00	15675.0
228.00	2283.0
229.00	3551.0
229.90	380.0
231.00	1091.0
233.90	974.0
234.90	1249.0
235.90	803.0
237.00	1590.0
238.90	630.0
239.90	453.0
240.90	1038.0
242.00	1940.0
243.00	2216.0
244.00	29856.0
245.10	4008.0
246.00	5765.0
247.00	1141.0
249.00	998.0
253.00	817.0
254.00	1249.0
255.00	153344.0
256.00	22256.0
257.00	1469.0
258.00	8345.0
259.00	1027.0
264.90	3258.0
266.00	456.0
272.00	435.0
273.00	4542.0
274.00	11807.0
275.00	69176.0
276.00	9164.0

Instrument :

BNA_N

ClientSampleId :

DFTPP154

277.00	5199.0
278.00	840.0
283.00	600.0
284.00	535.0
285.00	897.0
292.00	334.0
293.00	1449.0
293.90	338.0
295.00	494.0
296.00	18104.0
297.00	2169.0
303.00	2201.0
304.00	671.0
307.80	352.0
314.00	868.0
315.00	2318.0
316.00	1275.0
321.00	697.0
322.00	487.0
323.00	6904.0
324.00	1204.0
326.90	1198.0
328.00	632.0
332.00	587.0
333.00	607.0
334.00	4109.0
335.00	1039.0
341.00	905.0
345.90	1247.0
346.90	319.0
352.00	1786.0
353.00	1324.0
354.00	2370.0
355.00	399.0
365.00	8426.0
365.90	1631.0
370.90	516.0
372.00	3065.0
372.90	738.0
382.90	1184.0
383.90	395.0
389.90	595.0
390.90	393.0
400.90	314.0
402.00	1603.0
402.90	2321.0
404.00	686.0
420.90	1926.0
421.90	1997.0
423.00	15932.0
424.00	3373.0
424.90	457.0
441.00	42440.0
442.00	315584.0
443.10	54592.0
444.00	5054.0
445.00	310.0

Instrument :
BNA_N
ClientSampleId :
DFTPP154