

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN070918\
 Data File : BN002004.D
 Acq On : 09 Jul 2018 18:35
 Operator : JU/SJ
 Sample : J3892-10
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SVOC-GPC2-BLANK

Quant Time: Jul 10 02:08:42 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN061918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 10 01:03:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	376178	20.00	ng/ul	0.00
18) Naphthalene-d8	10.48	136	1769539	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.33	164	1155458	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.08	188	2771831	20.00	ng/ul	0.00
75) Chrysene-d12	21.28	240	3186967	20.00	ng/ul	-0.02
83) Perylene-d12	23.51	264	3395650	20.00	ng/ul	-0.02

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL
5) Phenol-d5	0.00	99	0	0.00	ng/ul
7) Bis-(2-Chloroethyl)ether-d	0.00	67	0d	0.00	ng/ul
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul
13) 4-Methylphenol-d8	0.00	113	0	0.00	ng/ul
19) Nitrobenzene-d5	0.00	128	0	0.00	ng/ul
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul
26) 2,4-Dichlorophenol-d3	0.00	165	0	0.00	ng/ul
29) 4-Chloroaniline-d4	0.00	131	0d	0.00	ng/ul
43) Dimethylphthalate-d6	0.00	166	0	0.00	ng/ul
46) Acenaphthylene-d8	0.00	160	0	0.00	ng/ul
51) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul
57) Fluorene-d10	0.00	176	0d	0.00	ng/ul
62) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul
70) Anthracene-d10	0.00	188	0d	0.00	ng/ul
76) Pyrene-d10	0.00	212	0	0.00	ng/ul
87) Benzo(a)pyrene-d12	0.00	264	0d	0.00	ng/ul

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070918\
Data File : BN002004.D
Acq On : 09 Jul 2018 18:35
Operator : JU/SJ
Sample : J3892-10
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SVOC-GPC2-BLANK

Quant Time: Jul 10 02:08:42 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN061918.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 10 01:03:06 2018
Response via : Initial Calibration

