

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN071019\
Data File : BN006774.D
Acq On : 10 Jul 2019 17:40
Operator : HP/JU
Sample : PB121119BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK19

Manual Integrations
APPROVED

mohammad
7/11/2019 9:10:14 AM

Quant Time: Jul 11 02:37:37 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN062719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 11 02:31:55 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	2521	0.40	ng/ul	-0.03
2) Naphthalene-d8	10.36	136	9033	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.24	164	4841	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.00	188	10576m	0.40	ng/ul	-0.01
16) Chrysene-d12	21.24	240	8787m	0.40	ng/ul	0.00
20) Perylene-d12	23.46	264	8182m	0.40	ng/ul	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	11.98	152	3322	0.24	ng/ul	0.00
14) Fluoranthene-d10	19.05	212	8499m	0.27	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN071019\
Data File : BN006774.D
Acq On : 10 Jul 2019 17:40
Operator : HP/JU
Sample : PB121119BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK19

Manual Integrations
APPROVED

mohammad
7/11/2019 9:10:14 AM

Quant Time: Jul 11 02:37:37 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN062719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 11 02:31:55 2019
Response via : Initial Calibration

