

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN071019\
 Data File : BN006781.D
 Acq On : 10 Jul 2019 21:44
 Operator : HP/JU
 Sample : K3596-05
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 JLJF9

Manual Integrations
 APPROVED

mohammad
 7/11/2019 9:10:50 AM

Quant Time: Jul 11 03:51:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN062719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 11 02:31:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	2607	0.40	ng/ul	-0.02
2) Naphthalene-d8	10.36	136	9651	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.24	164	5350	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.00	188	11174m	0.40	ng/ul	-0.01
16) Chrysene-d12	21.23	240	9073m	0.40	ng/ul	-0.01
20) Perylene-d12	23.45	264	13545	0.40	ng/ul	-0.04
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.97	152	5849	0.39	ng/ul	-0.02
14) Fluoranthene-d10	19.05	212	8217m	0.25	ng/ul	0.00
Target Compounds						
3) Naphthalene	10.41	128	829	0.030	ng/ul#	Qvalue 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN071019\
Data File : BN006781.D
Acq On : 10 Jul 2019 21:44
Operator : HP/JU
Sample : K3596-05
Misc :
ALS Vial : 10 Sample Multiplier: 1




Instrument :
BNA_N
ClientSampleId :
JLJF9

Manual Integrations
APPROVED

mohammad
7/11/2019 9:10:50 AM

Quant Time: Jul 11 03:51:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN062719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 11 02:31:55 2019
Response via : Initial Calibration





#3

Naphthalene

Concen: 0.030 ng/ul

RT: 10.41 min Scan# 2325

Delta R.T. -0.01 min

Lab File: BN006781.D

Acq: 10 Jul 2019 21:44

Instrument :

BNA_N

ClientSampleId :

JLJF9

Tgt Ion:	128	Resp:	829
Ion Ratio	Lower	Upper	
128	100		
129	61.4	9.4	14.0#
127	68.9	10.6	16.0#

Manual Integrations
APPROVED

mohammad
7/11/2019 9:10:50 AM

