

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071119\
 Data File : BN006796.D
 Acq On : 11 Jul 2019 14:08
 Operator : HP/JU
 Sample : K3578-06
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 JLJG6

Manual Integrations
 APPROVED

mohammad
 7/12/2019 10:26:32 PM

Quant Time: Jul 12 07:24:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN062719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 12 06:39:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	3318	0.40	ng/ul	-0.01
2) Naphthalene-d8	10.35	136	12295	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.24	164	9907	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.00	188	14905m	0.40	ng/ul	-0.01
16) Chrysene-d12	21.23	240	11376m	0.40	ng/ul	-0.01
20) Perylene-d12	23.45	264	10547m	0.40	ng/ul	-0.02
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.96	152	8161	0.43	ng/ul	-0.02
14) Fluoranthene-d10	19.04	212	18363m	0.42	ng/ul	0.00
Target Compounds						
3) Naphthalene	10.40	128	4132	0.117	ng/ul#	Ovalue 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071119\
Data File : BN006796.D
Acq On : 11 Jul 2019 14:08
Operator : HP/JU
Sample : K3578-06
Misc :
ALS Vial : 9 Sample Multiplier: 1




Instrument :
BNA_N
ClientSampleId :
JLJG6

Manual Integrations
APPROVED

mohammad
7/12/2019 10:26:32 PM

Quant Time: Jul 12 07:24:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN062719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 12 06:39:20 2019
Response via : Initial Calibration





#3

Naphthalene

Concen: 0.117 ng/ul

RT: 10.40 min Scan# 2323

Delta R.T. -0.01 min

Lab File: BN006796.D

Acq: 11 Jul 2019 14:08

Instrument :

BNA_N

ClientSampleId :

JLJG6

Tot Ion:128 Resp: 4132

Ion Ratio Lower Upper

128 100

129 104.0 9.4 14.0#

127 41.1 10.6 16.0#

Manual Integrations
APPROVED

mohammad
7/12/2019 10:26:32 PM

