


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071221\
 Data File : BN015483.D
 Acq On : 12 Jul 2021 09:44
 Operator : CG/JU
 Sample : DFTPP155
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP155

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062121.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jul 12 11:02:52 2021



AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2468

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.0	74269	PASS
68	69	0.00	2	1.7	1339	PASS
69	198	0.00	100	34.7	78128	PASS
70	69	0.00	2	0.6	446	PASS
127	198	10	80	45.6	102616	PASS
197	198	0.00	2	0.6	1343	PASS
198	198	100	100	100.0	225173	PASS
199	198	5	9	6.7	15061	PASS
275	198	10	60	23.8	53587	PASS
365	198	1	100	2.6	5875	PASS
441	198	0.01	100	12.5	28047	PASS
442	198	50	100	87.2	196448	PASS
443	442	15	24	17.3	33899	PASS

m/z	Abundance
37.00	407.0
38.00	839.0
39.00	5204.0
40.00	897.0
44.00	1486.0
49.00	862.0
50.00	17432.0
51.00	72776.0
52.00	3940.0
54.90	452.0
56.00	2101.0
57.00	5538.0
60.90	974.0
62.00	941.0
63.00	3145.0
64.00	402.0
65.00	1571.0
68.00	1438.0
69.00	76792.0
70.00	494.0
73.00	786.0
74.00	7919.0
75.00	12231.0
76.00	3870.0
77.00	91400.0
78.00	6272.0
79.00	5444.0
80.00	4097.0
81.00	6133.0
82.00	1411.0
83.00	1638.0
83.90	514.0
85.00	913.0
85.90	2027.0
86.90	798.0
88.00	353.0
91.00	1548.0
92.00	1754.0
93.00	9865.0
93.90	605.0
96.00	447.0
98.00	7238.0
99.00	7155.0
100.00	745.0
101.00	3566.0
103.00	1417.0
103.90	2030.0
105.00	1988.0
106.00	746.0
107.00	26656.0
108.00	4063.0
109.00	724.0
110.00	51128.0
111.00	7779.0
111.90	932.0
113.00	389.0
116.00	1472.0
117.00	20456.0
118.00	1536.0
122.00	1809.0
123.00	2497.0
124.00	1465.0
125.00	1061.0
126.20	604.0
127.00	94032.0
128.00	7091.0
129.00	33880.0
130.00	3107.0
130.90	608.0
134.00	909.0
135.00	2893.0
136.00	1190.0
137.00	1799.0
137.90	327.0
139.90	411.0
141.00	4658.0
142.00	1464.0
142.90	1050.0
146.00	685.0
147.00	2382.0
148.00	5318.0
149.00	1326.0
149.90	388.0
151.10	659.0
151.90	411.0
153.00	1469.0
154.00	1150.0
155.00	2376.0
156.00	3375.0
157.10	764.0
158.00	1013.0

Instrument :
BNA_N
ClientSampleId :
DFTPP155

159.00	429.0
160.00	1507.0
160.90	2098.0
162.00	599.0
165.00	1626.0
165.90	1198.0
167.00	9531.0
168.00	7499.0
169.00	1033.0
170.90	343.0
172.00	810.0
172.90	899.0
174.00	1921.0
175.00	3825.0
175.90	783.0
177.00	1676.0
177.90	584.0
179.00	6380.0
180.00	5293.0
180.90	2088.0
183.80	439.0
185.00	3159.0
186.00	24344.0
187.00	7554.0
187.90	864.0
188.90	1275.0
190.90	789.0
192.00	2455.0
193.00	2358.0
193.90	530.0
195.10	350.0
195.90	5073.0
196.80	1069.0
197.00	1119.0
198.00	187072.0
198.90	13925.0
199.90	1066.0
201.50	827.0
202.90	1105.0
204.00	6794.0
205.00	11012.0
206.00	47440.0
207.00	7666.0
208.00	1627.0
208.90	523.0
211.00	1822.0
214.90	652.0
215.90	807.0
217.00	12198.0
217.90	1576.0
221.00	7609.0
221.90	1655.0
223.00	2701.0
224.00	25472.0
225.00	6698.0
226.00	721.0
227.00	9524.0
228.00	1569.0
229.00	2156.0
231.00	844.0
234.00	791.0
235.00	849.0
236.00	607.0
237.00	970.0
239.00	378.0
240.90	555.0
242.00	1380.0
243.00	1634.0
244.00	18640.0
245.00	2282.0
245.90	3790.0
246.90	609.0
248.90	544.0
252.90	378.0
254.00	670.0
255.00	92800.0
256.00	15021.0
256.90	1477.0
258.00	5317.0
259.00	832.0
264.90	2042.0
266.00	352.0
270.80	316.0
273.00	2603.0
274.00	7709.0
275.00	40184.0
276.00	5718.0
277.00	3510.0
277.90	772.0
282.90	430.0
284.00	309.0
284.90	441.0
293.00	787.0
296.00	10519.0
297.00	1526.0

Instrument :

BNA_N

ClientSampleId :

DFTPP155

302.90	1077.0
304.10	456.0
314.00	382.0
314.90	1230.0
316.00	698.0
320.90	312.0
323.00	3203.0
324.00	752.0
326.90	612.0
328.00	379.0
333.00	405.0
334.00	1939.0
335.00	531.0
341.00	357.0
346.00	742.0
352.00	739.0
352.90	718.0
354.00	1268.0
365.00	4460.0
365.90	610.0
372.00	1875.0
372.90	342.0
382.90	434.0
402.00	766.0
403.00	967.0
421.00	863.0
422.00	800.0
423.00	7140.0
424.00	2745.0
441.10	15062.0
442.00	113952.0
443.00	23464.0
443.90	1995.0

Instrument :
BNA_N
ClientSampleId :
DFTPP155

m/z	Abundance
40.00	926.0
44.00	1443.0
49.00	470.0
83.90	378.0
206.90	918.0

Instrument :
BNA_N
ClientSampleId :
DFTPP155

m/z	Abundance
37.00	353.0
38.00	1046.0
39.00	6350.0
40.00	966.0
41.00	361.0
44.00	1428.0
49.00	1004.0
50.00	21928.0
51.00	87736.0
52.00	4501.0
55.00	469.0
56.00	2506.0
57.00	6377.0
61.00	1142.0
62.00	1206.0
63.00	3466.0
64.00	447.0
64.90	1722.0
68.00	1498.0
69.00	91256.0
70.00	403.0
73.00	931.0
74.00	9200.0
75.00	15018.0
76.00	5013.0
77.00	110240.0
78.00	7895.0
79.00	7167.0
80.00	5090.0
81.00	7657.0
82.00	1849.0
83.00	1602.0
84.00	739.0
85.00	1174.0
86.00	2135.0
87.00	960.0
87.90	444.0
91.00	1622.0
92.00	1908.0
93.00	12307.0
94.00	876.0
96.00	641.0
98.00	9204.0
99.00	8245.0
100.00	689.0
101.00	4362.0
103.00	1501.0
104.00	2435.0
105.00	2841.0
106.00	977.0
107.00	33056.0
108.00	5436.0
109.10	1072.0
110.00	65440.0
111.00	9844.0
112.00	1077.0
112.90	422.0
116.00	1591.0
117.00	25472.0
118.00	2045.0
119.80	436.0
122.00	2281.0
123.00	3375.0
124.00	1403.0
124.90	1594.0
127.00	122720.0
128.00	9404.0
129.00	44160.0
130.00	3830.0
130.90	785.0
132.00	455.0
133.90	1158.0
135.00	3636.0
136.00	1514.0
137.00	2292.0
137.90	424.0
140.00	446.0
141.00	5673.0
142.00	2076.0
143.00	1450.0
143.90	307.0
145.10	311.0
146.00	1053.0
147.00	3085.0
148.00	6826.0
149.00	1671.0
150.00	482.0
151.10	838.0
152.00	445.0
152.90	1984.0
154.00	1564.0

Instrument :
BNA_N
ClientSampleId :
DFTPP155

m/z	Abundance
38.10	828.0
39.00	4306.0
40.00	1012.0
44.00	1712.0
49.00	790.0
50.00	15140.0
51.00	62296.0
52.00	2993.0
55.00	436.0
56.00	1543.0
57.00	4717.0
60.90	940.0
62.00	1103.0
63.00	2307.0
64.00	391.0
65.10	1190.0
68.00	1080.0
69.00	66336.0
70.00	441.0
73.00	697.0
74.00	6989.0
75.00	10678.0
76.00	3430.0
77.00	81992.0
78.00	5599.0
79.00	4831.0
80.00	4054.0
81.00	5550.0
82.00	1441.0
83.00	1339.0
83.90	606.0
85.00	965.0
86.00	1685.0
87.00	693.0
87.90	444.0
91.00	1093.0
92.00	1492.0
93.00	8569.0
93.90	631.0
96.00	517.0
98.00	6751.0
99.00	6035.0
100.00	586.0
101.00	2981.0
103.00	1039.0
104.00	2033.0
105.00	1834.0
106.00	683.0
107.00	25264.0
108.00	3713.0
108.90	745.0
110.00	51048.0
111.00	6996.0
112.00	999.0
115.90	1357.0
117.00	20272.0
117.90	1790.0
122.00	1669.0
123.00	2605.0
124.00	1007.0
124.90	1229.0
126.20	399.0
127.00	91096.0
128.00	6724.0
129.00	34328.0
130.00	2560.0
131.00	633.0
133.90	945.0
135.00	2715.0
136.00	1235.0
137.00	1765.0
140.00	377.0
140.90	4953.0
142.00	1672.0
142.90	1150.0
144.00	323.0
145.90	903.0
146.90	2191.0
148.00	4925.0
149.00	1200.0
150.10	314.0
150.90	615.0
152.00	361.0
153.00	1434.0
154.00	1037.0
155.00	2675.0
156.00	3956.0
157.00	1039.0
158.00	752.0
158.90	604.0
159.90	1592.0

Instrument :
BNA_N
ClientSampleId :
DFTPP155

160.90	2178.0
162.00	630.0
164.90	1883.0
166.00	1538.0
167.00	9802.0
168.00	6016.0
169.00	962.0
170.90	514.0
171.90	817.0
173.00	1185.0
174.00	2176.0
175.00	4382.0
175.90	1222.0
177.00	1766.0
177.90	632.0
179.00	7343.0
180.00	5575.0
181.00	2269.0
182.00	335.0
183.90	528.0
185.00	3651.0
186.00	28576.0
187.00	8283.0
187.90	1119.0
189.00	1756.0
191.00	1015.0
192.00	2577.0
193.00	2650.0
193.90	708.0
194.90	439.0
196.00	6283.0
196.80	1316.0
197.10	1265.0
197.90	219712.0
199.00	14211.0
200.00	1165.0
201.40	951.0
202.90	1497.0
204.00	7611.0
205.00	12095.0
206.00	53456.0
207.00	8001.0
207.90	1822.0
208.90	675.0
209.90	758.0
210.20	681.0
211.00	2039.0
215.00	578.0
215.90	1159.0
217.00	13833.0
217.90	1897.0
221.00	9108.0
221.80	2170.0
223.00	3241.0
224.10	28560.0
225.00	7743.0
226.00	932.0
227.00	12821.0
228.00	1714.0
229.00	2539.0
229.90	353.0
231.00	932.0
233.90	797.0
234.90	1039.0
235.90	626.0
237.00	1067.0
238.90	459.0
239.90	364.0
240.90	743.0
242.00	1499.0
243.00	1657.0
244.00	25248.0
245.00	3252.0
246.00	4755.0
246.90	883.0
248.90	920.0
252.90	536.0
255.00	129392.0
256.00	18136.0
257.00	1234.0
257.90	6224.0
259.00	1044.0
265.00	2967.0
266.00	451.0
271.00	341.0
272.10	394.0
273.00	3631.0
274.00	10310.0
275.00	58984.0
276.00	7860.0
277.00	4930.0
278.00	649.0
283.00	612.0
284.00	465.0
284.90	838.0

Instrument :
BNA_N
ClientSampleId :
DFTPP155

293.00	1233.0
296.00	13135.0
297.00	2030.0
301.00	313.0
302.00	333.0
303.00	1771.0
304.10	384.0
314.00	856.0
314.90	1726.0
315.90	1150.0
321.00	445.0
321.90	358.0
323.00	5317.0
324.10	801.0
327.00	1120.0
327.90	495.0
331.80	339.0
333.00	472.0
334.00	3557.0
335.00	785.0
341.00	532.0
346.00	1199.0
352.00	1858.0
353.00	1189.0
354.00	1762.0
355.00	384.0
364.90	6706.0
365.90	1118.0
370.90	539.0
372.00	2918.0
372.90	646.0
383.00	779.0
390.00	485.0
390.90	312.0
401.90	1414.0
403.00	1534.0
404.00	730.0
421.10	1436.0
422.00	1928.0
423.00	11799.0
424.00	3048.0
441.00	32576.0
442.00	236736.0
443.10	36896.0
444.00	4252.0

Instrument :
BNA_N
ClientSampleId :
DFTPP155

155.00	3295.0
156.00	5304.0
157.00	1042.0
157.90	1167.0
159.00	734.0
160.00	1824.0
161.00	3292.0
161.90	820.0
164.00	343.0
164.90	2025.0
166.00	1873.0
167.00	11922.0
168.00	8079.0
169.00	1260.0
169.90	434.0
171.00	630.0
171.90	1085.0
173.00	1242.0
174.00	2547.0
175.00	4569.0
176.00	1584.0
177.00	2092.0
178.00	866.0
179.00	8562.0
180.00	6013.0
181.00	3166.0
182.00	323.0
182.80	377.0
184.00	635.0
185.00	3977.0
186.00	35560.0
187.00	9922.0
188.00	1133.0
189.00	1962.0
191.00	1194.0
192.00	2806.0
193.00	3416.0
194.00	845.0
195.00	374.0
196.00	7146.0
197.10	1644.0
198.00	268736.0
199.00	17048.0
200.00	1443.0
201.50	1034.0
202.90	1577.0
204.00	8665.0
205.00	14859.0
206.00	62248.0
207.00	9534.0
208.00	2387.0
209.00	673.0
210.00	1185.0
211.00	2434.0
212.00	416.0
215.00	723.0
216.00	1390.0
217.00	16440.0
217.90	2427.0
221.00	11389.0
221.70	2048.0
223.00	3923.0
224.00	36304.0
225.00	8413.0
226.00	1148.0
227.00	14570.0
228.00	1901.0
229.00	3265.0
229.90	335.0
231.00	1271.0
234.00	750.0
234.90	1112.0
236.00	767.0
237.00	1389.0
238.90	512.0
239.90	457.0
240.90	828.0
242.00	2009.0
243.00	2075.0
244.10	25768.0
245.00	3702.0
246.00	5048.0
246.90	1282.0
249.00	1097.0
251.00	301.0
252.00	408.0
253.00	577.0
255.00	141632.0
256.00	20632.0
257.00	1885.0
258.00	7946.0
258.90	1424.0
265.00	3106.0
266.00	531.0
271.00	353.0

Instrument :
BNA_N
ClientSampleId :
DFTPP155

271.90	444.0
273.00	4119.0
274.00	10924.0
275.00	61592.0
276.00	8519.0
277.00	5134.0
277.90	979.0
283.00	535.0
283.90	393.0
285.00	955.0
289.00	311.0
292.00	426.0
292.90	1155.0
294.00	375.0
295.90	16149.0
297.00	2257.0
303.00	2394.0
303.90	439.0
313.90	742.0
314.90	1738.0
315.90	1062.0
321.00	447.0
321.90	344.0
323.00	4908.0
324.00	1130.0
327.00	1028.0
328.00	574.0
331.90	390.0
332.90	573.0
334.00	3417.0
335.00	821.0
341.00	674.0
345.90	1162.0
346.90	312.0
352.00	1620.0
352.90	1300.0
354.00	2086.0
365.00	6460.0
366.00	1076.0
371.00	418.0
371.90	1959.0
373.00	760.0
383.00	842.0
389.90	363.0
402.00	1349.0
403.00	1449.0
404.00	746.0
421.00	1827.0
422.00	1478.0
423.00	10970.0
424.00	3495.0
425.00	404.0
441.00	36504.0
442.00	238656.0
443.00	41336.0
444.00	4088.0

Instrument :
BNA_N
ClientSampleId :
DFTPP155